

101.701-24/06– including *Taq* pol., IFU-01  
 101.701-24u/06u – without *Taq* pol., IFU-02

Visit [www.olerup-ssp.com](http://www.olerup-ssp.com) for  
 “Instructions for Use” (IFU)

Lot No.: **91S**

Lot-specific information

## CERTIFICATE OF ANALYSIS

### Olerup SSP® HLA-A-B-DR SSP Combi Tray

**Product number:** 101.701-24/06 – including *Taq* pol.  
 101.701-24u/06u – without *Taq* pol.  
**Lot number:** 91S  
**Expiry date:** 2016-May-01  
**Number of tests:** 24 tests – Product No. 101.701-24/24u  
 6 tests – Product No. 101.701-06/06u  
**Number of wells per test:** 95 + 1

### Well specifications:

| Well No. | Production No. | Well No. | Production No. | Well No. | Production No. |
|----------|----------------|----------|----------------|----------|----------------|
| 1        | 2013-217-01    | 9        | 2012-117-09    | 17       | 2012-117-17    |
| 2        | 2012-117-02    | 10       | 2012-117-10    | 18       | 2013-217-18    |
| 3        | 2013-265-03    | 11       | 2012-117-11    | 19       | 2013-217-19    |
| 4        | 2012-117-04    | 12       | 2013-220-12    | 20       | 2012-117-20    |
| 5        | 2013-265-05    | 13       | 2012-117-13    | 21       | 2013-217-21    |
| 6        | 2012-117-06    | 14       | 2013-217-14    | 22       | 2012-117-22    |
| 7        | 2012-117-07    | 15       | 2013-217-15    | 23       | 2012-117-23    |
| 8        | 2012-117-08    | 16       | 2012-117-16    | 24       | 2012-117-24    |

The specificity of each primer solution of the kit has been tested against 48 well characterized IHWC cell line DNAs.

Additional 5'- and 3'-primers in primer solutions 4, 5, 7, 8, 12, 14 to 17 and 20 to 23 were tested by separately adding one 3'-primer, respectively one 5'-primer. Additional 5'-primers in primer solution 1 and 19 were tested by separately adding one 3'-primer. Additional 3'-primers in primer solutions 3 and 18 were tested by separately adding one 5'-primer.

In primer solutions 2, 11, 12 and 15 one or two 5'-primers were not possible to test, and in primer solutions 3, 6, 8, 12, 13, 16 and 18 one 3'-primer was not possible to test.

| Well No. | Production No. | Well No. | Production No. | Well No. | Production No. |
|----------|----------------|----------|----------------|----------|----------------|
| 25       | 2013-187-01    | 41       | 2013-187-17    | 57       | 2013-187-33    |
| 26       | 2013-187-02    | 42       | 2013-218-18    | 58       | 2013-187-34    |
| 27       | 2013-218-03    | 43       | 2013-187-19    | 59       | 2013-187-35    |
| 28       | 2013-187-04    | 44       | 2013-187-20    | 60       | 2013-187-36    |
| 29       | 2013-187-05    | 45       | 2013-187-21    | 61       | 2013-187-37    |
| 30       | 2013-208-06    | 46       | 2013-208-22    | 62       | 2013-187-38    |
| 31       | 2013-187-07    | 47       | 2013-208-23    | 63       | 2013-187-39    |
| 32       | 2013-221-08    | 48       | 2013-208-24    | 64       | 2013-187-40    |
| 33       | 2013-266-09    | 49       | 2013-187-25    | 65       | 2013-187-41    |
| 34       | 2013-187-10    | 50       | 2013-187-26    | 66       | 2013-218-42    |
| 35       | 2013-264-11    | 51       | 2013-187-27    | 67       | 2013-187-43    |
| 36       | 2013-187-12    | 52       | 2013-187-28    | 68       | 2013-187-44    |
| 37       | 2013-187-13    | 53       | 2013-264-29    | 69       | 2013-187-45    |
| 38       | 2013-187-14    | 54       | 2013-187-30    | 70       | 2013-187-46    |
| 39       | 2013-187-15    | 55       | 2013-187-31    | 71       | 2013-187-47    |
| 40       | 2013-187-16    | 56       | 2013-187-32    | 72       | 2013-218-48    |

101.701-24/06– including *Taq* pol., IFU-01  
 101.701-24u/06u – without *Taq* pol., IFU-02

Visit [www.olerup-ssp.com](http://www.olerup-ssp.com) for  
 “Instructions for Use” (IFU)

Lot No.: **91S**

#### Lot-specific information

The specificity of each primer solution of the kit has been tested against 48 well characterized IHWC cell line DNAs.

Additional 5'-primers and 3'-primers in primer solutions 27, 30, 38, 43, 53, 55 and 64 were tested by separately adding one additional 3'-primer, respectively one additional 5'-primer. Additional 3'-primers in primer solutions 26, 32, 35, 37, 39 and 42 were tested by separately adding one additional 5'-primer. Additional 5'-primers in primer solutions 25, 45, 49 and 72 were tested by separately adding one additional 3'-primer.

In primer mixes 55, 66 and 67 one 5'-primer was not possible to test, and in primer mixes 27, 53, 62 and 66 one or two 3'-primer was not possible to test.

| Well No. | Production No. | Well No. | Production No. | Well No. | Production No. |
|----------|----------------|----------|----------------|----------|----------------|
| 73       | 2013-267-01    | 81       | 2013-267-09    | 89       | 2013-267-17    |
| 74       | 2013-267-02    | 82       | 2013-267-10    | 90       | 2013-267-18    |
| 75       | 2013-267-03    | 83       | 2013-267-11    | 91       | 2013-267-19    |
| 76       | 2013-267-04    | 84       | 2013-267-12    | 92       | 2013-267-20    |
| 77       | 2013-267-05    | 85       | 2013-267-13    | 93       | 2013-267-29    |
| 78       | 2013-267-06    | 86       | 2013-267-14    | 94       | 2013-267-30    |
| 79       | 2013-267-07    | 87       | 2013-267-15    | 95       | 2013-267-31    |
| 80       | 2013-267-08    | 88       | 2013-267-16    |          |                |

The specificity of each primer solution of the kit has been tested against 48 well characterized IHWC cell line DNAs.

Additional 5'- and 3'-primers in primer solutions 81, 83 and 90 were tested by separately adding one 3'-primer, respectively one 5'-primer.

Additional 5'-primers in primer solutions 78, 84, 87 and 88 were tested by separately adding one 3'-primer. Additional 3'-primers in primer solutions 73, 75, 76, 82 and 92 were tested by separately adding one 5'-primer.

One, two or three of the 5'-primers in primer solution 73, 75, 76, 80 to 82, 85, 87, 88 and 92 were not possible to test. One or two of the 3'-primers in primer solution 73, 75, 76, 81, 85, 87 and 94 were not possible to test.

The negative control primer pairs, **Production No. 2013-271-01**, can detect contamination with PCR products diluted  $10^{-7}$ .

**Results:** No false positive or false negative amplifications were obtained.

**Date of approval:** 2014-March-03

**Approved by:**

*Karin Mattsson*

**Production Quality Control**

101.701-24/06– including *Taq* pol., IFU-01  
101.701-24u/06u – without *Taq* pol., IFU-02

Visit [www.olerup-ssp.com](http://www.olerup-ssp.com) for  
“Instructions for Use” (IFU)

Lot No.: **91S**

Lot-specific information

## Declaration of Conformity

**Product name:** Olerup SSP® HLA-A-B-DR SSP Combi Tray  
**Product number:** 101.701-24/24u, 06/06u  
**Lot number:** 91S

**Intended use:** HLA-A, HLA-B and HLA-DR low resolution histo-compatibility testing

**Manufacturer:** Olerup SSP AB  
Franzengatan 5  
SE-112 51 Stockholm, Sweden  
**Phone:** +46-8-717 88 27  
**Fax:** +46-8-717 88 18

We, Olerup SSP AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) ISO 9001:2008 and ISO 13485:2012, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex II List B, conformity assessed using Annex IV, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at Olerup SSP AB, Franzengatan 5, SE-112 51 Stockholm, Sweden.

Notified Body: Lloyd's Register Quality Assurance Limited, Hiramford, Middlemarch Office Village, Siskin Drive, Coventry CV3 4FJ, United Kingdom. (Notified Body number: 0088.)

Stockholm, Sweden  
2015-January-23



Daniel Malica  
Head of QA and Regulatory Affairs