

CERTIFICATE OF ANALYSIS

Olerup SSP® HLA-B*27 unit dose single well SSP

Product number: 101.911-96u – without Taq polymerase
Lot number: 66M
Expiry date: 2014-February-01
Number of tests: 96
Number of wells per test: 1

Well specifications:

Well No.	Production No.
1	2011-903-01

The specificity of the primer solutions of the kit has been tested against 48 well characterized cell line DNAs.

Results: No false positive or false negative amplifications were obtained.

Date of approval: 2011-August-26

Approved by:



Quality Control, Supervisor

Lot No.: **66M**

Lot-specific Information

www.olerup-ssp.com

Declaration of Conformity

Product name: Olerup SSP® HLA-B*27 unit dose single well
Product number: 101.911-96u
Lot number: 66M

Intended use: HLA-B*27 low resolution histocompatibility testing

Manufacturer: Olerup SSP AB
Franzengatan 5
SE-112 51 Stockholm, Sweden
Phone: +46-8-717 88 27
Fax: +46-8-717 88 18

We, Olerup SSP AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) ISO 9001:2008 and ISO 13485:2003, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex II List B, conformity assessed using Annex IV, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at Olerup SSP AB, Franzengatan 5, SE-112 51 Stockholm, Sweden.

The Authorized Representative located within the Community is: Olerup SSP AB.

Notified Body: Lloyd's Register Quality Assurance Limited, Hiramford, Middlemarch Office Village, Siskin Drive, Coventry CV3 4FJ, United Kingdom. (Notified Body number: 0088.)

Stockholm, Sweden
2011-August-26


Ann-Cathrin Jareman
Head of QA and Regulatory Affairs