

CERTIFICATE OF ANALYSIS

Olerup SSP® DNA Size Marker


Product number: 103.202-100/-500
Lot number: 8D5
Volume: Product No. 103.202-100 – 2 x 500 µl
Product No. 103.202-500 – 10 x 500 µl
Concentration: 20 ng/µl
Expiry date: 2018-03-01

10 µl DNA Size Marker was size-separated on a 2% agarose gel stained by ethidium bromide.

Results: Seven fragments of 50, 100, 200, 300, 400, 500 and 1000 bp, were visible on a UV transilluminator.

Date of approval: 2016-03-18

Approved by:



Production Quality Control

Declaration of Conformity

Product name: DNA Size Marker
Product number: 103.202-100/500
Lot number: 8D5

Intended use: DNA Size Marker

Manufacturer: Olerup SSP AB
Franzengatan 5
SE-112 51 Stockholm, Sweden
Phone: +46-8-717 88 27
Fax: +46-8-717 88 18

We, Olerup SSP AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) ISO 9001:2008 and EN ISO 13485:2012, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex III, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at Olerup SSP AB, Franzengatan 5, SE-112 51 Stockholm, Sweden.

Stockholm, Sweden
2016-Mar-21



Daniel Malica
Head of QA and Regulatory Affairs