

Lot No.: 5R2

Lot-specific information

Declaration of Conformity

Product name: DNA Size Marker
Product number: 103.202-100/500
Lot number: 5R2

Intended use: DNA Size Marker

Manufacturer: *CareDx AB*
Franzéngatan 5
SE-112 51 Stockholm, Sweden
Phone: +46-8-508 939 00
Fax: +46-8-717 88 18

We, CareDx AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) EN ISO 13485:2016, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex III, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at *CareDx AB*, Franzéngatan 5, SE-112 51 Stockholm, Sweden.

The Authorized Representative located within the Community is: *CareDx AB*.

Stockholm, Sweden

Date:


2022-12-15

Quality Assurance



Lot No.: **5R2**

Lot-specific information

CERTIFICATE OF ANALYSIS

Olerup SSP[®] DNA Size Marker

Product number: 103.202-100/500

Lot number: 5R2

Volume: Product No. 103.202-100 – 2 x 500 µl

Product No. 103.202-500 – 10 x 500 µl

Concentration: 20 ng/µl

Expiry date: 2024-12-01

10 µl DNA Size Marker was size separated on a 2% agarose gel stained by ethidium bromide.

Results: Seven fragments of 50, 100, 200, 300, 400, 500 and 1000 bp, were visible on a UV transilluminator.

Date of approval: 2022-12-15

Approved by:



Production Quality Control

