

Lot No.: **4H1**

Lot-specific information

CERTIFICATE OF ANALYSIS

Olerup SSP[®] HLA-A*74

Product number:

101.433-06 – including *Taq* polymerase

101.433-06u – without *Taq* polymerase

Lot number:

4H1

Expiry date:

2021-07-01

Number of tests:

6

Number of wells per test:

15+1

Well specifications:

Well No.	Production No.	Well No.	Production No.
1	2019-024-01	9	2019-024-09
2	2019-024-02	10	2019-024-10
3	2019-024-03	11	2019-024-11
4	2019-024-04	12	2019-024-12
5	2019-024-05	13	2019-024-13
6	2019-024-06	14	2019-024-14
7	2019-024-07	15	2019-024-15
8	2019-024-08		

The negative control primer pairs, **Production No. 2018-947-01**, can detect contamination with PCR products diluted 10⁻⁷.

Results of Quality Control: No false positive or false negative amplifications obtained.

Approved by: *Rebecca Salme*

2019-02-20

Production Quality Control



0088

For *In Vitro* Diagnostic Use

MA100 v01 CoA_DoC IVD Annex II List B

Date: February 2019, Rev. No: 00

101.433-06 – including *Taq* polymerase, IFU-01Visit www.labproducts.caredx.com for101.433-06u – without *Taq* polymerase, IFU-02

“Instructions for Use” (IFU)

Lot No.: 4H1**Lot-specific information**

Declaration of Conformity

Product name: Olerup SSP® HLA-A*74**Product number:** 101.433-06/06u**Lot number:** 4H1**Intended use:** HLA-A*74 high resolution histocompatibility testing**Manufacturer:** Olerup SSP AB
Franzengatan 5
SE-112 51 Stockholm, Sweden
Phone: +46-8-508 939 00
Fax: +46-8-717 88 18

We, Olerup SSP AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) EN ISO 13485:2016, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex II List B, conformity assessed using Annex IV, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at Olerup SSP AB, Franzengatan 5, SE-112 51 Stockholm, Sweden.

The Authorized Representative located within the Community is: Olerup SSP AB.

Notified Body: Lloyd's Register Quality Assurance Limited, 1 Trinity Park, Bickenhill Lane, Birmingham B37 7ES, United Kingdom.
(Notified Body number: 0088.)

Stockholm, Sweden

Date: 2019 02 21



Emil Jonsson
Head of Quality Assurance

**0088**

For *In Vitro* Diagnostic Use
MA100 v01 CoA_DoC IVD Annex II List B
Date: February 2019, Rev. No: 00