

102.101-01 – including *Taq* polymerase
 102.101-01u – without *Taq* polymerase

Visit www.caredx.com for
 for "Instructions for Use" (IFU)

Lot No.: **2R8**

Lot-specific information

CERTIFICATE OF ANALYSIS

Olerup SSP® HLA Wipe Test – Negative Control

Product number: 102.101-01 – including *Taq* polymerase
 Product number: 102.101-01u – without *Taq* polymerase
 Lot number: 2R8
 Expiry date: 2026-04-01
 Number of tests: 96
 Number of wells per test: 1-2

Well specification:

Well No.	Production No.
1	2022-404-01

The negative control primer solution has been tested in a dilution series of the corresponding PCR products, 1 to 10³ down to 1 to 10⁹. The Positive Control DNA has been tested with the HLA Wipe Test kit and gives rise to PCR amplicons.

The negative control primer pairs can detect contamination with the corresponding PCR products diluted 1 to 10⁸, equivalent to 0.1 ng starting material.

Results of Quality Control: No false positive or false negative amplifications obtained.

Date of approval: 2022-11-18

Approved by:



Production Quality Control

