

CERTIFICATE OF ANALYSIS**Olerup SSP® KIR HLA Ligand SSP****Product number:**104.201-12 – including *Taq* polymerase104.201-12u –without *Taq* polymerase**Lot number:**

88M

Expiry date:

2014-May-01

Number of tests:

12

Number of wells per test:

5

Well specifications:

Well No.	Production No.
1	2011-827-01
2	2011-827-02
3	2011-827-03
4	2011-827-04
5	2011-827-05

The specificity of each primer solution of the kit has been tested against 48 IHWS cell line DNAs.

Results: No false positive or false negative amplifications were obtained.

Date of approval: 2011-December-02

Approved by:



Production Quality Control

Declaration of Conformity

Product name: Olerup SSP® KIR HLA Ligand

Product number: 104.201-12/12u

Lot number: 88M

Intended use: Determination of HLA-C, HLA-B^{Bw4+} and HLA-A^{Bw4+} KIR ligand sequence motifs.

Manufacturer: Olerup SSP AB
Franzengatan 5
SE-112 51 Stockholm, Sweden
Phone: +46-8-717 88 27
Fax: +46-8-717 88 18

We, Olerup SSP AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) ISO 9001:2008 and ISO 13485:2003, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex III, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at Olerup SSP AB, Franzengatan 5, SE-112 51 Stockholm, Sweden.

Stockholm, Sweden
2011-December-02



Ann-Cathrin Jareman
Head of QA and Regulatory Affairs