

101.704-48/12 – including *Taq* pol., IFU-01
101.704-48u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **6D2**

Lot-specific information

Olerup SSP® DQ-DR SSP Combi Tray

Product number:	101.704-48/12 – including <i>Taq</i> pol. 101.704-48u/12u – without <i>Taq</i> pol.
Lot number:	6D2
Expiry date:	2018-08-01
Number of tests:	48 tests – Product No. 101.704-48/48u 12 tests – Product No. 101.704-12/12u
Number of wells per test:	45 + 1
Storage - pre-aliquoted primers:	dark at -20°C
- PCR Master Mix:	-20°C
- Adhesive PCR seals	RT
- Product Insert	RT

This Product Description is only valid for Lot No. 6D2.

Complete product documentation consists of generic Instructions for Use (IFU), lot specific Product Insert, Worksheet and Certificate.

CHANGES COMPARED TO THE PREVIOUS OLERUP SSP® DQ-DR SSP COMBI TRAY LOT (19Y)

The format of the Product Insert and Worksheet have been changed.

The DQ-DR kit has been redesigned and improved with regards to allelic detection and discrimination and facilitated interpretation.

One well has been added to DQ low, well 14.

The DQ low resolution specificity and interpretation tables have been updated for the HLA-DQB1 alleles described since the previous *Olerup SSP®* DQ-DR Combi Tray lot was made (**Lot No. 19Y**). The kit design is based on IMGT/HLA database 3.22.0.

The primers of the wells detailed below have been exchanged, added or modified compared to the previous lot.

Well	5'-primer	3'-primer	rationale
1	Added	-	5'-primer added for the DQB1*05:105 allele.
4	-	Added	3'-primer added for the DQB1*03:154 allele.
7	-	Added	3'-primer added for the DQB1*03:189 allele.
10	-	Added	3'-primer added for the DQB1*05:98 allele.
14	New	New	New primer pair for the DQB1*06:189 allele.

101.704-48/12 – including *Taq* pol., IFU-01
101.704-48u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: 6D2

Lot-specific information

The DR low resolution specificity and interpretation tables have been updated for the HLA-DRB1 alleles described since the previous *Olerup* SSP® DQ-DR Combi Tray lot was made (**Lot No. 19Y**). The kit design is based on IMGT/HLA database 3.22.0.

The primers of the wells detailed below have been exchanged, modified or added compared to the previous lot.

Well	5'-primer	3'-primer	rationale
17	Added	Added	5'-primer added for improved amplification of the DRB1*15:69 allele, 3'-primers added for the DRB1*15:115N and DRB1*15:127 alleles.
18	Added	-	5'-primer added for the DRB1*12:57 allele.
19	-	Added	3'-primer added for the allelic resolution of the DRB1*13:125 allele.
23	Added	Added	5'-primer added for the DRB1*07:64 allele, 3'-primer added for the DRB1*07:61 allele.
26	-	Added	3'-primer added for the DRB1*10:11 allele.
28	-	Added	3'-primer added for the DRB1*12:58 allele.
30	Added	-	5'-primer added for the DRB1*13:183 allele.
35	Added	-	5'-primer added for the DRB1*08:57 allele.
46	-	-	Negative control moved from well 45.

Change in revision R01 compared to R00:

1. Primer mix 19 does not amplify the DRB1*03:97 and *11:53 alleles. This has been corrected in the Specificity and Interpretation Tables.

Change in revision R02 compared to R01:

1. Primer mix 35 does not amplify the DRB1*08:14 allele. This has been corrected in the Specificity and Interpretation Tables.

Changes in revision R03 compared to R02:

1. Due to sharing of sequence motifs in codon 38 and 47, DRB3*01:14 will also be amplified in primer mixes 19, 20 and 31 and DRB3*01:23 in mix 19, in addition to primer mix 43.

2. The DRB1*13:02:02 allele is amplified in primer mix 20.

The Specificity and Interpretation Tables have been changed.

Changes in revision R04 compared to R03:

1. Primer mix 17 does not amplify the DRB1*16:05:01-16:05:02 and 16:07 alleles. This has been corrected in the Specificity and Interpretation Tables.

Change in revision R05 compared to R04:

1. Primer mix 20 does not amplify the DRB1*14:137N and 14:152N alleles. Primer mix 29 does not amplify the DRB1*14:137N allele. This has been corrected in the Specificity and Interpretation Tables.

101.704-48/12 – including *Taq* pol., IFU-01
101.704-48u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **6D2**

Lot-specific information

Well **46** contains Negative Control primer pairs, that will amplify more than 95% of the Olerup SSP[®] HLA Class I, DRB, DQB1, DPB1 and DQA1 amplicons as well as all the amplicons generated by the control primer pairs matching the human growth hormone gene.

HLA-specific PCR product sizes range from 75 to 200 base pairs.
The PCR product generated by the positive control primer pair is 430 base pairs.

Length of PCR product	105	200	105	80	75	80	85
5'-primer¹	164	340	440	45	45	43	36
	5'-CAC ^{3'}	5'-Agg ^{3'}	5'-TTA ^{3'}	5'-Tgg ^{3'}	5'-Tgg ^{3'}	5'-Tgg ^{3'}	5'-TAC ^{3'}
							36
							5'-TAT ^{3'}
3'-primer²	231	2nd I	507	59	58	57	47
	5'-TgC ^{3'}	5'-AAA ^{3'}	5'-TTg ^{3'}	5'-CTC ^{3'}	5'-ggC ^{3'}	5'-CTC ^{3'}	5'-ACA ^{3'}
							48
							5'-gCA ^{3'}
							48
							5'-gCC ^{3'}
							52
							5'-TgT ^{3'}
A*	+	+	+				
B*	+	+	+				
C*	+	+	+				
DRB1				+	+		
DRB3				+	+		
DRB5				+			
DQB1					+		
DPB1						+	
DQA1							+

¹The nucleotide position for HLA class I genes and the codon for HLA class II genes, in the 2nd or 3rd exon, matching the specificity-determining 3'-end of the primer is given. Nucleotide and codon numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.

²The nucleotide position for HLA class I genes and the codon for HLA class II genes, in the 2nd or 3rd exon or the 2nd intron, matching the specificity-determining 3'-end of the primer is given in the anti-sense direction. Nucleotide and codon numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.

101.704-48/12 – including *Taq* pol., IFU-01
101.704-48u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **6D2**

Lot-specific information

PRODUCT DESCRIPTION

DQ-DR SSP Combi Tray

CONTENT

The primer set contains 5'- and 3'-primers for grouping the DQB1 alleles in to the serological groups DQ2 to DQ9.

The primer set contains 5'- and 3'-primers for grouping the DRB1*01:01 to DRB1*10:16 alleles into the corresponding serological groups DR1 to DR18 as well as primer pairs for recognizing the DRB3, DRB4 and DRB5 groups of alleles.

Please note that DQB1 amplifications usually are somewhat less pronounced than e.g. DRB and DQA1 amplifications even when using the same DNA preparation and exactly the same experimental procedures.

PLATE LAYOUT

Each test consists of 46 PCR reactions in a 48 well cut PCR plate. Wells 47 and 48 are empty.

1	2	3	4	5	6	7	8
9	10	11	12	13	14	15	16
17	18	19	20	21	22	23	24
25	26	27	28	29	30	31	NC
33	34	35	36	37	38	39	40
41	42	43	44	45	NC	empty	empty

The 48 well cut PCR plate is marked with 'DQ-DR'.

Well No. 1 is marked with the Lot No. '6D2' in silver/gray ink.

Wells 1 to 14 – DQ low resolution primers.

Wells 15 to 45 – DR low resolution primers.

Well 46 – Negative Control (NC).

A faint row of numbers is seen between wells 1 and 2 or wells 7 and 8 of the PCR trays. These stem from the manufacture of the trays, and should be disregarded.

The PCR plates are covered with a PCR-compatible foil.

Please note: When removing each 48 well PCR plate, make sure that the remaining plates stay covered. Use a scalpel or a similar instrument to carefully cut the foil between the plates.

INTERPRETATION

Only the DQB1 alleles will be amplified by the 14 wells of the DQ low resolution primer set, **wells 1 to 14**. Thus, the interpretation of DQ low resolution typings is not influenced by the DQB2 and DQB3 genes.

Only HLA-DRB alleles will be amplified by the 31 wells of the DR low resolution primer set, **wells 15 to 45**. Thus, the interpretation of DR low resolution typings is not influenced by other HLA class II genes.

101.704-48/12 – including *Taq* pol., IFU-01
101.704-48u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **6D2**

Lot-specific information

UNIQUELY IDENTIFIED ALLELES

All the DQB1 alleles, i.e. **DQB1*05:01 to 05:109**, **DQB1*06:01 to 06:197**, **DQB1*02:01 to 02:64**, **DQB1*03:01 to 03:215** and **DQB1*04:01 to 04:32**, recognized by the HLA Nomenclature Committee in October 2015^{1,2} will be amplified by the primers in the DQ low resolution SSP primer set, **wells 1 to 14**. The DQB1 alleles will be grouped into their corresponding serological specificities³, i.e.:

DQ5(1) =	DQB1*05:01:01-05:05
DQ6(1) =	DQB1*06:01:01-06:44
DQ2 =	DQB1*02:01:01-02:05
DQ3 =	DQB1*03:06, 03:10, 03:14
DQ7(3) =	DQB1*03:01:01-03:01:06, 03:04, 03:09, 03:13, 03:16, 03:19
DQ8(3) =	DQB1*03:02:01-03:02:05, 03:05:01-03:05:04, 03:07-03:08, 03:11, 03:18
DQ9(3) =	DQB1*03:03:02:01-03:03:04, 03:12, 03:15, 03:17, 03:20
DQ4 =	DQB1*04:01-04:02

¹DQB1 alleles listed on the IMGT/HLA web page 2015-October-10, release 3.22.0, www.ebi.ac.uk/imgt/hla.

²Alleles that have been deleted from or renamed in the official WHO HLA Nomenclature up to and including the last IMGT/HLA database release can be retrieved from web page <http://hla.alleles.org/alleles/deleted.html>.

³The serological split of the DQB1*05:05 to 05:109 alleles, the DQB1*06:06 to 06:07, 06:10, 06:13, 06:15-06:24 and 06:27 to 06:197 alleles, the DQB1*02:04-02:64 alleles, the DQB1*03:07-03:09 and 03:11 to 03:215 alleles and the DQB1*04:03:01-04:32 alleles is not known. In this table we have used the expert-assigned serological grouping in Tissue Antigens (2009) 73:95-170, and also inferred the serological grouping from the naming of the sequence-defined allele.

All the HLA-DRB1, -DRB3, -DRB4¹ and -DRB5 alleles, i.e. **DRB1*01:01 to DRB1*16:37**, **DRB3*01:01 to DRB3*03:03**, **DRB4*01:01 to DRB4*01:10** and **DRB5*01:01 to DRB5*02:06**, recognized by the HLA Nomenclature Committee in October 2015^{2,3} will be amplified by the primers in the DR low resolution SSP kit. The HLA-DRB alleles will be grouped into their corresponding serological specificities⁴.

¹The DRB4*02:01N and DRB4*03:01N null alleles will not be amplified by the DR low resolution primer set.

²DRB alleles listed on the IMGT/HLA web page 2015-October-10, release 3.22.0, www.ebi.ac.uk/imgt/hla.

³Alleles that have been deleted from or renamed in the official WHO HLA Nomenclature up to and including the last IMGT/HLA database release can be retrieved from web page <http://hla.alleles.org/alleles/deleted.html>.

⁴The serological split of all DRB1 alleles is not known. In this table we use the expert-assigned serological grouping in Tissue Antigens (2009) 73:95-170 and the serological grouping of the sequence-defined allele.

101.704-48/12 – including *Taq* pol., IFU-01
101.704-48u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **6D2**

Lot-specific information
SPECIFICITY TABLE

DQ low resolution primer set

**Specificities and sizes of the PCR products of the 14 primer mixes
of the DQ low resolution primer set**

Primer Mix	Size of spec. PCR product ¹	Size of control band ²	DQ serology ³	Amplified DQB1 alleles ⁴
1⁸	135 bp, 225 bp	515 bp	5	*05:01:01:01-05:01:15, 05:01:17-05:59, 05:61-05:81, 05:83-05:97, 05:99-05:109
2⁸	135 bp, 185 bp, 220 bp, 270 bp	515 bp	1, 5, 6	*06:01:01-06:145, 06:147-06:168, 06:170-06:197
3	210 bp	430 bp	2	*02:01:01-02:64
4^{6,8}	130 bp, 220 bp	515 bp	3, 7	*03:01:01:01-03:01:31, 03:04:01-03:04:02, 03:09-03:10:02, 03:13-03:14:02, 03:16, 03:19, 03:21-03:22, 03:24, 03:27-03:29, 03:35-03:36, 03:42, 03:44, 03:46-03:60, 03:69, 03:71, 03:73, 03:75-03:77, 03:80, 03:82-03:84N, 03:92-03:94, 03:101-03:103, 03:108-03:109, 03:114-03:116, 03:118N-03:122, 03:127-03:131, 03:133-03:135, 03:138-03:140, 03:142-03:144, 03:147-03:148, 03:150, 03:152, 03:154, 03:157-03:160, 03:162-03:167, 03:169-03:173, 03:180, 03:182-03:183, 03:186-03:188, 03:191-03:198, 03:201-03:202, 03:206-03:208
5⁶	130 bp, 220 bp	515 bp	6, 8	*03:02:01-03:02:19, 03:05:01-03:05:04, 03:07-03:08, 03:11, 03:18, 03:32, 03:37, 03:45, 03:61, 03:63-03:64, 03:66N-03:68, 03:70, 03:85, 03:104, 03:106-03:107, 03:125, 03:132, 03:146, 03:153, 03:161, 03:174-03:175, 03:178-03:179, 03:181, 03:184-03:185, 03:189-03:190, 03:199, 03:203-03:205, 03:210-03:211, 03:213N-03:215, 06:29, 06:123, 06:139
6⁶	135 bp	515 bp	2, 3, 6, 9	*02:03, 03:03:02:01-03:03:13, 03:06, 03:12, 03:15, 03:20, 03:25-03:26, 03:30-03:31, 03:33-03:34, 03:38-03:41, 03:43, 03:65, 03:74, 03:79, 03:86-03:91Q, 03:95N-03:99Q, 03:104-03:105, 03:111-03:113, 03:117, 03:123-03:124, 03:126, 03:136-03:137, 03:141, 03:145, 03:149, 03:155-03:156, 03:168, 03:176-03:177, 03:200, 03:209, 03:212, 04:03:01-04:03:02, 06:03:10, 06:51:01, 06:66, 06:96, 06:168, 06:172
7^{5,6}	85 bp	515 bp	3, 7, 8, 9	*03:01:01:01-03:01:06, 03:01:08-03:02:17, 03:02:19-03:05:04, 03:07, 03:09-03:24, 03:26-03:57, 03:59-03:64, 03:66N-03:103,

101.704-48/12 – including *Taq* pol., IFU-01
 101.704-48u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
 “Instructions for Use” (IFU)

Lot No.: **6D2**

Lot-specific information

				03:105-03:106, 03:108-03:136, 03:138-03:193, 03:195-03:215
8^{6,8}	160 bp, 205 bp	430 bp	4	*03:132, 04:01:01-04:02:07, 04:02:09-04:29, 04:31-04:32
9	225 bp	430 bp	4	*04:01:01-04:03:02, 04:06-04:21, 04:22 ^w , 04:23, 04:25N-04:32
10	215 bp	430 bp	5	*05:01:01:01-05:03:09, 05:03:11-05:03:17, 05:05:01-05:33, 05:35-05:43, 05:45-05:51, 05:53, 05:55-05:71, 05:73-05:76, 05:78-05:104, 05:106-05:109, 06:156, 06:162, 06:169
11	185 bp	430 bp	8, 9	*03:03:11, 03:05:01, 03:05:03, 03:17:01, 03:61, 03:72, 03:100, 03:181
12⁶	185 bp	430 bp	3, 4, 7, 8, 9	*03:01:01:01-03:01:01:03, 03:01:03-03:01:07, 03:01:09-03:02:02, 03:02:04-03:02:12, 03:02:14-03:03:02:04, 03:03:04-03:04:02, 03:05:03-03:17:01, 03:18-03:19, 03:21-03:22, 03:23:02-03:36, 03:38-03:60, 03:62-03:71, 03:74, 03:76-03:98, 03:101-03:103, 03:106-03:108, 03:110-03:111, 03:113-03:117, 03:119-03:131, 03:133-03:153, 03:155, 03:157-03:161, 03:163-03:180, 03:182, 03:184-03:188, 03:190-03:203, 03:205-03:215, 04:01:03
13⁷	185 bp	515 bp	6	*04:10, 06:02:01-06:02:25, 06:14:01-06:16, 06:19:01-06:20, 06:23-06:24, 06:33, 06:37, 06:46-06:50, 06:51:02, 06:68, 06:70-06:84, 06:95, 06:97, 06:107, 06:109, 06:111-06:117, 06:122, 06:124-06:127, 06:136-06:138, 06:146:01-06:147, 06:150-06:152, 06:156, 06:159, 06:161-06:163, 06:166, 06:173-06:175, 06:178-06:179N, 06:182-06:183, 06:188, 06:192, 06:197
14	185 bp	430 bp	1, 5, 6	*06:02:01-06:02:12, 06:02:14-06:03:10, 06:03:12-06:03:21, 06:05:02 [?] -06:06 [?] , 06:08:01-06:08:03, 06:10-06:11:03, 06:13:01-06:14:03, 06:16, 06:18:01-06:20, 06:23-06:24, 06:26N-06:27:02, 06:29-06:33, 06:37, 06:40-06:41, 06:44, 06:47-06:51:02, 06:59-06:65, 06:67-06:68, 06:70-06:78, 06:80-06:84, 06:87, 06:90-06:91, 06:95-06:97, 06:99:01-06:99:02, 06:106-06:107, 06:109-06:117, 06:122-06:128, 06:130-06:131, 06:133-06:134, 06:136-06:139, 06:141, 06:143, 06:145, 06:147-06:148, 06:150-06:152, 06:154, 06:156, 06:159, 06:161-06:163, 06:165-06:166, 06:169-06:170, 06:173-06:176, 06:178-06:179N, 06:182-06:185, 06:187-06:188, 06:190-06:192, 06:195-06:197

¹Alleles are assigned by the presence of specific PCR product(s). However, the sizes of the specific PCR products may be helpful in the interpretation of DQ low resolution SSP subtypings. When the primers in a primer mix can give rise to HLA-specific PCR products of more than one length this is indicated if the size difference is more than 20 base pairs. Size differences of

101.704-48/12 – including *Taq* pol., IFU-01
101.704-48u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: 6D2

Lot-specific information

20 base pairs or less are not given. For high resolution SSP kits, the alleles listed are specified according to amplicon length.

Nonspecific amplifications, i.e. a ladder or a smear of bands, may sometimes be seen. GC-rich primers have a higher tendency of giving rise to nonspecific amplifications than other primers.

PCR fragments longer than the control bands may sometimes be observed. Such bands should be disregarded and do not influence the interpretation of the SSP typings.

PCR fragments migrating faster than the control bands, but slower than a 400 bp fragment may be seen in some gel read-outs. Such bands can be disregarded and do not influence the interpretation of the SSP typings.

Some primers may give rise to primer oligomer artifacts. Sometimes this phenomenon is an inherent feature of the primer pair(s) of a primer mix. More often it is due to other factors such as too low amount of DNA in the PCR reactions, taking too long time in setting up the PCR reactions, working at elevated room temperature or using thermal cyclers that are not pre-heated.

²The internal positive control primer pairs amplify segments of the human growth hormone gene. The internal positive control bands are 430 or 515 base pairs respectively, well distribution as outlined in the table. Well number 1 contains the longer, 515 bp, internal positive control band. The well distribution of the internal controls can help in orientation of the kit on gel photo, as well as allow for kit identification. In the presence of a specific amplification the intensity of the control band often decreases. In the presence of a specific amplification the intensity of the control band often decreases.

³The serological reactivity of all DQ alleles is not known. In this table we use the expert-assigned serological grouping in Tissue Antigens (2009) 73:95-170 and the serological grouping of the sequence-defined allele. The DQB1*03:10 allele has been assigned type DQ7 by NMDP.

⁴For several DQB1 alleles 1st and/or 3rd exon(s) and beyond, as well as intron nucleotide sequences, are not available. In these instances it is not known whether some of the primers of the SSP sets are completely matched with the target sequences or not. Assumption is made that unknown sequences in these regions are conserved within allelic groups.

⁵HLA-specific PCR products shorter than 125 base pairs have a lower intensity and are less sharp than longer PCR products.

⁶Primer mixes 4 to 8 and 12 may give rise to a lower yield of HLA-specific PCR product than the other DQ low primer mixes, most pronounced in primer mixes 4, 7 and 12.

⁷Primer mix 13 may have a tendency of unspecific amplification.

⁸The primer pairs in wells 1, 2, 4 and 8 will in some samples give rise to two HLA-specific PCR fragments.

‘w’, may be weakly amplified.

‘?’, nucleotide sequence information not available for the primer matching sequence.

101.704-48/12 – including *Taq* pol., IFU-01
101.704-48u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **6D2**

Lot-specific information

SPECIFICITY TABLE

DR low resolution primer set

Specificities and sizes of the PCR products of the 31+1 primer
mixes of the DR low resolution primer set

Primer Mix	Size of spec. PCR product ¹	Size of control band ²	DR serology ³	Amplified HLA-DRB alleles ⁴
15^{6,8}	210 bp, 230 bp, 260 bp	515 bp	1	*01:01:01-01:02:12, 01:04-01:38, 01:40N-01:72
16	200 bp	430 bp	1, 103	*01:03, 01:39N, 01:42, 01:61
17^{8,11}	210 bp, 235 bp	430 bp	2, 15, 16	*15:01:01:01-15:127
18¹¹	210 bp	430 bp	11, 16	*11:30, 12:57, 16:01:01-16:05:02, 16:07-16:37
19^{5,6,7}	125 bp, 225 bp	430 bp	3, 11, 17, 18	*03:01:01:01-03:16, 03:18-03:26, 03:28-03:34, 03:36-03:96, 03:98-03:125, 11:07, 11:27:02-11:27:03, 11:84:02-11:84:03, 11:103:01-11:103:02, 11:105, 11:107, 11:125, 11:173, 13:33:01, 13:61:02, 13:94:01, 13:96:01, 14:38:02, 15:25, 15:37:01, 15:100, 15:104
20^{5,6,7}	75 bp, 205 bp	430 bp	3, 6, 11, 13, 14, 17	*03:01:01:01-03:01:23, 03:01:25, 03:04:01-03:06, 03:08-03:16, 03:18-03:20, 03:22-03:23, 03:25-03:26, 03:28, 03:30-03:31, 03:33-03:34, 03:36-03:37, 03:43-03:48, 03:50-03:52, 03:54-03:68N, 03:70-03:73, 03:75-03:86, 03:89, 03:91-03:93, 03:95-03:96, 03:98-03:100:02, 03:104, 03:106-03:110, 03:112-03:114, 03:116-03:118, 03:121-03:125, 08:40, 11:02:01-11:03:02, 11:11:01, 11:11:03, 11:14:01-11:14:02, 11:16, 11:20-11:21, 11:36, 11:40-11:41, 11:48, 11:59, 11:63:01-11:63:02, 11:65:01-11:65:02, 11:68, 11:70, 11:73, 11:76, 11:79-11:80, 11:83, 11:85-11:87, 11:93, 11:118, 11:122, 11:124, 11:127, 11:131-11:132, 11:135, 11:138-11:139, 11:142, 11:151, 11:153, 11:161, 11:168, 11:171, 11:176, 11:182, 11:184, 13:01:01-13:04, 13:08, 13:10, 13:15-13:17, 13:19-13:20, 13:22-13:24, 13:27-13:29, 13:31-13:41, 13:43, 13:45, 13:48, 13:51-13:54, 13:57, 13:59, 13:61:01-13:61:02, 13:63-13:66:02, 13:68-13:76, 13:78-13:81, 13:83-13:85, 13:87-13:99, 13:101-13:102, 13:104-13:107, 13:109, 13:111-13:117, 13:120-13:131, 13:133, 13:135, 13:137N-13:145, 13:147-13:149, 13:151-13:153, 13:155, 13:159, 13:162, 13:165-13:168, 13:170-13:180, 13:182, 13:184-13:188, 13:190-13:191,

101.704-48/12 – including *Taq* pol., IFU-01
101.704-48u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **6D2**

Lot-specific information

				13:193-13:194, 13:196, 13:198, 13:200N-13:202, 13:204-13:205, 13:207-13:209, 14:16, 14:19, 14:21, 14:45, 14:82, 14:95, 14:109, 14:120, 14:122, 14:132
21^{5,6}	85 bp, 210 bp	430 bp	3, 6, 11, 13, 14, 1403, 18	*03:02:01-03:03, 03:27, 03:29, 03:38, 03:53, 03:74, 03:88, 03:90, 03:102-03:103, 03:115, 03:119, 11:13:01 ^w -11:13:02 ^w , 11:26, 11:34, 13:15, 13:19, 13:26:01-13:26:02, 13:44, 13:53, 13:57, 13:85-13:86, 13:104, 13:193, 13:198, 13:206, 14:02:01-14:03:02, 14:06:01-14:06:03, 14:09, 14:12:01-14:13, 14:17-14:21, 14:24, 14:27, 14:29-14:30, 14:32:01 ^w -14:32:02 ^w , 14:33, 14:40-14:41, 14:47-14:49, 14:51, 14:63, 14:65 ^w , 14:67, 14:77-14:78, 14:80-14:81, 14:83, 14:85, 14:89, 14:94, 14:98, 14:102, 14:106, 14:108-14:109, 14:115, 14:119, 14:121, 14:135, 14:146, 14:154, 14:156, 14:159-14:160, 14:165
22^{5,6}	100 bp, 175 bp	430 bp	3, 4	*04:01:01-04:05:11, 04:05:13-04:202
23¹¹	235 bp	430 bp	7	*07:01:01:01-07:01:18, 07:03-07:66
24⁶	170 bp, 215 bp, 250 bp	515 bp	8, 11, 12, 14	*08:01:01-08:19, 08:21-08:56, 08:58-08:66, 08:68-08:77, 11:67, 11:193, 12:04, 12:16:01-12:16:03, 12:22, 12:39, 12:49, 14:11, 14:15, 14:68:01-14:68:02, 14:93, 14:148
25^{5,6}	85 bp, 100 bp, 135 bp, 160 bp, 195 bp	430 bp	3, 9, 11	*03:08, 03:65, 03:112, 09:01:02-09:27, 11:07, 11:103:01-11:103:02, 11:105, 11:107, 11:125, 11:173
26	175 bp	430 bp	10, 11, 13	*03:76, 03:125, 10:01:01-10:16, 11:59, 11:80, 11:83, 11:87, 11:135, 11:142, 11:182, 13:27, 13:41, 13:71, 13:129, 13:176
27^{5,6}	100 bp, 170 bp	430 bp	3, 8, 11, 13, 14	*03:08, 03:65, 08:31, 08:41, 08:63, 08:75, 11:01:01-11:11:01, 11:11:03-11:70, 11:72-11:191, 11:193-11:194
28^{5,6}	85 bp, 110 bp	430 bp	12	*08:32, 08:53, 12:01:01-12:56, 12:58, 14:52
29	220 bp	430 bp	6, 8, 11, 13, 14, 1403	*03:76, 03:125, 08:20-08:21, 11:01:01-11:04:12, 11:06:01-11:06:03, 11:08:01-11:11:01, 11:11:03-11:12:03, 11:14:01-11:16, 11:18-11:21, 11:23:01-11:25, 11:27:01-11:33, 11:35-11:51, 11:54:01-11:54:02, 11:56-11:66, 11:68, 11:70, 11:72-11:81, 11:83-11:88, 11:90-11:97, 11:99-11:102:02, 11:106, 11:108-11:124, 11:126-11:135, 11:137-11:142, 11:144-11:158, 11:160-11:172, 11:174-11:191, 11:194, 12:57, 13:01:01-13:02:01, 13:02:03-13:08, 13:10-13:16, 13:18-13:43, 13:45-13:85, 13:87-13:115, 13:117-13:128, 13:130-13:145, 13:147-13:166, 13:168-13:174, 13:176-13:182, 13:184-13:191, 13:194-13:205, 13:207-13:209, 14:03:01-14:03:02, 14:12:01-

101.704-48/12 – including *Taq* pol., IFU-01
101.704-48u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **6D2**

Lot-specific information

				14:12:02, 14:16, 14:19, 14:21-14:22, 14:25, 14:27, 14:40, 14:53, 14:63, 14:67, 14:69, 14:74, 14:77-14:78, 14:84-14:85, 14:98, 14:102, 14:105, 14:109, 14:115-14:116, 14:128, 14:135, 14:144, 14:156, DRB3*02:27
30^{6,8}	200 bp, 225 bp	430 bp	6, 8, 11, 12, 13, 14	*03:125, 08:01:01-08:01:05, 08:02:01-08:02:04, 08:04:01-08:09, 08:11, 08:16-08:17, 08:20-08:22, 08:24, 08:26, 08:28, 08:31, 08:39, 08:41-08:44, 08:50, 08:52, 08:54-08:55, 08:57, 08:59, 08:64, 08:67, 08:70, 08:72-08:73, 08:75, 08:77, 11:01:01-11:01:17, 11:01:20-11:06:03, 11:09-11:11:01, 11:11:03-11:12:03, 11:14:01-11:16, 11:20-11:21, 11:23:01-11:25, 11:27:01-11:30, 11:32-11:33, 11:35-11:41, 11:43-11:44, 11:46:01-11:51, 11:54:01-11:56, 11:58:01-11:63:02, 11:65:01-11:70, 11:72, 11:74:01-11:78, 11:80-11:88, 11:90-11:97, 11:99-11:102:02, 11:106, 11:108-11:118, 11:120-11:124, 11:126-11:129, 11:133-11:135, 11:137-11:142, 11:144-11:152, 11:154-11:158, 11:160-11:172, 11:174-11:183, 11:185-11:190, 11:193-11:194, 12:02:01-12:02:06, 12:13, 12:15-12:16:03, 12:18-12:21, 12:23, 12:26-12:27, 12:31N-12:33, 12:37, 12:42-12:45, 12:49-12:52, 12:55-12:57, 13:01:01-13:02:01, 13:02:03-13:02:12, 13:04-13:05:03, 13:07:01-13:09, 13:11:01-13:11:02, 13:14:01-13:24, 13:26:01-13:29, 13:31-13:32, 13:34-13:36, 13:38-13:43, 13:45-13:55, 13:57, 13:59, 13:61:01-13:65, 13:67-13:76, 13:78-13:80, 13:83-13:84, 13:87, 13:91-13:93, 13:96:01-13:100, 13:102-13:109, 13:111-13:114, 13:116-13:117, 13:121, 13:123-13:132, 13:135-13:136, 13:138-13:150, 13:153, 13:155, 13:158-13:160, 13:162, 13:164-13:166, 13:168-13:169, 13:171:01-13:171:02, 13:173, 13:175, 13:177, 13:179, 13:182-13:187, 13:189-13:192, 13:195, 13:197-13:205, 13:207-13:210, 14:15-14:16, 14:22, 14:24-14:25, 14:27, 14:37, 14:53, 14:73, 14:105, 14:128
31⁷	175 bp	430 bp	3, 6, 11, 13, 14, 1403, 17, 18	*03:01:01:01-03:07, 03:09, 03:11:01-03:41:02, 03:43-03:45, 03:47-03:63, 03:66-03:68N, 03:70-03:86, 03:88-03:91, 03:93-03:110, 03:112-03:125, 08:20, 13:01:01-13:16, 13:18-13:42, 13:44, 13:46-13:66:02, 13:68-13:102, 13:104-13:115, 13:117-13:121, 13:123-13:158, 13:161-13:164, 13:166-13:170, 13:171:02-13:174, 13:176-13:178, 13:180-13:182, 13:184-13:190, 13:192-13:194, 13:196, 13:198-13:210, 14:02:01-14:03:02, 14:05:01-14:06:03, 14:09, 14:12:01-14:14, 14:17-14:21, 14:23:01,

101.704-48/12 – including *Taq* pol., IFU-01
101.704-48u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **6D2**

Lot-specific information

				14:23:03-14:24, 14:27, 14:29-14:30, 14:33, 14:36-14:37, 14:40-14:45, 14:47-14:48, 14:51, 14:56, 14:59, 14:63-14:65, 14:67, 14:77-14:78, 14:80-14:81, 14:83-14:85, 14:89, 14:91, 14:94-14:96, 14:98, 14:100, 14:102-14:103, 14:106, 14:108-14:109, 14:115-14:116, 14:121, 14:123, 14:127:01-14:127:02, 14:134-14:136, 14:144, 14:154-14:156, 14:159-14:161, 14:165, 14:167
32^{5,6}	100 bp, 150 bp, 195 bp, 240 bp	430 bp	4, 6, 8, 13, 14, 1404	*04:62, 04:69, 04:73, 04:105:01-04:105:02, 04:122, 04:146, 08:08, 11:69, 11:82, 13:45, 13:197, 14:01:01-14:01:02, 14:01:04, 14:04:01-14:04:03, 14:07:01-14:07:02, 14:10, 14:16, 14:22, 14:25-14:26, 14:28, 14:31-14:32:03, 14:35, 14:37-14:39, 14:49-14:50, 14:53-14:54:01, 14:54:03-14:55, 14:57-14:58, 14:60-14:62, 14:68:01-14:71, 14:73-14:76, 14:79, 14:82, 14:86-14:88, 14:90, 14:93, 14:99, 14:101, 14:104-14:105, 14:107, 14:110-14:114, 14:117-14:120, 14:122, 14:124-14:125, 14:128-14:129, 14:131, 14:137N-14:140, 14:142-14:143, 14:145-14:147, 14:149-14:153, 14:157-14:158, 14:163-14:164, 14:166N, 14:168 ^w , DRB4*01:03:01:02N
33^{5,6,8,9}	110 bp, 145 bp, 170 bp	430 bp	3, 6, 9, 11, 13, 14, 1404	*03:10, 09:01:02-09:01:05, 09:01:07-09:02:02, 09:04-09:22, 09:24-09:27, 11:13:01-11:13:02, 11:17, 11:52, 13:43, 13:159, 13:171:01, 13:179, 13:191, 13:193, 14:01:01-14:02:02, 14:04:01-14:11, 14:13-14:14, 14:16-14:18, 14:19 ^w , 14:20, 14:21 ^w , 14:22-14:23:04, 14:26, 14:28-14:36, 14:38:01-14:39, 14:41, 14:43-14:52, 14:54:01-14:54:03, 14:54:04 ^w , 14:54:05-14:57, 14:59-14:62, 14:64-14:65, 14:68:01-14:68:02, 14:70-14:76, 14:79-14:83, 14:86-14:88, 14:90-14:97, 14:99-14:101, 14:103-14:108, 14:109 ^w , 14:110-14:114, 14:117-14:127:02, 14:129-14:134, 14:137N-14:140, 14:142-14:143, 14:145-14:155, 14:157-14:160, 14:162-14:167, 15:27, 15:34, 15:66:01-15:66:02, 15:122 ^w
34^{5,6}	110 bp, 150 bp, 180 bp, 215 bp	430 bp	2, 3, 4, 6, 8, 11, 13, 14, 1403, 1404, 16	*03:10, 08:09, 08:20-08:21, 08:32, 08:35, 08:36:02, 08:53, 08:68 ^w , 08:70 ^w , 11:13:01-11:13:02, 11:17, 11:23:01-11:23:02, 11:25, 11:31, 11:45, 11:52, 11:55, 11:64, 11:89, 11:96, 11:119, 11:148, 11:159, 13:13, 13:18, 13:43, 13:45, 13:47, 13:55, 13:119, 13:144, 13:146, 13:154, 13:156, 13:158-13:159, 13:164, 13:171:01, 13:179, 13:191, 13:197, 14:01:01-14:01:04, 14:03:01-14:05:04, 14:07:01-14:08, 14:10-14:12:02, 14:14-14:16, 14:18, 14:22-14:23:04, 14:25-14:28, 14:31-14:32:03, 14:34-14:36, 14:38:01-14:40, 14:42-14:45, 14:49-14:50, 14:53-14:65, 14:67-14:79, 14:81-14:82, 14:84-14:93,

101.704-48/12 – including *Taq* pol., IFU-01
101.704-48u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: 6D2

Lot-specific information

				14:95-14:97, 14:99-14:105, 14:107, 14:110-14:120, 14:122-14:140, 14:142-14:158, 14:160-14:164, 14:166N-14:168, 15:21 ^w , 16:04 ^w , 16:18 ^w
35	170 bp	430 bp	8, 12, 13, 14	*03:92, 08:01:01-08:04:05, 08:04:07-08:06, 08:09-08:10, 08:12-08:13, 08:16-08:18, 08:21-08:24, 08:26-08:30:03, 08:32-08:33, 08:35-08:40, 08:42-08:51, 08:53-08:62, 08:64-08:74, 08:76-08:77, 12:09, 12:46, 12:48 ^w , 13:17, 13:116, 13:175, 14:15, 14:52, 14:126:01
36^{5,6,8}	75 bp, 175 bp, 265 bp	515 bp	2, 4, 8, 11, 13, 14	*04:12, 04:18, 04:25, 04:58, 08:04:01, 08:04:02 ^w -08:04:03 ^w , 08:04:04-08:04:07, 08:06, 08:10, 08:12, 08:14, 08:20, 08:22, 08:28, 08:54, 08:59, 08:67, 08:70, 08:75, 11:25, 11:67, 11:119, 13:18, 13:144, 13:156, 13:164, 14:12:01-14:12:02, 14:15, 14:78, 14:84, 14:156, 15:21
37	135 bp	430 bp	8, 12, 13	*03:92, 08:17, 08:28, 08:37, 08:45, 08:63, 11:67, 11:193, 12:01:01-12:20, 12:22-12:37, 12:39-12:41, 12:43-12:52, 12:54-12:56, 12:58, 13:17, 13:175, 14:138
38	180 bp	430 bp	3, 6, 11, 13, 14	*03:12, 13:03:01-13:04, 13:12:01-13:13, 13:21:01-13:21:02, 13:30, 13:32-13:33:03, 13:38, 13:48-13:49, 13:55, 13:58, 13:65-13:66:02, 13:75, 13:81, 13:88-13:90, 13:93-13:95, 13:101, 13:108, 13:115, 13:118, 13:120, 13:122, 13:133-13:134, 13:139, 13:151-13:152, 13:154, 13:161, 13:164, 13:167, 13:169, 13:174, 13:188, 13:194, 14:13, 14:63, 14:65, 14:78, 14:85, DRB5*01:08N
39	150 bp	430 bp	2, 4, 6, 11, 13	*03:125, 11:16, 11:20, 11:40, 11:59, 11:63:02, 11:118, 11:122, 13:01:01-13:02:01, 13:02:03-13:02:12, 13:08, 13:15-13:16, 13:19-13:20, 13:27-13:29, 13:31-13:32, 13:34-13:36, 13:39-13:41, 13:43, 13:51-13:53, 13:57, 13:59, 13:61:01-13:61:02, 13:63-13:65, 13:67-13:69, 13:71-13:74, 13:76, 13:78-13:80, 13:83-13:84, 13:87, 13:91-13:93, 13:96:01-13:97:01, 13:98-13:99, 13:102-13:107, 13:109, 13:111-13:114, 13:117, 13:121, 13:123-13:128, 13:130-13:131, 13:135, 13:138-13:143, 13:145, 13:147-13:148, 13:153, 13:155, 13:159-13:160, 13:165-13:166, 13:168, 13:171:01-13:171:02, 13:173, 13:175, 13:177, 13:179, 13:182-13:187, 13:190-13:191, 13:195, 13:198, 13:200N-13:202, 13:205, 13:207-13:210, 14:16, 14:57, 15:10
40	145 bp	430 bp	3, 6, 11, 13, 14, 17	*03:01:01:01-03:01:22, 03:01:24-03:01:25, 03:04:01-03:05:03, 03:07-03:16, 03:18-03:25, 03:28-03:30, 03:32-03:34, 03:36-03:37, 03:39-03:40, 03:43-03:52, 03:54-03:59, 03:61-03:68N, 03:70-03:73, 03:76-03:79,

101.704-48/12 – including *Taq* pol., IFU-01
101.704-48u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **6D2**

Lot-specific information

				03:81, 03:83-03:84, 03:86, 03:89, 03:91, 03:93-03:101, 03:104-03:110, 03:112-03:114, 03:116-03:121, 03:123-03:125, 11:01:01-11:01:06, 11:01:08-11:04:12, 11:06:01-11:11:01, 11:11:03-11:16, 11:18-11:21, 11:23:01-11:29:02, 11:31-11:36, 11:38-11:40, 11:42-11:47, 11:49:01-11:49:02, 11:51, 11:53-11:55, 11:57-11:66, 11:68-11:70, 11:72-11:82, 11:84:01-11:87, 11:89-11:103:02, 11:105-11:106, 11:108-11:144, 11:146-11:148, 11:150-11:169N, 11:172-11:182, 11:185-11:188, 11:190-11:191, 11:194, 13:01:01-13:01:08, 13:01:10-13:02:12, 13:04-13:06, 13:09-13:11:02, 13:14:01-13:16, 13:18, 13:20-13:25, 13:27-13:31, 13:34, 13:39, 13:41-13:46, 13:50:01-13:52, 13:54, 13:56-13:57, 13:59, 13:61:01-13:64, 13:66:01-13:66:02, 13:68-13:69, 13:71, 13:73-13:75, 13:77-13:80, 13:82-13:83, 13:86-13:87, 13:91-13:93, 13:96:01-13:100, 13:102, 13:104-13:107, 13:109-13:114, 13:117, 13:119, 13:121, 13:123-13:132, 13:136-13:148, 13:150, 13:153-13:158, 13:162-13:163, 13:165-13:166, 13:168-13:173, 13:176, 13:178-13:179, 13:181, 13:184-13:187, 13:190, 13:192-13:193, 13:196-13:197, 13:199-13:210, 14:17, 14:21, 14:30, 14:33, 14:35, 14:42, 14:53, 14:64-14:65, 14:72, 14:95, 14:128, 14:132, 14:134
41 ⁶	145 bp, 210 bp, 240 bp	430 bp	6, 8, 11, 13, 14, 1403, 1404	*08:09, 08:20-08:21, 08:35, 11:13:01-11:13:02, 11:17, 11:23:01-11:23:02, 11:25, 11:45, 11:52, 11:55, 11:64, 11:89, 11:119, 11:149, 11:159, 13:08, 13:13, 13:18, 13:47, 13:55, 13:119, 13:144, 13:146, 13:154, 13:156, 13:158, 13:164, 14:01:01-14:01:04, 14:02:02 [?] , 14:03:01-14:05:04, 14:07:01-14:08, 14:09 [?] , 14:10-14:12:02, 14:14-14:15, 14:16 [?] , 14:18, 14:19 [?] -14:20 [?] , 14:22 [?] , 14:23:01-14:23:04, 14:25 [?] , 14:26-14:28, 14:29 [?] -14:30 [?] , 14:31-14:32:02, 14:33 [?] , 14:34-14:36, 14:37 [?] , 14:38:01-14:40, 14:41 [?] , 14:42-14:45, 14:47 [?] -14:49 [?] , 14:50, 14:51 [?] -14:53 [?] , 14:54:01-14:65, 14:67-14:68:02, 14:69 [?] , 14:70-14:72, 14:73 [?] -14:74 [?] , 14:75, 14:76 [?] , 14:77-14:78, 14:79 [?] -14:80 [?] , 14:81-14:82, 14:83 [?] , 14:84-14:93, 14:94 [?] , 14:95-14:97, 14:98 [?] , 14:99-14:103, 14:104 [?] -14:109 [?] , 14:110, 14:111 [?] , 14:112-14:118, 14:119 [?] , 14:120, 14:121 [?] , 14:122-14:127:02, 14:128 [?] , 14:129-14:145, 14:146 [?] , 14:147-14:158, 14:159 [?] , 14:160-14:164, 14:165 [?] , 14:166N-14:168
42	140 bp	430 bp	6, 13, 14	*03:76, 13:10, 13:85, 13:120, 13:170, 13:176, 13:178, 13:193, 13:196, 14:02:01-14:02:02, 14:06:01-14:06:03, 14:09, 14:13, 14:17,

101.704-48/12 – including *Taq* pol., IFU-01
101.704-48u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **6D2**

Lot-specific information

				14:19, 14:21, 14:29-14:30, 14:33, 14:46-14:48, 14:51-14:52, 14:80, 14:83, 14:106, 14:108-14:109, 14:121, 14:165
43⁶	160 bp, 240 bp	430 bp	52	*14:141, DRB3*01:01:02:01-01:17, DRB3*02:01-02:29N, DRB3*03:01:01-03:03
44¹⁰	215 bp	430 bp	53	DRB4*01:01:01-01:10
45	175 bp	430 bp	51	DRB5*01:01:01-01:16, DRB5*02:02-02:06
46^{8,12}	-	-	-	Negative Control

¹Alleles are assigned by the presence of specific PCR product(s). However, the sizes of the specific PCR products may be helpful in the interpretation of DR low resolution SSP subtypings. When the primers in a primer mix can give rise to HLA-specific PCR products of more than one length this is indicated if the size difference is more than 20 base pairs. Size differences of 20 base pairs or less are not given. For high resolution SSP kits, the alleles listed are specified according to amplicon length.

Nonspecific amplifications, i.e. a ladder or a smear of bands, may sometimes be seen. GC-rich primers have a higher tendency of giving rise to nonspecific amplifications than other primers.

PCR fragments longer than the control bands may sometimes be observed. Such bands should be disregarded and do not influence the interpretation of the SSP typings.

PCR fragments migrating faster than the control bands, but slower than a 400 bp fragment may be seen in some gel read-outs. Such bands can be disregarded and do not influence the interpretation of the SSP typings.

Some primers may give rise to primer oligomer artifacts. Sometimes this phenomenon is an inherent feature of the primer pair(s) of a primer mix. More often it is due to other factors such as too low amount of DNA in the PCR reactions, taking too long time in setting up the PCR reactions, working at elevated room temperature or using thermal cyclers that are not pre-heated.

²The internal positive control primer pairs amplify segments of the human growth hormone gene. The internal positive control bands are 430 or 515 base pairs respectively, well distribution as outlined in the table. Well number 15 contains the longer, 515 bp, internal positive control band. The well distribution of the internal controls can help in orientation of the kit on gel photo, as well as allow for kit identification. In the presence of a specific amplification the intensity of the control band often decreases. In the presence of a specific amplification the intensity of the control band often decreases.

³The serological split of all DRB1 alleles is not known. In this table we use the expert-assigned serological grouping in Tissue Antigens (2009) 73:95-170 and the serological grouping of the sequence-defined allele.

⁴For several DRB1 alleles 1st and/or 3rd exon(s) and beyond, as well as intron nucleotide sequences, are not available. In these instances it is not known whether some of the primers of the SSP sets are completely matched with the target sequences or not. Assumption is made that unknown sequences in these regions are conserved within allelic groups.

⁵HLA-Specific PCR fragments shorter than 125 base pairs have a lower intensity and are less sharp than longer PCR bands.

⁶Individual alleles can give rise to two differently sized specific PCR fragments in primer mixes 15, 19 to 22, 24, 25, 27, 28, 30, 32 to 34, 36, 41 and 43.

⁷Due to sharing of sequence motifs in codon 38 and 47, DRB3*01:14 will also be amplified in primer mixes 19, 20 and 31 and DRB3*01:23 in mix 19, in addition to primer mix 43.

⁸Primer mixes 15, 17, 30, 33, 36 and 46 may have a tendency of giving rise to primer oligomer formation.

⁹Primer mix 33 has a tendency of giving rise to an intense primer cloud due to the high number of primers present in the primer mix.

¹⁰The DRB4*01:03:01:02N allele is amplified by primer mix 44, whereas the DRB4*02:01N and DRB4*03:01N null alleles are not amplified by this primer pair.

¹¹Primer mixes 17, 18 and 23 may have a tendency of unspecific amplification.

101.704-48/12 – including *Taq* pol., IFU-01
101.704-48u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: 6D2

Lot-specific information

¹²Primer mix 46 contains a negative control, which will amplify more than 95% of HLA amplicons as well as the amplicons generated by the control primer pairs matching the human growth hormone gene. HLA-specific PCR product sizes range from 75 to 200 base pairs and the PCR product generated by the HGH positive control primer pair is 430 base pairs.

‘w’, might be weakly amplified.

‘?’, nucleotide sequence information not available for the primer matching sequence.

101.704-48/12 – including *Taq* pol., IFU-01
101.704-48u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: 6D2

Lot-specific information

DQ LOW PRIMER SPECIFICATION

Well No.	1	2	3	4	5	6	7	8	9	10	11	12
Length of spec.	135	135	210	130	130	135	85	160	225	215	185	185
PCR product	225	185		220	220			205				
		220										
		270										
Length of int.	515	515	430	515	515	515	515	430	430	430	430	430
pos. control ¹												
5'-primer(s) ²	25(170) 5'-gCA 3'	9(122) 5'-gTT 3'	29(184) 5'-gAg 3'	26(173) 5'-TTA 3'	28(179) 5'-gAC 3'	26(173) 5'-TCT 3'	71(309) 5'-ACC 3'	23(164) 5'-gCT 3'	9(122) 5'-gTT 3'	29(184) 5'-gAC 3'	21(159) 5'-ACC 3'	38(210) 5'-gCA 3'
	26(173) 5'-ggA 3'	24(169) 5'-TgT 3'	30(185) 5'-AAg 3'		28(179) 5'-gAC 3'		71(309) 5'-ACC 3'	38(210) 5'-gCg 3'				
	26(173) 5'-ggg 3'	26(173) 5'-TTA 3'			28(179) 5'-gAC 3'							
		26(173) 5'-TCT 3'										
3'-primer(s) ³	57(266) 5'-CAA 3'	57(266) 5'-CAA 3'	86(353) 5'-gCT 3'	55(260) 5'-gCg 3'	57(266) 5'-Cgg 3'	57(266) 5'-CgT 3'	86(353) 5'-gCT 3'	77(327) 5'-ACg 3'	69(304) 5'-CTC 3'	87(356) 5'-ggA 3'	69(304) 5'-CCT 3'	86(353) 5'-gCT 3'
	87(356) 5'-ggT 3'	86(353) 5'-ACg 3'		86(353) 5'-gCT 3'	57(266) 5'-CAg 3'		86(354) 5'-AgT 3'			87(356) 5'-ggT 3'		
	87(356) 5'-ggT 3'	86(353) 5'-ACC 3'		86(354) 5'-AgT 3'	57(266) 5'-Cgg 3'		86(355) 5'-gAC 3'					
	88(361) 5'-CCT 3'	86(354) 5'-TAT 3'			87(356) 5'-ggg 3'		87(356) 5'-ggC 3'					
		86(354) 5'-AAA 3'					87(358) 5'-gCC 3'					
		86(354) 5'-AAg 3'										
		86(354) 5'-AAT 3'										
Well No.	1	2	3	4	5	6	7	8	9	10	11	12

101.704-48/12 – including *Taq* pol., IFU-01
101.704-48u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: 6D2

Lot-specific information

Well No.	13	14
Length of spec. PCR product	185	185
Length of int. pos. control ¹	515	430
5'-primer(s) ²	9(122) 5' -gTT 3'	38(209) 5' -CgC 3'
3'-primer(s) ³	57(266) 5' -CAT 3'	86(353) 5' -ACg 3'
	58(270) 5' -TCC 3'	
Well No.	13	14

¹The internal positive control primer pairs amplify segments of the human growth hormone gene. The internal positive control bands are 430 or 515 base pairs respectively, well distribution as outlined in the table. Well number 1 contains the longer, 515 bp, internal positive control band. The well distribution of the internal controls can help in orientation of the kit on gel photo, as well as allow for kit identification. In the presence of a specific amplification the intensity of the control band often decreases.

²The nucleotide position matching the specificity-determining 3'-end of the primer is given. Nucleotide numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.

³The nucleotide position matching the specificity-determining 3'-end of the primer is given in the anti-sense direction. Nucleotide numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.

101.704-48/12 – including *Taq* pol., IFU-01
101.704-48u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: 6D2

Lot-specific information

DR LOW PRIMER SPECIFICATION

Well No.	15	16	17	18	19	20	21	22	23	24	25	26
Length of spec.	210	200	210	210	125	75	85	100	235	170	85	175
PCR product	230		235		225	205	210	175		215	100	
	260									250	135	
											160	
											195	
Length of int.	515	430	430	430	430	430	430	430	430	515	430	430
pos. control ¹												
5'-primer(s) ²	12(124) 5'-A.T 3'	14(129) 5'-gAA 3'	11(121) 5'-CTg 3'	11(119) 5'-gCT 3'	13(125) 5'-gTC 3'	13(125) 5'-gTC 3'	13(125) 5'-gTC 3'	13(125) 5'-ACA 3'	10(118) 5'-AgA 3'	15(133) 5'-gTT 3'	26(165) 5'-TAT 3'	26(164) 5'-gTA 3'
	14(129) 5'-gAA 3'		13(126) 5'-AgA 3'	13(126) 5'-AAg 3'	47(227) 5'-gTT 3'	15(133) 5'-gTT 3'		13(125) 5'-ACC 3'	13(127) 5'-gTA 3'	15(133) 5'-gTT 3'	36(196) 5'-AgA 3'	30(178) 5'-gCg 3'
			13(126) 5'-AAg 3'	13(126) 5'-Agg 3'				13(125) 5'-ATA 3'	13(127) 5'-ATA 3'	15(133) 5'-gCT 3'	58(261) 5'-gAg 3'	
			13(126) 5'-Agg 3'					13(125) 5'-gTC 3'	13(127) 5'-ATA 3'			
3'-primer(s) ³	66(286) 5'-gAg 3'	66(286) 5'-gAT 3'	66(286) 5'-gAT 3'	66(286) 5'-gAA 3'	73(305) 5'-ggC 3'	26(164) 5'-ggT 3'	28(171) 5'-CTC 3'	32(184) 5'-gTg 3'	77(319) 5'-CgC 3'	58(260) 5'-CCT 3'	57(257) 5'-CgA 3'	72(303) 5'-gCA 3'
	66(286) 5'-gAg 3'		66(286) 5'-gAT 3'	66(286) 5'-gAg 3'	73(305) 5'-ggC 3'	66(286) 5'-gAT 3'	69(295) 5'-CTg 3'	58(260) 5'-Cgg 3'	77(319) 5'-gTA 3'	73(307) 5'-Cag 3'	73(305) 5'-ggC 3'	73(307) 5'-CgC 3'
	66(286) 5'-gAT 3'		69(295) 5'-Tg 3'	70(296) 5'-TgT 3'	73(305) 5'-ggC 3'	71(299) 5'-gCT 3'			77(319) 5'-CAA 3'	86(344) 5'-CAC 3'	77(319) 5'-CAC 3'	
	70(297) 5'-CTg 3'		69(295) 5'-CTg 3'	71(301) 5'-ggC 3'	74(308) 5'-CCC 3'				77(319) 5'-CAC 3'			
	71(299) 5'-gCg 3'		70(298) 5'-CgC 3'		74(310) 5'-CAA 3'							
	77(317) 5'-AgT 3'		71(299) 5'-gCT 3'		77(317) 5'-AgT 3'							
	86(344) 5'-CCA 3'		73(305) 5'-Cgg 3'									
			73(305) 5'-ggC 3'									
			77(317) 5'-AgT 3'									
Well No.	15	16	17	18	19	20	21	22	23	24	25	26

101.704-48/12 – including *Taq* pol., IFU-01
101.704-48u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **6D2**

Lot-specific information

Well No.	27	28	29	30	31	32	33	34	35	36	37	38
Length of spec. PCR product	100	85	220	200	175	100	110	110	170	75	135	180
	170	110		225		150	145	150		175		
						195	170	180		265		
						240		215				
Length of int. pos. control ¹	430	430	430	430	430	430	430	430	430	515	430	430
5'-primer(s) ²	13(125) 5'-gTC 3'	12(124) 5'-Cgg 3'	10(116) 5'-gCT 3'	10(116) 5'-gCT 3'	13(125) 5'-gTC 3'	5(101) 5'-CAA 3'	26(164) 5'-gTA 3'	13(125) 5'-gTC 3'	13(125) 5'-ggg 3'	12(122) 5'-TAG 3'	15(133) 5'-gTT 3'	13(125) 5'-gTC 3'
	15(133) 5'-gTC 3'	15(133) 5'-gTT 3'	12(122) 5'-TAT 3'	12(122) 5'-TAT 3'		37(197) 5'-gTT 3'	34(189) 5'-CAG 3'	34(189) 5'-CAG 3'	15(133) 5'-gTT 3'	74(308) 5'-CCT 3'		107(409) 5'-AgA 3'
	38(200) 5'-CgT 3'		13(125) 5'-gTC 3'	13(125) 5'-gTC 3'		37(197) 5'-gTA 3'		36(196) 5'-AgC 3'	15(133) 5'-gCT 3'			
				13(125) 5'-gTg 3'								
				15(133) 5'-gTT 3'								
				15(133) 5'-gTC 3'								
3'-primer(s) ³	58(260) 5'-CCT 3'	28(171) 5'-CTC 3'	69(295) 5'-gTC 3'	66(286) 5'-gAA 3'	58(260) 5'-Cgg 3'	42(213) 5'-TCA 3'	57(257) 5'-CAG 3'	57(257) 5'-CAG 3'	56(256) 5'-gCT 3'	56(256) 5'-gCT 3'	47(227) 5'-ggA 3'	56(256) 5'-gCT 3'
	58(260) 5'-CCT 3'	29(175) 5'-gTg 3'	71(299) 5'-gCT 3'	70(298) 5'-CgC 3'	58(260) 5'-CAG 3'	57(257) 5'-CAG 3'	69(295) 5'-CTg 3'	59(265) 5'-gTg 3'	57(257) 5'-CAT 3'	86(344) 5'-CCA 3'		159(565) 5'-CAT 3'
	58(260) 5'-CCT 3'	37(199) 5'-CAG 3'	71(299) 5'-ACT 3'	70(298) 5'-CTC 3'		70(298) 5'-CgC 3'	70(296) 5'-TCC 3'	70(296) 5'-TCC 3'	57(257) 5'-CAT 3'			
								73(307) 5'-CAG 3'				
Well No.	27	28	29	30	31	32	33	34	35	36	37	38

Well No.	39	40	41	42	43	44	45
Length of spec. PCR product	150	145	145	140	160	215	175
			210		240		
			240				
Length of int. pos. control ¹	430	430	430	430	430	430	430
5'-primer(s) ²	34(189) 5'-CAG 3'	13(125) 5'-gTC 3'	13(125) 5'-gTC 3'	36(196) 5'-AgA 3'	10(116) 5'-gCT 3'	28(170) 5'-gAT 3'	13(125) 5'-gTA 3'
	34(189) 5'-CAG 3'		37(197) 5'-gTT 3'		10(116) 5'-gCT 3'		
			114(429) 5'-CTg 3'		37(199) 5'-TCC 3'		
3'-primer(s) ³	70(298) 5'-CTC 3'	47(227) 5'-ggA 3'	70(296) 5'-TCC 3'	69(295) 5'-CTg 3'	51(239) 5'-CCC 3'	86(346) 5'-CTC 3'	57(258) 5'-gCg 3'
		47(229) 5'-CCA 3'	73(307) 5'-CAG 3'	70(298) 5'-CTT 3'	77(317) 5'-AAT 3'	86(346) 5'-CTT 3'	58(260) 5'-CCT 3'
			181(630) 5'-CTT 3'				
Well No.	39	40	41	42	43	44	45

¹The internal positive control primer pairs amplify segments of the human growth hormone gene. The internal positive control bands are 430 or 515 base pairs respectively, well distribution as outlined in the table. Well number 15 contains the longer, 515 bp, internal positive control band. The well distribution of the internal controls can help in orientation of the kit on gel photo, as well as allow for kit identification. In the presence of a specific amplification the intensity of the control band often decreases.

²The nucleotide position matching the specificity-determining 3'-end of the primer is given.

101.704-48/12 – including *Taq* pol., IFU-01
101.704-48u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: 6D2

Lot-specific information

Nucleotide numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.

³The nucleotide position matching the specificity-determining 3'-end of the primer is given in the anti-sense direction. Nucleotide numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.

101.704-48/12 – including *Taq* pol., IFU-01

101.704-48u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **6D2**

Lot-specific information

CELL LINE VALIDATION SHEET																			
DQ low resolution primer set ²																			
					Prod. No.:	Well													
						1	2	3	4	5	6	7	8	9	10	11	12	13	14
						201663501	201663502	201663503	201663504	201663505	201663506	201663507	201663508	201663509	201663510	201663511	201663512	201664713	201664714
	IHWC cell line ¹		DQB1																
1	9001	SA	*05:01			+	-	-	-	-	-	-	-	-	+	-	-	-	-
2	9280	LK707	*06:01	*02:02		-	+	+	-	-	-	-	-	-	-	-	-	-	-
3	9011	E4181324	*06:01			-	+	-	-	-	-	-	-	-	-	-	-	-	-
4	9275	GU373	*02:01			-	-	+	-	-	-	-	-	-	-	-	-	-	-
5	9009	KAS011	*05:02			+	-	-	-	-	-	-	-	-	+	-	-	-	-
6	9353	SM	*03:02	*06:01		-	+	-	-	+	-	+	-	-	-	-	+	-	-
7	9020	QBL	*02:01			-	-	+	-	-	-	-	-	-	-	-	-	-	-
8	9025	DEU	*03:01			-	-	-	+	-	-	+	-	-	-	-	+	-	-
9	9026	YAR	*03:02			-	-	-	-	+	-	+	-	-	-	-	+	-	-
10	9107	LKT3	*04:01			-	-	-	-	-	-	-	+	+	-	-	-	-	-
11	9051	PITOUT	*02:02			-	-	+	-	-	-	-	-	-	-	-	-	-	-
12	9052	DBB	*03:03			-	-	-	-	-	+	+	-	-	-	-	+	-	-
13	9004	JESTHOM	*05:01			+	-	-	-	-	-	-	-	-	+	-	-	-	-
14	9071	OLGA	*04:02			-	-	-	-	-	-	-	+	+	-	-	-	-	-
15	9075	DKB	*03:03			-	-	-	-	-	+	+	-	-	-	-	+	-	-
16	9037	SWEIG007	*03:01			-	-	-	+	-	-	+	-	-	-	-	+	-	-
17	9282	CTM 3953540	*02:01	*06:03		-	+	+	-	-	-	-	-	-	-	-	-	-	+
18	9257	32367	*06:02	*02:02		-	+	+	-	-	-	-	-	-	-	-	-	+	+
19	9038	BM16	*03:01			-	-	-	+	-	-	+	-	-	-	-	+	-	-
20	9059	SLE005	*06:04			-	+	-	-	-	-	-	-	-	-	-	-	-	-
21	9064	AMALA	*03:01			-	-	-	+	-	-	+	-	-	-	-	+	-	-
22	9056	KOSE	*05:03	*06:04		+	+	-	-	-	-	-	-	-	+	-	-	-	-
23	9124	IHL	*05:03	*06:01		+	+	-	-	-	-	-	-	-	+	-	-	-	-
24	9035	JBUSH	*03:01			-	-	-	+	-	-	+	-	-	-	-	+	-	-
25	9049	IBW9	*02:02			-	-	+	-	-	-	-	-	-	-	-	-	-	-
26	9285	WT49	*02:01			-	-	+	-	-	-	-	-	-	-	-	-	-	-
27	9191	CH1007	*04:01	*05:01		+	-	-	-	-	-	-	+	+	+	-	-	-	-
28	9320	BEL5GB	*02:02	*03:01		-	-	+	+	-	-	+	-	-	-	-	+	-	-
29	9050	MOU	*02:02			-	-	+	-	-	-	-	-	-	-	-	-	-	-
30	9021	RSH	*04:02			-	-	-	-	-	-	-	+	+	-	-	-	-	-
31	9019	DUCAF	*02:01			-	-	+	-	-	-	-	-	-	-	-	-	-	-
32	9297	HAG	*03:01			-	-	-	+	-	-	+	-	-	-	-	+	-	-
33	9098	MT14B	*03:02			-	-	-	-	+	-	+	-	-	-	-	+	-	-
34	9104	DHIF	*03:01			-	-	-	+	-	-	+	-	-	-	-	+	-	-
35	9302	SSTO	*03:05			-	-	-	-	+	-	+	-	-	-	+	-	-	-
36	9024	KT17	*03:02			-	-	-	-	+	-	+	-	-	-	-	+	-	-
37	9065	HHKB	*06:03			-	+	-	-	-	-	-	-	-	-	-	-	-	+
38	9099	LZL	*03:01			-	-	-	+	-	-	+	-	-	-	-	+	-	-
39	9315	CML	*02:01	*03:01		-	-	+	+	-	-	+	-	-	-	-	+	-	-
40	9134	WHONP199	*02:02	*03:03		-	-	+	-	-	+	+	-	-	-	-	+	-	-
41	9055	H0301	*06:09			-	+	-	-	-	-	-	-	-	-	-	-	-	-
42	9066	TAB089	*06:01			-	+	-	-	-	-	-	-	-	-	-	-	-	-
43	9076	T7526	*03:03			-	-	-	-	-	+	+	-	-	-	-	+	-	-
44	9057	TEM	*05:03			+	-	-	-	-	-	-	-	-	+	-	-	-	-
45	9239	SHJO	*02:02			-	-	+	-	-	-	-	-	-	-	-	-	-	-
46	9013	SCHU	*06:02			-	+	-	-	-	-	-	-	-	-	-	-	+	+
47	9045	TUBO	*03:01			-	-	-	+	-	-	+	-	-	-	-	+	-	-
48	9303	TER-ND	*05:01			+	-	-	-	-	-	-	-	-	+	-	-	-	-

101.704-48/12 – including *Taq* pol., IFU-01

101.704-48u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **6D2**

Lot-specific information

CELL LINE VALIDATION SHEET																															
DR low resolution primer set ²																															
					Prod. No.:	Well																									
						15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30										
						201553601	201553602	201664603	201663304	201663305	201553606	201553607	201553608	201663309	201553610	201554911	201553612	201553613	201557414	201553615	201553616										
	IHCW cell line ¹		DRB1																												
1	9001	SA	*01:01		+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
2	9280	LK707	*15:02	*04:05	-	-	+	-	-	-	-	-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
3	9011	E4181324	*15:02		-	-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
4	9275	GU373	*03:01		-	-	-	-	+	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
5	9009	KAS011	*16:01		-	-	-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
6	9353	SM	*04:07	*08:03	-	-	-	-	-	-	-	-	+	-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
7	9020	QBL	*03:01		-	-	-	-	+	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
8	9025	DEU	*04:01		-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
9	9026	YAR	*04:02		-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
10	9107	LKT3	*04:05		-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
11	9051	PITOUT	*07:01		-	-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
12	9052	DBB	*07:01		-	-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
13	9004	JESTHOM	*01:01		+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
14	9071	OLGA	*08:02		-	-	-	-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	+
15	9075	DKB	*09:01		-	-	-	-	-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-
16	9037	SWEIG007	*11:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	+	-	+	+	-	-	-	-	-	-	-	-	-
17	9282	CTM3953540	*03:01	*13:01	-	-	-	-	+	+	-	-	-	-	-	-	-	-	-	-	-	+	+	-	-	-	-	-	-	-	-
18	9257	32367	*09:01	*11:01	-	-	-	-	-	-	-	-	-	-	-	-	+	-	+	-	+	-	+	+	-	-	-	-	-	-	-
19	9038	BM16	*12:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-	-	-	-	-
20	9059	SLE005	*13:02		-	-	-	-	-	-	+	-	-	-	-	-	-	-	-	-	+	+	-	-	-	-	-	-	-	-	-
21	9064	AMALA	*14:02		-	-	-	-	-	-	-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
22	9056	KOSE	*13:02	*14:54	-	-	-	-	-	+	-	-	-	-	-	-	-	-	-	-	-	+	+	-	-	-	-	-	-	-	-
23	9124	IHL	*08:03	*14:14	-	-	-	-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
24	9035	JBUSH	*11:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	+	-	+	+	-	-	-	-	-	-	-	-	-
25	9049	IBW9	*07:01		-	-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
26	9285	WT49	*03:01		-	-	-	-	+	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
27	9191	CH1007	*04:05	*10:01	-	-	-	-	-	-	-	-	+	-	-	-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-
28	9320	BEL5GB	*04:16	*07:01	-	-	-	-	-	-	-	-	+	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
29	9050	MOU	*07:01		-	-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
30	9021	RSH	*03:02		-	-	-	-	+	-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
31	9019	DUCAF	*03:01		-	-	-	-	+	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
32	9297	HAG	*13:03		-	-	-	-	-	-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
33	9098	MT14B	*04:04		-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
34	9104	DHIF	*11:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	+	-	+	+	-	-	-	-	-	-	-	-	-
35	9302	SSTO	*04:03		-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
36	9024	KT17	*04:03	*04:06	-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
37	9065	HHKB	*13:01		-	-	-	-	-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
38	9099	LZL	*14:02		-	-	-	-	-	-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
39	9315	CML	*03:01	*04:01	-	-	-	-	+	+	-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
40	9134	WHONP199	*07:01	*09:01	-	-	-	-	-	-	-	-	-	+	-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
41	9055	H0301	*13:02		-	-	-	-	-	-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
42	9066	TAB089	*08:03		-	-	-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
43	9076	T7526	*09:01		-	-	-	-	-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-
44	9057	TEM	*14:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
45	9239	SHJO	*07:01		-	-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
46	9013	SCHU	*15:01		-	-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
47	9045	TUBO	*11:04	*12:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	+	+	+	+	+	+	+	+	+	+	+	+	+
48	9303	TER-ND	*01:03		-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-

101.704-48/12 – including *Taq* pol., IFU-01
101.704-48u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **6D2**

Lot-specific information

CELL LINE VALIDATION SHEET																				
DR low resolution primer set ¹																				
					Prod. No.:	Well														
						31	32	33	34	35	36	37	38	39	40	41	42	43	44	45
						201663317	201553618	201553619	201553620	201557421	201663322	201557423	201557424	201557425	201557426	201557427	201557428	201553629	201553630	201553631
	IHCW cell line		DRB1																	
1	9001	SA	*01:01			-	-	-	-	-	-	-	-	-	-	-	-			
2	9280	LK707	*15:02	*04:05		-	-	-	-	-	-	-	-	-	-	+	+			
3	9011	E4181324	*15:02			-	-	-	-	-	-	-	-	-	-	-	+			
4	9275	GU373	*03:01			+	-	-	-	-	-	-	-	-	+	-	-			
5	9009	KAS011	*16:01			-	-	-	-	-	-	-	-	-	-	-	+			
6	9353	SM	*04:07	*08:03		-	-	-	-	+	-	-	-	-	-	+	-			
7	9020	QBL	*03:01			+	-	-	-	-	-	-	-	-	+	-	-			
8	9025	DEU	*04:01			-	-	-	-	-	-	-	-	-	-	+	-			
9	9026	YAR	*04:02			-	-	-	-	-	-	-	-	-	-	+	-			
10	9107	LKT3	*04:05			-	-	-	-	-	-	-	-	-	-	+	-			
11	9051	PITOUT	*07:01			-	-	-	-	-	-	-	-	-	-	+	-			
12	9052	DBB	*07:01			-	-	-	-	-	-	-	-	-	-	+	-			
13	9004	JESTHOM	*01:01			-	-	-	-	-	-	-	-	-	-	-	-			
14	9071	OLGA	*08:02			-	-	-	-	+	-	-	-	-	-	-	-			
15	9075	DKB	*09:01			-	-	+	-	-	-	-	-	-	-	+	-			
16	9037	SWEIG007	*11:01			-	-	-	-	-	-	-	-	-	+	-	-			
17	9282	CTM3953540	*03:01	*13:01		+	-	-	-	-	-	-	-	+	+	-	-			
18	9257	32367	*09:01	*11:01		-	-	+	-	-	-	-	-	-	+	+	-			
19	9038	BM16	*12:01			-	-	-	-	-	-	+	-	-	+	-	-			
20	9059	SLE005	*13:02			+	-	-	-	-	-	-	-	+	+	-	-			
21	9064	AMALA	*14:02			+	-	+	-	-	-	-	-	-	-	+	-			
22	9056	KOSE	*13:02	*14:54		+	+	+	+	-	-	-	-	+	+	+	-			
23	9124	IHL	*08:03	*14:14		+	-	+	+	+	-	-	-	-	+	-	-			
24	9035	JBUSH	*11:01			-	-	-	-	-	-	-	-	-	+	-	-			
25	9049	IBW9	*07:01			-	-	-	-	-	-	-	-	-	-	+	-			
26	9285	WT49	*03:01			+	-	-	-	-	-	-	-	-	+	-	-			
27	9191	CH1007	*04:05	*10:01		-	-	-	-	-	-	-	-	-	-	+	-			
28	9320	BEL5GB	*04:16	*07:01		-	-	-	-	-	-	-	-	-	-	-	+			
29	9050	MOU	*07:01			-	-	-	-	-	-	-	-	-	-	+	-			
30	9021	RSH	*03:02			+	-	-	-	-	-	-	-	-	-	+	-			
31	9019	DUCAF	*03:01			+	-	-	-	-	-	-	-	-	+	-	-			
32	9297	HAG	*13:03			+	-	-	-	-	-	+	-	-	-	+	-			
33	9098	MT14B	*04:04			-	-	-	-	-	-	-	-	-	-	+	-			
34	9104	DHIF	*11:01			-	-	-	-	-	-	-	-	+	-	+	-			
35	9302	SSTO	*04:03			-	-	-	-	-	-	-	-	-	-	-	+			
36	9024	KT17	*04:03	*04:06		-	-	-	-	-	-	-	-	-	-	-	+			
37	9065	HHKB	*13:01			+	-	-	-	-	-	-	-	+	+	-	-			
38	9099	LZL	*14:02			+	-	+	-	-	-	-	-	-	-	+	-			
39	9315	CML	*03:01	*04:01		+	-	-	-	-	-	-	-	-	+	+	+			
40	9134	WHONP199	*07:01	*09:01		-	-	+	-	-	-	-	-	-	-	-	+			
41	9055	H0301	*13:02			+	-	-	-	-	-	-	-	+	+	-	-			
42	9066	TAB089	*08:03			-	-	-	-	+	-	-	-	-	-	-	-			
43	9076	T7526	*09:01			-	-	+	-	-	-	-	-	-	-	-	+			
44	9057	TEM	*14:01			-	+	+	+	-	-	-	+	-	-	+	-			
45	9239	SHJO	*07:01			-	-	-	-	-	-	-	-	-	-	-	+			
46	9013	SCHU	*15:01			-	-	-	-	-	-	-	-	-	-	-	+			
47	9045	TUBO	*11:04	*12:01		-	-	-	-	-	+	-	-	+	-	-	-			
48	9303	TER-ND	*01:03			-	-	-	-	-	-	-	-	-	-	-	-			

101.704-48/12 – including *Taq* pol., IFU-01
101.704-48u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: 6D2

Lot-specific information

¹The provided cell line HLA specificities are retrieved from the <http://www.ihwg.org/hla> web site.
 The specificity of an individual cell line may thus be subject to change.

²The specificity of each primer solution in the kit has been tested against 48 well characterized cell line DNAs and where applicable, additional cell line DNAs.

One 5'-primer and one or more 3'-primers in primer solution 2, 18, 25, 32 and 41 were tested by separately adding additional 5'-primers or 3'-primers.

One or more additional 3'-primers in primer solution 1, 10, 15, 17, 19, 23, 24, 34 and 36 were tested by separately adding another 5'-primer.

One 5'-primer in primer solutions 7, 20, 26, 29 and 30 was tested by separately adding additional 3'-primers.

In primer solutions 1, 2, 4, 5, 7, 13, 15, 17 to 20, 23, 27, 29, 35, 40 and 44 one or more 3'-primers were not possible to test, and in primer solutions 1 to 3, 5, 15, 17, 18, 22 to 24, 27 to 30, 34 to 36 and 41 one or more 5'-primers were not possible to test.

101.704-48/12 – including *Taq* pol., IFU-01
101.704-48u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **6D2**

Lot-specific information

101.704-48/12 – including *Taq* pol., IFU-01
101.704-48u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **6D2**

Lot-specific information

101.704-48/12 – including *Taq* pol., IFU-01
 101.704-48u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: 6D2

Lot-specific information

ADDRESSES:

Manufacturer:

Olerup SSP AB, Franzengatan 5, SE-112 51 Stockholm, Sweden.

Tel: +46-8-717 88 27

Fax: +46-8-717 88 18

E-mail: olerup-se@caredx.com

Web page: <http://www.olerup.com>

Distributed by:

Olerup GmbH, Löwengasse 47 / 6, AT-1030 Vienna, Austria.

Tel: +43-1-710 15 00

Fax: +43-1-710 15 00 10

E-mail: olerup-at@caredx.com

Web page: <http://www.olerup.com>

Olerup Inc., 901 S. Bolmar St., Suite R, West Chester, PA 19382

Tel: 1-877-OLERUP1

Fax: 610-344-7989

E-mail: olerup-us@caredx.com

Web page: <http://www.olerup.com>

For information on *Olerup* distributors worldwide, contact **Olerup GmbH**.