

101.101.48/12 – including *Taq* pol., IFU-01
101.101.48u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **91Y**

Lot-specific information

Olerup SSP® DR low resolution

Product number:	101.101-48/12 – including <i>Taq</i> pol. 101.101-48u/12u – without <i>Taq</i> pol.
Lot number:	91Y
Expiry date:	2018-April-01
Number of tests:	48 tests – Product No. 101.101-48/48u 12 tests – Product No. 101.101-12/12u
Number of wells per test:	31 + 1
Storage - pre-aliquoted primers:	dark at -20°C
- PCR Master Mix:	-20°C
- Adhesive PCR seals	RT
- Product Insert	RT

This Product Description is only valid for Lot No. 91Y.

Complete product documentation consists of generic Instructions for Use (IFU), lot specific Product Insert, Worksheet and Certificate.

CHANGES COMPARED TO THE PREVIOUS OLERUP SSP® DR LOW RESOLUTION LOT (09Y)

A well containing Negative Control primer pairs has been added.

The format of the Product Insert and Worksheet have been changed.

The DR low resolution specificity and interpretation tables have been updated for the HLA-DRB1 alleles described since the previous *Olerup SSP®* DR low resolution lot was made (**Lot No. 09Y**). The kit design is based on IMGT/HLA database 3.21.1.

As of lot series V, the Specificity Table is included in the lot-specific Product Insert, and the Interpretation Table is included in the Worksheet.

101.101.48/12 – including *Taq* pol., IFU-01
101.101.48u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **91Y**

Lot-specific information

The primers of the wells detailed below have been exchanged, added or modified compared to the previous lot.

Well	5'-primer	3'-primer	rationale
3	Added	Added	5'-primer added for the DRB1*15:01:22 allele, 3'-primer added for the DRB1*15:115N allele.
4	-	Exchanged	3'-primer exchanged for improved HLA-specific amplification.
6	-	Added	3'-primer added for the DRB1*13:193 allele.
9	-	Added	3'-primer added for the DRB*07:61 allele.
11	Added	-	5'-primer added for the DRB1*09:23 allele.
12	-	Added	3'-primer added for the DRB1*10:11 allele.
14	-	Added	3'-primer added for the DRB1*12:58 allele.
16	Added	-	5'-primer added for the DRB1*13:183 allele.
21	Added	-	5'-primers added for the DRB1*08:01:06 and DRB1*08:57 alleles.
22	-	-	Exchanged positive control primer pair for decreased tendency of primer oligomer formation.

Change in revision R01 compared to R00:

1. Primer mix 5 does not amplify the DRB1*03:11:01, *03:97, *11:53 and *15:100 alleles. This has been corrected in the Specificity and Interpretation Tables.

Change in revision R02 compared to R01:

1. Primer mix 12 has pronounced tendencies of primer oligomer formation. A foot note has been added in the Specificity Table.

Change in revision R03 compared to R02:

1. Primer mix 21 does not amplify the DRB1*08:14 allele. This has been corrected in the Specificity and Interpretation Tables.

Changes in revision R04 compared to R03:

1. Due to sharing of sequence motifs in codon 38 and 47, DRB3*01:23 will be amplified in primer mix 5, in addition to primer mix 21.

2. The DRB1*13:02:02 allele is amplified in primer mix 6.

The Specificity and Interpretation Tables have been changed.

Change in revision R05 compared to R04:

1. Primer mix 6 does not amplify the DRB1*14:137N and 14:152N alleles. Primer mix 15 does not amplify the DRB1*14:137N allele. This has been corrected in the Specificity and Interpretation Tables.

101.101.48/12 – including *Taq* pol., IFU-01
101.101.48u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **91Y**

Lot-specific information

Well **32** contains Negative Control primer pairs, that will amplify more than 95% of the *Olerup SSP®* HLA Class I, DRB, DQB1, DPB1 and DQA1 amplicons as well as all the amplicons generated by the control primer pairs matching the human growth hormone gene.

HLA-specific PCR product sizes range from 75 to 200 base pairs.
The PCR product generated by the positive control primer pair is 430 base pairs.

Length of PCR product	105	200	105	80	75	80	85
5'-primer¹	164	340	440	45	45	43	36
	5'-CAC ^{3'}	5'-Agg ^{3'}	5'-TTA ^{3'}	5'-Tgg ^{3'}	5'-Tgg ^{3'}	5'-Tgg ^{3'}	5'-TAC ^{3'}
							36
							5'-TAT ^{3'}
3'-primer²	231	2nd I	507	59	58	57	47
	5'-TgC ^{3'}	5'-AAA ^{3'}	5'-TTg ^{3'}	5'-CTC ^{3'}	5'-ggC ^{3'}	5'-CTC ^{3'}	5'-ACA ^{3'}
							48
							5'-gCA ^{3'}
							48
							5'-gCC ^{3'}
							52
							5'-TgT ^{3'}
A*	+	+	+				
B*	+	+	+				
C*	+	+	+				
DRB1				+	+		
DRB3				+	+		
DRB5				+			
DQB1					+		
DPB1						+	
DQA1							+

¹The nucleotide position for HLA class I genes and the codon for HLA class II genes, in the 2nd or 3rd exon, matching the specificity-determining 3'-end of the primer is given. Nucleotide and codon numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.

²The nucleotide position for HLA class I genes and the codon for HLA class II genes, in the 2nd or 3rd exon or the 2nd intron, matching the specificity-determining 3'-end of the primer is given in the anti-sense direction. Nucleotide and codon numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.

101.101.48/12 – including *Taq* pol., IFU-01
101.101.48u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **91Y**

Lot-specific information

PRODUCT DESCRIPTION

DR low resolution

CONTENT

The primer set contains 5'- and 3'-primers for grouping the DRB1*01:01 to DRB1*10:15 alleles into the corresponding serological groups DR1 to DR18 as well as primer pairs for recognizing the DRB3, DRB4 and DRB5 groups of alleles.

PLATE LAYOUT

Each test consists of 32 PCR reactions in a 32 well cut PCR plate.

1	2	3	4	5	6	7	8
9	10	11	12	13	14	15	16
17	18	19	20	21	22	23	24
25	26	27	28	29	30	31	NC

The 32 well cut PCR plate is marked with ‘DR low’ in silver/gray ink.

Well No. 1 is marked with the Lot No. ‘91Y’.

Wells 1 to 31 – DR low resolution primers.

Well 32 – Negative Control.

A faint row of numbers is seen between wells 1 and 2 or wells 7 and 8 of the PCR trays. These stem from the manufacture of the trays, and should be disregarded.

The PCR plates are covered with a PCR-compatible foil.

Please note: When removing each 32 well PCR plate, make sure that the remaining plates stay covered. Use a scalpel or a similar instrument to carefully cut the foil between the plates.

INTERPRETATION

Only HLA-DRB alleles will be amplified by the 31 wells of the DR low resolution primer set, **wells 1 to 31**. Thus, the interpretation of DR low resolution typings is not influenced by other HLA class II genes.

UNIQUELY IDENTIFIED ALLELES

All the HLA-DRB1, -DRB3, -DRB4¹ and -DRB5 alleles, i.e. **DRB1*01:01 to 10:15, DRB3*01:01:02:01 to DRB3*03:03, DRB4*01:01:01:01 to DRB4*01:09, DRB5*01:01:01 to DRB5*02:06**, recognized by the HLA Nomenclature Committee in August 2015^{2,3} will be amplified by the primers in the DR low resolution SSP kit⁴. The HLA-DRB alleles will be grouped into their corresponding serological specificities.

¹The DRB4*02:01N and DRB4*03:01N null alleles will not be amplified by the DR low resolution primer set.

²DRB alleles listed on the IMGT/HLA web page 2015-August-11, release 3.21.1, www.ebi.ac.uk/imgt/hla.

³Alleles that have been deleted from or renamed in the official WHO HLA Nomenclature up to and including the last IMGT/HLA database release can be retrieved from web page <http://hla.alleles.org/alleles/deleted.html>.

⁴The DRB1*03:76 and the DRB1*13:176 alleles give rise to identical amplification patterns with the DR low resolution primer set. These alleles can be separated by the respective high resolution primer sets.

101.101.48/12 – including *Taq* pol., IFU-01
101.101.48u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **91Y**

Lot-specific information
SPECIFICITY TABLE

DR low resolution primer set

Specificities and sizes of the PCR products of the 31+1 primer mixes of the DR low resolution primer set

Primer Mix	Size of spec. PCR product ¹	Size of control band ²	DR serology ³	Amplified HLA-DRB alleles ^{4,6}
1^{6,7}	210 bp, 235 bp, 260 bp	515 bp	1	*01:01:01-01:02:11, 01:04-01:38, 01:40N-01:71
2	200 bp	430 bp	1/103	*01:03, 01:39N, 01:42, 01:61
3^{6,7,9}	210 bp, 230 bp	430 bp	2, 15	*15:01:01-01:15:120, 15:122-15:124
4⁹	210 bp	430 bp	16	*16:01:01-16:05:02, 16:07-16:36
5^{5,6,7,12}	120 bp, 220 bp	430 bp	3, 11, 17, 18	*03:01:01-01:03:10, 03:12-03:75, 03:77-03:96, 03:98-03:118, 03:120-03:121, 11:07, 11:103-01-11:103:02, 11:105, 11:107, 11:125, 11:173, 15:25
6^{5,6,12}	75 bp, 210 bp	430 bp	3, 6, 11, 13, 14, 17	*03:01:01-01:03:01:23, 03:01:25, 03:04:01-03:06, 03:08-03:16, 03:18-03:20, 03:22-03:23, 03:25-03:26, 03:28, 03:30-03:31, 03:33-03:34, 03:36-03:37, 03:43-03:48, 03:50-03:52, 03:54-03:68N, 03:70-03:73, 03:75-03:86, 03:89, 03:91-03:93, 03:95-03:96, 03:98-03:100:02, 03:104, 03:106-03:110, 03:112-03:114, 03:116-03:118, 03:121, 08:40, 11:02:01-11:03:02, 11:11:01, 11:11:03, 11:14:01-11:14:02, 11:16, 11:20-11:21, 11:36, 11:40-11:41, 11:48, 11:59, 11:63:01-11:63:02, 11:65:01-11:65:02, 11:68, 11:70, 11:73, 11:76, 11:79-11:80, 11:83, 11:85-11:87, 11:93, 11:118, 11:122, 11:124, 11:127, 11:131-11:132, 11:135, 11:138-11:139, 11:142, 11:151, 11:153, 11:161, 11:168, 11:171, 11:176, 11:182, 11:184, 13:01:01-13:04, 13:08, 13:10, 13:15-13:17, 13:19-13:20, 13:22-13:24, 13:27-13:29, 13:31-13:41, 13:43, 13:45, 13:48, 13:51-13:54, 13:57, 13:59, 13:61:01-13:61:02, 13:63-13:66:02, 13:68-13:76, 13:78-13:81, 13:83-13:85, 13:87-13:99, 13:101-13:102, 13:104-13:107, 13:109, 13:111-13:117, 13:120-13:131, 13:133, 13:135, 13:137N-13:145, 13:147-13:149, 13:151-13:153, 13:155, 13:159, 13:162, 13:165-13:168, 13:170-13:180, 13:182, 13:184-13:188, 13:190-13:191, 13:193-13:194, 13:196, 13:198, 13:200N-13:202, 13:204, 14:16, 14:19, 14:21, 14:45, 14:82, 14:95, 14:109, 14:120, 14:122, 14:132
7^{5,6}	85 bp, 210 bp	430 bp	3, 6, 11, 13, 14, 1403, 18	*03:02:01-03:03, 03:27, 03:29, 03:38, 03:53, 03:74, 03:88, 03:90, 03:102-03:103, 03:115, 11:13:01 ^w -11:13:02 ^w , 11:26, 11:34, 13:15, 13:19, 13:26:01-13:26:02, 13:44, 13:53, 13:57, 13:85-13:86, 13:104, 13:193, 13:198, 14:02:01-14:03:02, 14:06:01-14:06:03, 14:09, 14:12:01-14:13, 14:17-14:21, 14:24, 14:27, 14:29-14:30, 14:32:01 ^w -14:32:02 ^w , 14:33, 14:40-14:41, 14:47-14:49, 14:51, 14:63, 14:65 ^w , 14:67, 14:77-14:78, 14:80-14:81, 14:83, 14:85, 14:89, 14:94, 14:98, 14:102, 14:106, 14:108-14:109, 14:115, 14:119, 14:121, 14:135, 14:146, 14:154, 14:156, 14:159-14:160

101.101.48/12 – including *Taq* pol., IFU-01
101.101.48u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **91Y**

Lot-specific information

8^{5,6,7}	100 bp, 175 bp	430 bp	3, 4	*04:01:01-04:05:11, 04:05:13-04:194
9⁹	235 bp	430 bp	7	*07:01:01:01-07:01:18, 07:03-07:61
10⁶	170 bp, 215 bp, 250 bp	515 bp	8, 11, 12, 14	*08:01:01-08:19, 08:21-08:56, 08:58-08:66, 08:68-08:76, 11:67, 12:04, 12:16:01-12:16:03, 12:22, 12:39, 12:49, 14:11, 14:15, 14:68:01-14:68:02, 14:93, 14:148
11^{5,6,7}	90 bp, 135 bp, 165 bp, 190 bp	430 bp	3, 9, 11	*03:08, 03:65, 03:112, 09:01:02-09:24, 09:26, 11:07, 11:103:01- 11:103:02, 11:105, 11:107, 11:125, 11:173
12⁷	175 bp	430 bp	10, 11, 13	*03:76, 10:01:01-10:15, 11:59, 11:80, 11:83, 11:87, 11:135, 11:142, 11:182, 13:27, 13:41, 13:71, 13:129, 13:176
13^{5,6,7}	100 bp, 170 bp	430 bp	3, 8, 11, 13, 14	*03:08, 03:65, 08:31, 08:41, 08:63, 08:75, 11:01:01-11:11:01, 11:11:03-11:70, 11:72-11:176, 11:178-11:184
14^{5,6}	90 bp, 110 bp	430 bp	12, 14	*08:32, 08:53, 12:01:01-12:56, 12:58, 14:52
15¹¹	220 bp	430 bp	6, 8, 11, 13, 14, 1403	*03:76, 08:20-08:21, 11:01:01-11:04:12, 11:06:01-11:06:03, 11:08:01-11:11:01, 11:11:03-11:12:02, 11:14:01-11:16, 11:18- 11:21, 11:23:01-11:25, 11:27:01-11:33, 11:35-11:51, 11:54:01- 11:54:02, 11:56-11:66, 11:68, 11:70, 11:72-11:81, 11:83-11:88, 11:90-11:97, 11:99-11:102:02, 11:106, 11:108-11:124, 11:126- 11:135, 11:137-11:142, 11:144-11:158, 11:160-11:172, 11:174- 11:176, 11:178-11:184, 12:57, 13:01:01-13:02:01, 13:02:03- 13:08, 13:10-13:16, 13:18-13:43, 13:45-13:85, 13:87-13:115, 13:117-13:128, 13:130-13:145, 13:147-13:166, 13:168-13:174, 13:176-13:182, 13:184-13:191, 13:194-13:204, 14:03:01- 14:03:02, 14:12:01-14:12:02, 14:16, 14:19, 14:21-14:22, 14:25, 14:27, 14:40, 14:53, 14:63, 14:67, 14:69, 14:74, 14:77-14:78, 14:84-14:85, 14:98, 14:102, 14:105, 14:109, 14:115-14:116, 14:128, 14:135, 14:144, 14:156, DRB3*02:27
16^{6,7}	205 bp, 225 bp	430 bp	6, 8, 11, 12, 13, 14	*08:01:01-08:01:05, 08:02:01-08:02:04, 08:04:01-08:09, 08:11, 08:16-08:17, 08:20-08:22, 08:24, 08:26, 08:28, 08:31, 08:39, 08:41-08:44, 08:50, 08:52, 08:54-08:55, 08:57, 08:59, 08:64, 08:67, 08:70, 08:72-08:73, 08:75, 11:01:01-11:01:17, 11:01:20- 11:06:03, 11:09-11:11:01, 11:11:03-11:12:02, 11:14:01-11:16, 11:20-11:21, 11:23:01-11:25, 11:27:01-11:30, 11:32-11:33, 11:35-11:41, 11:43-11:44, 11:46:01-11:51, 11:54:01-11:56, 11:58:01-11:63:02, 11:65:01-11:70, 11:72, 11:74:01-11:78, 11:80-11:88, 11:90-11:97, 11:99-11:102:02, 11:106, 11:108- 11:118, 11:120-11:124, 11:126-11:129, 11:133-11:135, 11:137- 11:142, 11:144-11:152, 11:154-11:158, 11:160-11:172, 11:174- 11:176, 11:178-11:183, 12:02:01-12:02:06, 12:13, 12:15- 12:16:03, 12:18-12:21, 12:23, 12:26-12:27, 12:31N-12:33, 12:37, 12:42-12:45, 12:49-12:52, 12:55-12:57, 13:01:01-13:02:01, 13:02:03-13:02:12, 13:04-13:05:03, 13:07:01-13:09, 13:11:01- 13:11:02, 13:14:01-13:24, 13:26:01-13:29, 13:31-13:32, 13:34- 13:36, 13:38-13:43, 13:45-13:55, 13:57, 13:59, 13:61:01-13:65, 13:67-13:76, 13:78-13:80, 13:83-13:84, 13:87, 13:91-13:93, 13:96:01-13:100, 13:102-13:109, 13:111-13:114, 13:116-13:117, 13:121, 13:123-13:132, 13:135-13:136, 13:138-13:150, 13:153, 13:155, 13:158-13:160, 13:162, 13:164-13:166, 13:168-13:169, 13:171:01-13:171:02, 13:173, 13:175, 13:177, 13:179, 13:182- 13:187, 13:189-13:192, 13:195, 13:197-13:204, 14:15-14:16, 14:22, 14:24-14:25, 14:27, 14:37, 14:53, 14:73, 14:105, 14:128

101.101.48/12 – including *Taq* pol., IFU-01
101.101.48u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **91Y**

Lot-specific information

17¹²	175 bp	430 bp	3, 6, 11, 13, 14, 1403, 17, 18	*03:01:01:01-03:07, 03:09, 03:11:01-03:41:02, 03:43-03:45, 03:47-03:63, 03:66-03:68N, 03:70-03:86, 03:88-03:91, 03:93- 03:110, 03:112-03:118, 03:120-03:121, 08:20, 13:01:01-13:16, 13:18-13:42, 13:44, 13:46-13:66:02, 13:68-13:102, 13:104- 13:115, 13:117-13:121, 13:123-13:158, 13:161-13:164, 13:166- 13:170, 13:171:02-13:174, 13:176-13:178, 13:180-13:182, 13:184-13:190, 13:192-13:194, 13:196, 13:198-13:204, 14:02:01- 14:03:02, 14:05:01-14:06:03, 14:09, 14:12:01-14:14, 14:17- 14:21, 14:23:01, 14:23:03-14:24, 14:27, 14:29-14:30, 14:33, 14:36-14:37, 14:40-14:45, 14:47-14:48, 14:51, 14:56, 14:59, 14:63-14:65, 14:67, 14:77-14:78, 14:80-14:81, 14:83-14:85, 14:89, 14:91, 14:94-14:96, 14:98, 14:100, 14:102-14:103, 14:106, 14:108-14:109, 14:115-14:116, 14:121, 14:123, 14:127:01- 14:127:02, 14:134-14:136, 14:144, 14:154-14:156, 14:159- 14:160
18^{5,6,7,10}	100 bp, 150 bp, 195 bp, 240 bp	430 bp	4, 6, 8,13, 14, 1404	*04:62, 04:69, 04:73, 04:105:01-04:105:02, 04:122, 04:146, 08:08, 11:69, 11:82, 13:45, 13:197, 14:01:01-14:01:02, 14:01:04, 14:04:01-14:04:02, 14:07:01-14:07:02, 14:10, 14:16, 14:22, 14:25-14:26, 14:28, 14:31-14:32:03, 14:35, 14:37-14:39, 14:49- 14:50, 14:53-14:54:01, 14:54:03-14:55, 14:57-14:58, 14:60- 14:62, 14:68:01-14:71, 14:73-14:76, 14:79, 14:82, 14:86-14:88, 14:90, 14:93, 14:99, 14:101, 14:104-14:105, 14:107, 14:110- 14:114, 14:117-14:120, 14:122, 14:124-14:125, 14:128-14:129, 14:131, 14:137N-14:140, 14:142-14:143, 14:145-14:147, 14:149- 14:153, 14:157-14:158, 14:163, DRB4*01:03:01:02N
19^{5,6,8}	110 bp, 145 bp, 170 bp	430 bp	3, 4, 6, 9, 11, 13, 14, 1404	*03:10, 09:01:02-09:01:05, 09:01:07-09:02:02, 09:04-09:22, 09:24, 09:26, 11:13:01-11:13:02, 11:17, 11:52, 13:43, 13:159, 13:171:01, 13:179, 13:191, 13:193, 14:01:01-14:02:02, 14:04:01- 14:11, 14:13-14:14, 14:16-14:18, 14:19 ^w , 14:20, 14:21 ^w , 14:22- 14:23:04, 14:26, 14:28-14:36, 14:38:01-14:39, 14:41, 14:43- 14:52, 14:54:01-14:54:03, 14:54:04 ^w , 14:54:05-14:57, 14:59- 14:62, 14:64-14:65, 14:68:01-14:68:02, 14:70-14:76, 14:79- 14:83, 14:86-14:88, 14:90-14:97, 14:99-14:101, 14:103-14:108, 14:109 ^w , 14:110-14:114, 14:117-14:127:02, 14:129-14:134, 14:137N-14:140, 14:142-14:143, 14:145-14:155, 14:157-14:160, 14:162-14:163, 15:27, 15:34, 15:66:01-15:66:02, 15:122 ^w
20^{5,6}	110 bp, 150 bp, 180 bp, 220 bp	430 bp	2 ^w , 3, 6, 8, 11, 13, 14, 1403, 1404, 16 ^w	*03:10, 08:09, 08:20-08:21, 08:32, 08:35, 08:36:02, 08:53, 08:68 ^w , 08:70 ^w , 11:13:01-11:13:02, 11:17, 11:23:01-11:23:02, 11:25, 11:31, 11:45, 11:52, 11:55, 11:64, 11:89, 11:96, 11:119, 11:148, 11:159, 13:13, 13:18, 13:43, 13:45, 13:47, 13:55, 13:119, 13:144, 13:146, 13:154, 13:156, 13:158-13:159, 13:164, 13:171:01, 13:179, 13:191, 13:197, 14:01:01-14:01:04, 14:03:01- 14:05:04, 14:07:01-14:08, 14:10-14:12:02, 14:14-14:16, 14:18, 14:22-14:23:04, 14:25-14:28, 14:31-14:32:03, 14:34-14:36, 14:38:01-14:40, 14:42-14:45, 14:49-14:50, 14:53-14:65, 14:67- 14:79, 14:81-14:82, 14:84-14:93, 14:95-14:97, 14:99-14:105, 14:107, 14:110-14:120, 14:122-14:140, 14:142-14:158, 14:160, 14:162-14:163, 15:21 ^w , 16:04 ^w , 16:18 ^w
21	170 bp	430 bp	8, 12, 13, 14	*03:92, 08:01:01-08:04:05, 08:04:07-08:06, 08:09-08:10, 08:12- 08:13, 08:16-08:18, 08:21-08:24, 08:26-08:30:03, 08:32-08:33, 08:35-08:40, 08:42-08:51, 08:53-08:62, 08:64-08:74, 08:76, 12:09, 12:46, 12:48 ^w , 13:17, 13:116, 13:175, 14:15, 14:52, 14:126:01
22^{5,6}	75 bp, 175 bp,	515 bp	2, 4, 8, 11, 13, 14	*04:12, 04:18, 04:25, 04:58, 08:04:01, 08:04:02 ^w -08:04:03 ^w , 08:04:04-08:04:07, 08:06, 08:10, 08:12, 08:14, 08:20, 08:22,

101.101.48/12 – including *Taq* pol., IFU-01
101.101.48u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **91Y**

Lot-specific information

	265 bp			08:28, 08:54, 08:59, 08:67, 08:70, 08:75, 11:25, 11:67, 11:119, 13:18, 13:144, 13:156, 13:164, 14:12:01-14:12:02, 14:15, 14:78, 14:84, 14:156, 15:21
23	135 bp	430 bp	8, 12, 13	*03:92, 08:17, 08:28, 08:37, 08:45, 08:63, 11:67, 12:01:01-12:20, 12:22-12:37, 12:39-12:41, 12:43-12:52, 12:54-12:56, 12:58, 13:17, 13:175, 14:138
24	180 bp	430 bp	3, 6, 11, 13, 14	*03:12, 13:03:01-13:04, 13:12:01-13:13, 13:21:01-13:21:02, 13:30, 13:32-13:33:03, 13:38, 13:48-13:49, 13:55, 13:58, 13:65-13:66:02, 13:75, 13:81, 13:88-13:90, 13:93-13:95, 13:101, 13:108, 13:115, 13:118, 13:120, 13:122, 13:133-13:134, 13:139, 13:151-13:152, 13:154, 13:161, 13:164, 13:167, 13:169, 13:174, 13:188, 13:194, 14:13, 14:63, 14:65, 14:78, 14:85, DRB5*01:08N
25	150 bp	430 bp	2, 4, 6, 11, 13	*11:16, 11:20, 11:40, 11:59, 11:63:02, 11:118, 11:122, 13:01:01-13:02:01, 13:02:03-13:02:12, 13:08, 13:15-13:16, 13:19-13:20, 13:27-13:29, 13:31-13:32, 13:34-13:36, 13:39-13:41, 13:43, 13:51-13:53, 13:57, 13:59, 13:61:01-13:61:02, 13:63-13:65, 13:67-13:69, 13:71-13:74, 13:76, 13:78-13:80, 13:83-13:84, 13:87, 13:91-13:93, 13:96:01-13:97:01, 13:98-13:99, 13:102-13:107, 13:109, 13:111-13:114, 13:117, 13:121, 13:123-13:128, 13:130-13:131, 13:135, 13:138-13:143, 13:145, 13:147-13:148, 13:153, 13:155, 13:159-13:160, 13:165-13:166, 13:168, 13:171:01-13:171:02, 13:173, 13:175, 13:177, 13:179, 13:182-13:187, 13:190-13:191, 13:195, 13:198, 13:200N-13:202, 14:16, 14:57, 15:10
26	145 bp	430 bp	3, 6, 11, 13, 14, 17	*03:01:01:01-03:01:22, 03:01:24-03:01:25, 03:04:01-03:05:03, 03:07-03:16, 03:18-03:25, 03:28-03:30, 03:32-03:34, 03:36-03:37, 03:39-03:40, 03:43-03:52, 03:54-03:59, 03:61-03:68N, 03:70-03:73, 03:76-03:79, 03:81, 03:83-03:84, 03:86, 03:89, 03:91, 03:93-03:101, 03:104-03:110, 03:112-03:114, 03:116-03:118, 03:120-03:121, 11:01:01-11:01:06, 11:01:08-11:04:12, 11:06:01-11:11:01, 11:11:03-11:16, 11:18-11:21, 11:23:01-11:29:01, 11:31-11:36, 11:38-11:40, 11:42-11:47, 11:49:01-11:49:02, 11:51, 11:53-11:55, 11:57-11:66, 11:68-11:70, 11:72-11:82, 11:84:01-11:87, 11:89-11:103:02, 11:105-11:106, 11:108-11:144, 11:146-11:148, 11:150-11:169N, 11:172-11:176, 11:178-11:182, 13:01:01-13:01:08, 13:01:10-13:02:12, 13:04-13:06, 13:09-13:11:02, 13:14:01-13:16, 13:18, 13:20-13:25, 13:27-13:31, 13:34, 13:39, 13:41-13:46, 13:50:01-13:52, 13:54, 13:56-13:57, 13:59, 13:61:01-13:64, 13:66:01-13:66:02, 13:68-13:69, 13:71, 13:73-13:75, 13:77-13:80, 13:82-13:83, 13:86-13:87, 13:91-13:93, 13:96:01-13:100, 13:102, 13:104-13:107, 13:109-13:114, 13:117, 13:119, 13:121, 13:123-13:132, 13:136-13:148, 13:150, 13:153-13:158, 13:162-13:163, 13:165-13:166, 13:168-13:173, 13:176, 13:178-13:179, 13:181, 13:184-13:187, 13:190, 13:192-13:193, 13:196-13:197, 13:199-13:204, 14:17, 14:21, 14:30, 14:33, 14:35, 14:42, 14:53, 14:64-14:65, 14:72, 14:95, 14:128, 14:132, 14:134
27^{6,7,9}	145 bp, 210 bp, 235 bp	430 bp	6, 8, 11, 13, 14, 1403, 1404	*08:09, 08:20-08:21, 08:35, 11:13:01-11:13:02, 11:17, 11:23:01-11:23:02, 11:25, 11:45, 11:52, 11:55, 11:64, 11:89, 11:119, 11:149, 11:159, 13:08, 13:13, 13:18, 13:47, 13:55, 13:119, 13:144, 13:146, 13:154, 13:156, 13:158, 13:164, 14:01:01-14:01:04, 14:02:02 [?] , 14:03:01-14:05:04, 14:07:01-14:08, 14:09 [?] , 14:10-14:12:02, 14:14-14:15, 14:16 [?] , 14:18, 14:19 [?] -14:20 [?] , 14:22 [?] , 14:23:01-14:23:04, 14:25 [?] , 14:26-14:28, 14:29 [?] -14:30 [?] , 14:31-14:32:02, 14:33 [?] , 14:34-14:36, 14:37 [?] , 14:38:01-14:40,

101.101.48/12 – including *Taq* pol., IFU-01
101.101.48u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **91Y**

Lot-specific information

				14:41 [?] , 14:42-14:45, 14:47 [?] -14:49 [?] , 14:50, 14:51 [?] -14:53 [?] , 14:54:01-14:65, 14:67-14:68:02, 14:69 [?] , 14:70-14:72, 14:73 [?] - 14:74 [?] , 14:75, 14:76 [?] , 14:77-14:78, 14:79 [?] -14:80 [?] , 14:81-14:82, 14:83 [?] , 14:84-14:93, 14:94 [?] , 14:95-14:97, 14:98 [?] , 14:99-14:103, 14:104 [?] -14:109 [?] , 14:110, 14:111 [?] , 14:112-14:118, 14:119 [?] , 14:120, 14:121 [?] , 14:122-14:127:02, 14:128 [?] , 14:129-14:145, 14:146 [?] , 14:147-14:158, 14:159 [?] , 14:160, 14:162-14:163
28	140 bp	430 bp	6, 13, 14	*03:76, 13:10, 13:85, 13:120, 13:170, 13:176, 13:178, 13:193, 13:196, 14:02:01-14:02:02, 14:06:01-14:06:03, 14:09, 14:13, 14:17, 14:19, 14:21, 14:29-14:30, 14:33, 14:46-14:48, 14:51- 14:52, 14:80, 14:83, 14:106, 14:108-14:109, 14:121
29^{6,11,12}	160 bp, 240 bp	430 bp	52	*14:141, DRB3*01:01:02:01-01:16, DRB3*02:01-02:29N, DRB3*03:01:01-03:03
30^{7,10}	215 bp	430 bp	53	DRB4*01:01:01-01:09
31	175 bp	430 bp	51	DRB5*01:01:01-01:15, DRB5*02:02-02:06
32¹³	-	-		Negative control

¹Alleles are assigned by the presence of specific PCR product(s). However, the sizes of the specific PCR products may be helpful in the interpretation of DR low resolution SSP subtypings.

When the primers in a primer mix can give rise to HLA-specific PCR products of more than one length this is indicated if the size difference is more than 20 base pairs. Size differences of 20 base pairs or less are not given. For high resolution SSP kits, the alleles listed are specified according to amplicon length. Nonspecific amplifications, i.e. a ladder or a smear of bands, may sometimes be seen. GC-rich primers have a higher tendency of giving rise to nonspecific amplifications than other primers, e.g. the primers in wells 3, 18, 19 and 20.

PCR fragments longer than the control bands may sometimes be observed. Such bands should be disregarded and do not influence the interpretation of the SSP typings.

PCR fragments migrating faster than the control bands, but slower than a 400 bp fragment may be seen in some gel read-outs. Such bands can be disregarded and do not influence the interpretation of the SSP typings.

Some primers may give rise to primer oligomer artifacts. Sometimes this phenomenon is an inherent feature of the primer pair(s) of a primer mix. More often it is due to other factors such as too low amount of DNA in the PCR reactions, taking too long time in setting up the PCR reactions, working at elevated room temperature or using thermal cyclers that are not pre-heated.

²The internal positive control primer pairs amplify segments of the human growth hormone gene. The internal positive control bands are 430 or 515 base pairs respectively, well distribution as outlined in the table. Well number 1 contains the longer, 515 bp, internal positive control band. The well distribution of the internal controls can help in orientation of the kit on gel photo, as well as allow for kit identification. In the presence of a specific amplification the intensity of the control band often decreases.

³The serological reactivity of all DRB alleles is not known. In this table we use the information in the HLA Dictionary 2004 (on the www.ebi.ac.uk/imgt/hla web site and the expert-assigned serological grouping in Tissue Antigens (2009) 73:95-170.

⁴For several DRB alleles 1st and/or 3rd exon(s) and beyond, as well as intron nucleotide sequences, are not available. In these instances it is not known whether some of the primers of the SSP sets are completely matched with the target sequences or not. Assumption is made that unknown sequences in these regions are conserved within allelic groups.

⁵HLA-specific PCR products shorter than 125 base pairs have a lower intensity and are less sharp than longer PCR products.

⁶Individual alleles can give rise to two differently sized specific PCR fragments in primer mix 1, 3, 5 to 8, 10, 11, 13, 14, 16, 18 to 20, 22, 27 and 29.

⁷Primer mixes 1, 3, 5, 8, 11, 12, 13, 16, 18, 27 and 30 have a tendency to giving rise to primer oligomer formation, most pronounced in primer mix 12.

⁸Primer mix 19 has a tendency of primer oligomer formation and also has an intense primer cloud due to the high number of primers present in the primer mix.

⁹Primer mixes 3, 4, 9 and 27 may have tendencies of unspecific amplifications.

101.101.48/12 – including *Taq* pol., IFU-01
101.101.48u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: 91Y

Lot-specific information

¹⁰The DRB4*01:03:01:02N allele is amplified by the primer pairs in wells 18 and 30, whereas the DRB4*02:01N and DRB4*03:01N null alleles are not amplified by these primer pairs.

¹¹Due to sharing of sequence motifs, DRB3*02:27 is amplified by the primer pairs in well 15 in addition to primer mix 29.

¹²Due to sharing of sequence motifs in codon 38 and 47, DRB3*01:14 will also be amplified in primer mixes 5, 6 and 17 and DRB3*01:23 in mix 5, in addition to primer mix 29.

¹³Primer mix 32 contains a negative control, which will amplify more than 95% of HLA amplicons as well as the amplicons generated by the control primer pairs matching the human growth hormone gene. HLA-specific PCR product sizes range from 75 to 200 base pairs and the PCR product generated by the HGH positive control primer pair is 430 base pairs.

‘w’, might be weakly amplified.

‘?’, nucleotide sequence information not available for the primer matching sequence.

101.101.48/12 – including *Taq* pol., IFU-01
101.101.48u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **91Y**

Lot-specific information

PRIMER SPECIFICATION

Well No.	1	2	3	4	5	6	7	8	9	10	11	12
Length of spec.	210	200	210	210	120	75	85	100	235	170	90	175
PCR product	235		230		220	210	210	175		215	135	
	260									250	165	
											190	
Length of int.	515	430	430	430	430	430	430	430	430	515	430	430
pos. control ¹												
5'-primer(s) ²	12(124) 5'-A.T 3'	14(129) 5'-gAA 3'	13(126) 5'-Agg 3'	13(126) 5'-Agg 3'	13(125) 5'-gTC 3'	13(125) 5'-gTC 3'	13(125) 5'-gTC 3'	13(125) 5'-ACA 3'	13(127) 5'-gTA 3'	15(133) 5'-gTT 3'	26(165) 5'-TAT 3'	26(164) 5'-gTA 3'
	14(129) 5'-gAA 3'		13(126) 5'-AAg 3'	13(126) 5'-AAg 3'	47(227) 5'-gTT 3'	15(133) 5'-gTT 3'		13(125) 5'-ACC 3'	13(127) 5'-ATA 3'	15(133) 5'-gTT 3'	36(196) 5'-AgA 3'	30(178) 5'-gCg 3'
			13(126) 5'-AgA 3'					13(125) 5'-ATA 3'	13(127) 5'-ATA 3'	15(133) 5'-gCT 3'	58(261) 5'-gAg 3'	
								13(125) 5'-gTC 3'				
3'-primer(s) ³	66(286) 5'-gAg 3'	66(286) 5'-gAT 3'	66(286) 5'-gAT 3'	66(286) 5'-gAA 3'	73(305) 5'-ggC 3'	26(164) 5'-ggT 3'	28(171) 5'-CTC 3'	32(184) 5'-gTg 3'	77(319) 5'-CAA 3'	58(260) 5'-CCT 3'	57(257) 5'-CgA 3'	72(303) 5'-gCA 3'
	66(286) 5'-gAg 3'		66(286) 5'-gAT 3'	66(286) 5'-gAg 3'	73(305) 5'-ggC 3'	66(286) 5'-gAT 3'	69(295) 5'-CTg 3'	58(260) 5'-Cgg 3'	77(319) 5'-gTA 3'	73(307) 5'-CAG 3'	73(305) 5'-ggC 3'	73(307) 5'-CgC 3'
	66(286) 5'-gAT 3'		69(295) 5'-CTg 3'	70(296) 5'-TgT 3'	73(305) 5'-ggC 3'	71(299) 5'-gCT 3'			77(319) 5'-CAC 3'	86(344) 5'-CAC 3'	77(319) 5'-CAC 3'	
	70(297) 5'-CTg 3'		69(295) 5'-Tg 3'	71(301) 5'-ggC 3'	74(308) 5'-CCC 3'				77(319) 5'-CgC 3'			
	71(299) 5'-gCg 3'		70(298) 5'-CgC 3'		74(310) 5'-CAA 3'							
	77(317) 5'-AgT 3'		71(299) 5'-gCT 3'									
	86(344) 5'-CCA 3'		73(305) 5'-ggC 3'									
			77(317) 5'-AgT 3'									
Well No.	1	2	3	4	5	6	7	8	9	10	11	12

Well No.	13	14	15	16	17	18	19	20	21	22	23	24
Length of spec.	100	90	220	205	175	100	110	110	170	75	135	180
PCR product	170	110		225		150	145	150		175		
						195	170	180		265		
						240		220				
Length of int.	430	430	430	430	430	430	430	430	430	515	430	430
pos. control ¹												
5'-primer(s) ²	13(125) 5'-gTC 3'	12(124) 5'-Cgg 3'	10(116) 5'-gCT 3'	10(116) 5'-gCT 3'	13(125) 5'-gTC 3'	5(101) 5'-CAA 3'	26(164) 5'-gTA 3'	13(125) 5'-gTC 3'	13(125) 5'-ggg 3'	12(122) 5'-TAg 3'	15(133) 5'-gTT 3'	13(125) 5'-gTC 3'
	15(133) 5'-gTC 3'	15(133) 5'-gTT 3'	12(122) 5'-TAT 3'	12(122) 5'-TAT 3'		37(197) 5'-gTT 3'	34(189) 5'-CAG 3'	34(189) 5'-CAG 3'	15(133) 5'-gTT 3'	74(308) 5'-CCT 3'		107(409) 5'-AgA 3'
	38(200) 5'-CgT 3'		13(125) 5'-gTC 3'	13(125) 5'-gTC 3'		37(197) 5'-gTA 3'		36(196) 5'-AgC 3'	15(133) 5'-gCT 3'			
				13(125) 5'-gTg 3'								
				15(133) 5'-gTT 3'								
				15(133) 5'-gTC 3'								
3'-primer(s) ³	58(260) 5'-CCT 3'	28(171) 5'-CTC 3'	69(295) 5'-gTC 3'	66(286) 5'-gAA 3'	58(260) 5'-Cgg 3'	42(213) 5'-TCA 3'	57(257) 5'-CAG 3'	57(257) 5'-CAG 3'	56(256) 5'-gCT 3'	56(256) 5'-gCT 3'	47(227) 5'-ggA 3'	56(256) 5'-gCT 3'
	58(260) 5'-CCT 3'	29(175) 5'-gTg 3'	71(299) 5'-gCT 3'	70(298) 5'-CgC 3'	58(260) 5'-CAG 3'	57(257) 5'-CAG 3'	69(295) 5'-CTg 3'	59(265) 5'-gTg 3'	57(257) 5'-CAT 3'	86(344) 5'-CCA 3'		159(565) 5'-CAT 3'
	58(260) 5'-CCT 3'	37(199) 5'-CAG 3'	71(299) 5'-ACT 3'	70(298) 5'-CTC 3'		70(298) 5'-CgC 3'	70(296) 5'-TCC 3'	70(296) 5'-TCC 3'	57(257) 5'-CAT 3'			
								73(307) 5'-CAG 3'				
Well No.	13	14	15	16	17	18	19	20	21	22	23	24

101.101.48/12 – including *Taq* pol., IFU-01
101.101.48u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **91Y**

Lot-specific information

Well No.	25	26	27	28	29	30	31
Length of spec. PCR product	150	145	145	140	160	215	175
			210		240		
			235				
Length of int. pos. control ¹	430	430	430	430	430	430	430
5'-primer(s) ²	34(189) 5'-CAg 3'	13(125) 5'-gTC 3'	13(125) 5'-gTC 3'	36(196) 5'-AgA 3'	10(116) 5'-gCT 3'	28(170) 5'-gAT 3'	13(125) 5'-gTA 3'
	34(189) 5'-CAg 3'		37(197) 5'-gTT 3'		10(116) 5'-gCT 3'		
			114(429) 5'-CTg 3'		37(199) 5'-TCC 3'		
3'-primer(s) ³	70(298) 5'-CTC 3'	47(227) 5'-ggA 3'	70(296) 5'-TCC 3'	69(295) 5'-CTg 3'	51(239) 5'-CCC 3'	86(346) 5'-CTC 3'	57(258) 5'-gCg 3'
		47(229) 5'-CCA 3'	73(307) 5'-CAg 3'	70(298) 5'-CTT 3'	77(317) 5'-AAT 3'	86(346) 5'-CTT 3'	58(260) 5'-CCT 3'
			181(630) 5'-CTT 3'				
Well No.	25	26	27	28	29	30	31

¹The internal positive control primer pairs amplify segments of the human growth hormone gene. The internal positive control bands are 430 or 515 base pairs respectively, well distribution as outlined in the table. Well number 1 contains the longer, 515 bp, internal positive control band. The well distribution of the internal controls can help in orientation of the kit on gel photo, as well as allow for kit identification. In the presence of a specific amplification the intensity of the control band often decreases.

²The nucleotide position matching the specificity-determining 3'-end of the primer is given. Nucleotide numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.

³The nucleotide position matching the specificity-determining 3'-end of the primer is given in the anti-sense direction. Nucleotide numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot-specific information

CELL LINE VALIDATION SHEET																					
DR low resolution primer set ³																					
					Prod. No.:	Well ²															
						1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16
						201553601	201553602	201553603	201557404	201557405	201553606	201553607	201553608	201557409	201553610	201554911	201553612	201553613	201557414	201553615	201553616
	IHWC cell line ¹		DRB1																		
1	9001	SA	*01:01			+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
2	9280	LK707	*15:02	*04:05		-	-	+	-	-	-	-	+	-	-	-	-	-	-	-	-
3	9011	E4181324	*15:02				-	+	-	-	-	-	-	-	-	-	-	-	-	-	-
4	9275	GU373	*03:01			-	-	-	-	+	+	-	-	-	-	-	-	-	-	-	-
5	9009	KAS011	*16:01			-	-	-	+	-	-	-	-	-	-	-	-	-	-	-	-
6	9353	SM	*04:07	*08:03		-	-	-	-	-	-	-	+	-	+	-	-	-	-	-	-
7	9020	QBL	*03:01			-	-	-	-	+	+	-	-	-	-	-	-	-	-	-	-
8	9025	DEU	*04:01			-	-	-	-	-	-	-	+	-	-	-	-	-	-	-	-
9	9026	YAR	*04:02			-	-	-	-	-	-	-	+	-	-	-	-	-	-	-	-
10	9107	LKT3	*04:05			-	-	-	-	-	-	-	+	-	-	-	-	-	-	-	-
11	9051	PITOUT	*07:01			-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-
12	9052	DBB	*07:01			-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-
13	9004	JESTHOM	*01:01			+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
14	9071	OLGA	*08:02				-	-	-	-	-	-	-	-	+	-	-	-	-	-	+
15	9075	DKB	*09:01			-	-	-	-	-	-	-	-	-	-	+	-	-	-	-	-
16	9037	SWEIG007	*11:01			-	-	-	-	-	-	-	-	-	-	-	-	+	-	+	+
17	9282	CTM3953540	*03:01	*13:01		-	-	-	-	+	+	-	-	-	-	-	-	-	-	+	+
18	9257	32367	*09:01	*11:01		-	-	-	-	-	-	-	-	-	-	+	-	+	-	+	+
19	9038	BM16	*12:01			-	-	-	-	-	-	-	-	-	-	-	-	-	+	-	-
20	9059	SLE005	*13:02			-	-	-	-	-	+	-	-	-	-	-	-	-	-	+	+
21	9064	AMALA	*14:02			-	-	-	-	-	-	+	-	-	-	-	-	-	-	-	-
22	9056	KOSE	*13:02	*14:54		-	-	-	-	-	+	-	-	-	-	-	-	-	-	+	+
23	9124	IHL	*08:03	*14:14		-	-	-	-	-	-	-	-	-	+	-	-	-	-	-	-
24	9035	JBUSH	*11:01			-	-	-	-	-	-	-	-	-	-	-	-	+	-	+	+
25	9049	IBW9	*07:01			-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-
26	9285	WT49	*03:01			-	-	-	-	+	+	-	-	-	-	-	-	-	-	-	-
27	9191	CH1007	*04:05	*10:01		-	-	-	-	-	-	-	+	-	-	-	+	-	-	-	-
28	9320	BEL5GB	*04:16	*07:01		-	-	-	-	-	-	-	+	+	-	-	-	-	-	-	-
29	9050	MOU	*07:01			-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-
30	9021	RSH	*03:02			-	-	-	-	+	-	+	-	-	-	-	-	-	-	-	-
31	9019	DUCAF	*03:01			-	-	-	-	+	+	-	-	-	-	-	-	-	-	-	-
32	9297	HAG	*13:03			-	-	-	-	-	+	-	-	-	-	-	-	-	-	+	-
33	9098	MT14B	*04:04			-	-	-	-	-	-	-	+	-	-	-	-	-	-	-	-
34	9104	DHIF	*11:01			-	-	-	-	-	-	-	-	-	-	-	-	+	-	+	+
35	9302	SSTO	*04:03			-	-	-	-	-	-	-	+	-	-	-	-	-	-	-	-
36	9024	KT17	*04:03	*04:06		-	-	-	-	-	-	-	+	-	-	-	-	-	-	-	-
37	9065	HHKB	*13:01			-	-	-	-	-	+	-	-	-	-	-	-	-	-	+	+
38	9099	LZL	*14:02			-	-	-	-	-	-	+	-	-	-	-	-	-	-	-	-
39	9315	CML	*03:01	*04:01		-	-	-	-	+	+	-	+	-	-	-	-	-	-	-	-
40	9134	WHONP199	*07:01	*09:01		-	-	-	-	-	-	-	-	+	-	+	-	-	-	-	-
41	9055	H0301	*13:02			-	-	-	-	-	+	-	-	-	-	-	-	-	-	+	+
42	9066	TAB089	*08:03			-	-	-	-	-	-	-	-	-	+	-	-	-	-	-	-
43	9076	T7526	*09:01			-	-	-	-	-	-	-	-	-	-	+	-	-	-	-	-
44	9057	TEM	*14:01			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
45	9239	SHJO	*07:01			-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-
46	9013	SCHU	*15:01			-	-	+	-	-	-	-	-	-	-	-	-	-	-	-	-
47	9045	TUBO	*11:04	*12:01		-	-	-	-	-	-	-	-	-	-	-	-	+	+	+	+
48	9303	TER-ND	*01:03			-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-

101.101.48/12 – including *Taq* pol., IFU-01
101.101.48u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **91Y**

Lot-specific information

CELL LINE VALIDATION SHEET																							
DR low resolution primer set ³																							
					Prod. No.:	Well ²																	
						17	18	19	20	21	22	23	24	25	26	27	28	29	30	31			
						201557417	201553618	201553619	201553620	201557421	201557422	201557423	201557424	201557425	201557426	201557427	201557428	201553629	201553630	201553631			
	IHWC cell line ¹		DRB1																				
1	9001	SA	*01:01			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
2	9280	LK707	*15:02	*04:05		-	-	-	-	-	-	-	-	-	-	-	-	-	+	+	-	-	
3	9011	E4181324	*15:02			-	-	-	-	-	-	-	-	-	-	-	-	-	-	+	-	-	
4	9275	GU373	*03:01			+	-	-	-	-	-	-	-	-	+	-	-	+	-	-	-	-	
5	9009	KAS011	*16:01			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	+	-	
6	9353	SM	*04:07	*08:03		-	-	-	-	+	-	-	-	-	-	-	-	-	+	-	-	-	
7	9020	QBL	*03:01			+	-	-	-	-	-	-	-	-	+	-	-	+	-	-	-	-	
8	9025	DEU	*04:01			-	-	-	-	-	-	-	-	-	-	-	-	-	+	-	-	-	
9	9026	YAR	*04:02			-	-	-	-	-	-	-	-	-	-	-	-	-	-	+	-	-	
10	9107	LKT3	*04:05			-	-	-	-	-	-	-	-	-	-	-	-	-	+	-	-	-	
11	9051	PITOUT	*07:01			-	-	-	-	-	-	-	-	-	-	-	-	-	+	-	-	-	
12	9052	DBB	*07:01			-	+	-	-	-	-	-	-	-	-	-	-	-	+	-	-	-	
13	9004	JESTHOM	*01:01			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
14	9071	OLGA	*08:02			-	-	-	-	+	-	-	-	-	-	-	-	-	-	-	-	-	
15	9075	DKB	*09:01			-	-	+	-	-	-	-	-	-	-	-	-	-	+	-	-	-	
16	9037	SWEIG007	*11:01			-	-	-	-	-	-	-	-	-	+	-	-	+	-	-	-	-	
17	9282	CTM3953540	*03:01	*13:01		+	-	-	-	-	-	-	-	+	+	-	-	+	-	-	-	-	
18	9257	32367	*09:01	*11:01		-	-	+	-	-	-	-	-	-	+	-	-	+	+	-	-	-	
19	9038	BM16	*12:01			-	-	-	-	-	-	+	-	-	-	-	-	+	-	-	-	-	
20	9059	SLE005	*13:02			+	-	-	-	-	-	-	-	+	+	-	-	+	-	-	-	-	
21	9064	AMALA	*14:02			+	-	+	-	-	-	-	-	-	-	-	+	+	-	-	-	-	
22	9056	KOSE	*13:02	*14:54		+	+	+	+	-	-	-	-	+	+	+	-	+	-	-	-	-	
23	9124	IHL	*08:03	*14:14		+	-	+	+	+	-	-	-	-	-	+	-	+	-	-	-	-	
24	9035	JBUSH	*11:01			-	-	-	-	-	-	-	-	-	+	-	-	+	-	-	-	-	
25	9049	IBW9	*07:01			-	-	-	-	-	-	-	-	-	-	-	-	-	+	-	-	-	
26	9285	WT49	*03:01			+	-	-	-	-	-	-	-	-	+	-	-	+	-	-	-	-	
27	9191	CH1007	*04:05	*10:01		-	-	-	-	-	-	-	-	-	-	-	-	-	+	-	-	-	
28	9320	BEL5GB	*04:16	*07:01		-	-	-	-	-	-	-	-	-	-	-	-	-	+	-	-	-	
29	9050	MOU	*07:01			-	-	-	-	-	-	-	-	-	-	-	-	-	+	-	-	-	
30	9021	RSH	*03:02			+	-	-	-	-	-	-	-	-	-	-	-	+	-	-	-	-	
31	9019	DUCAF	*03:01			+	-	-	-	-	-	-	-	-	+	-	-	+	-	-	-	-	
32	9297	HAG	*13:03			+	-	-	-	-	-	-	+	-	-	-	-	+	-	-	-	-	
33	9098	MT14B	*04:04			-	-	-	-	-	-	-	-	-	-	-	-	-	+	-	-	-	
34	9104	DHIF	*11:01			-	-	-	-	-	-	-	-	-	+	-	-	+	-	-	-	-	
35	9302	SSTO	*04:03			-	-	-	-	-	-	-	-	-	-	-	-	-	+	-	-	-	
36	9024	KT17	*04:03	*04:06		-	-	-	-	-	-	-	-	-	-	-	-	-	+	-	-	-	
37	9065	HHKB	*13:01			+	-	-	-	-	-	-	-	+	+	-	-	+	-	-	-	-	
38	9099	LZL	*14:02			+	-	+	-	-	-	-	-	-	-	-	+	+	-	-	-	-	
39	9315	CML	*03:01	*04:01		+	-	-	-	-	-	-	-	-	+	-	-	+	+	-	-	-	
40	9134	WHONP199	*07:01	*09:01		-	-	+	-	-	-	-	-	-	-	-	-	-	+	-	-	-	
41	9055	H0301	*13:02			+	-	-	-	-	-	-	-	+	+	-	-	+	-	-	-	-	
42	9066	TAB089	*08:03			-	-	-	-	+	-	-	-	-	-	-	-	-	-	-	-	-	
43	9076	T7526	*09:01			-	-	+	-	-	-	-	-	-	-	-	-	-	+	-	-	-	
44	9057	TEM	*14:01			-	+	+	+	-	-	-	-	-	-	+	-	+	-	-	-	-	
45	9239	SHJO	*07:01			-	-	-	-	-	-	-	-	-	-	-	-	-	+	-	-	-	
46	9013	SCHU	*15:01			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	+	-	
47	9045	TUBO	*11:04	*12:01		-	-	-	-	-	-	+	-	-	+	-	-	+	-	-	-	-	
48	9303	TER-ND	*01:03			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	

¹The provided cell line HLA specificities are retrieved from the <http://www.ihwg.org/hla> web site. The specificity of an individual cell line may thus be subject to change.

101.101.48/12 – including *Taq* pol., IFU-01
101.101.48u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: 91Y

Lot-specific information

²The DRB4*01:03:01:02N allele is amplified by primer mix 18 in the 9052 (DBB) cell line.

³The specificity of each primer solution in the kit has been tested against 48 well characterized cell line DNAs and where applicable, additional cell line DNAs.

Additional 5'- and 3'-primers in primer solutions 11, 12, 18 and 27 were tested by separately adding one 3'-primer, respectively one 5'-primer.

Additional 5'-primers in primer solutions 6, 15 and 16 were tested by separately adding one 3'-primer. Additional 3'-primers in primer solutions 1, 3, 4, 9, 10, 20 and 22 were tested by separately adding one 5'-primer.

One, two or three of the 5'-primers in primer solution 1, 3, 4, 8 to 10, 13 to 16, 20 to 22 and 27 were not possible to test. One or two of the 3'-primers in primer solution 1, 3 to 6, 9, 13, 15, 21, 26 and 30 were not possible to test.

101.101.48/12 – including *Taq* pol., IFU-01
101.101.48u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **91Y**

Lot-specific information

ADDRESSES:

Manufacturer:

Olerup SSP AB, Franzengatan 5, SE-112 51 Stockholm, Sweden.

Tel: +46-8-717 88 27

Fax: +46-8-717 88 18

E-mail: olerup-se@caredx.com

Web page: <http://www.olerup.com>

Distributed by:

Olerup GmbH, Löwengasse 47 / 6, AT-1030 Vienna, Austria.

Tel: +43-1-710 15 00

Fax: +43-1-710 15 00 10

E-mail: olerup-at@caredx.com

Web page: <http://www.olerup.com>

Olerup Inc., 901 S. Bolmar St., Suite R, West Chester, PA 19382

Tel: 1-877-OLERUP1

Fax: 610-344-7989

E-mail: olerup-us@caredx.com

Web page: <http://www.olerup.com>

For information on *Olerup* distributors worldwide, contact **Olerup GmbH**.