

Olerup SSP® DRB1*03

Product number:	101.113-24u/04u – without <i>Taq</i> pol.
Lot number:	65F
Expiry date:	2011-March-01
Number of tests:	24 tests – Product No. 101.113-24u 4 tests – Product No. 101.113-04u
Number of wells per test:	30
Storage - pre-aliquoted primers:	dark at -20°C
- PCR Master Mix:	-20°C
- Adhesive PCR seals	RT
- Product Insert	RT

This Product Description is only valid for Lot No. 65F.

**CHANGES COMPARED TO THE PREVIOUS *OLERUP SSP*®
DRB1*03 Lot**

The DRB1*03 specificity and interpretation tables have been updated for the DRB alleles described since the previous *Olerup SSP*® DRB1*03 lot was made (**Lot No. 36E**).

One well has been added to the DRB1*03 kit,
well **30**.

The primers of the wells detailed below have been exchanged, added or modified compared to the previous lot.

Well	5'-primer	3'-primer	rationale
28	Exchanged	-	Exchanged 5'-primer for increased specificity of specific primer pair.
30	New	New	New primer pair for improved resolution of the DRB1*0335 and DRB1*0340 alleles.

PRODUCT DESCRIPTION

DRB1*03 SSP subtyping

CONTENT

The primer set contains 5'- and 3'-primers for identifying the DRB1*0301 to DRB1*0340 alleles.

PLATE LAYOUT

Each test consists of 30 PCR reactions in a 32 well cut PCR plate. Wells 31 and 32 are empty.

1	2	3	4	5	6	7	8
9	10	11	12	13	14	15	16
17	18	19	20	21	22	23	24
25	26	27	28	29	30	empty	empty

The 32 well cut PCR plate is marked with 'DRB1*03' in silver/gray ink.

Well No. 1 is marked with the Lot No. '65F'.

The PCR plates are heat-sealed with a PCR-compatible foil.

Please note: When removing each 32 well PCR plate, make sure that the remaining plates stay sealed. Use a scalpel or a similar instrument to carefully cut the foil between the plates.

INTERPRETATION

The interpretation of DRB1*03 PCR-SSP subtypings will be influenced by one DRB1*01 allele, one DRB1*04 allele, one DRB1*08 allele, four DRB1*09 alleles, several DRB1*11 alleles, three DRB1*12 alleles, one DRB1*15 allele, one DRB1*16 allele, as well as by many DRB1*13 and DRB1*14 alleles, when present on the other haplotype and also by the DRB3*0107, DRB3*0108, DRB3*0109, DRB3*0204, DRB3*0206, DRB3*0219, DRB3*0220, DRB3*0222 and DRB3*0303 alleles.

UNIQUELY IDENTIFIED ALLELES

All the phenotypically different DRB1*03 alleles, i.e. **DRB1*0301 to DRB1*0340**, recognized by the HLA Nomenclature Committee in January 2009¹ will give rise to unique amplification patterns by the primers in the DRB1*03 subtyping kit.

The DRB1*03 subtyping kit cannot distinguish the DRB1*030101 to 030106 alleles, the DRB1*030201 and 030202 alleles or the DRB1*030501 and 030502 alleles.

¹DRB1 alleles listed on the IMGT/HLA web page 2009-January-16, release 2.24.0, www.ebi.ac.uk/imgt/hla.

RESOLUTION IN HOMO- AND HETEROZYGOTES

The 40 DRB1*03 alleles can be combined in 820 homozygous and heterozygous combinations. 285 of these genotypes do not give rise to unique amplification patterns. The different sizes of the specific PCR products generated by the primers in vials 11, 14, 20 and 23 have not been considered in these calculations.

++++-+-+ +--+----- -+-----++ -----+	0302, 0323 = 0319, 0338 = 0323, 0338
++++-+-+ +--+----- -+-----+ -----+	0301, 0338 = 0302, 0313 = 0313, 0338
++-+-+--+ ++----- -+-+-----+ -----+	0303, 0325 = 0303, 0328
+--+--+--+ +-+----- -+-----+ -----+	0309, 0312 = 0309, 0316
+--+--+--+ +-+-----+ -+-----+ -----+	0309, 0315 = 0314, 0323
+--+--+--+ +-+----- -+-----+ -----+	0301, 0309 = 0305, 0323 = 0309, 0313 = 0309, 0319 = 0309, 0323
+--+--+--+ +-----+--+ -+-----+ -----+	0312, 0314 = 0314, 0316
+--+--+--+ +-----+--+ -+-----+ -----+	0305, 0312 = 0305, 0316
+--+--+--+ +-----+--+ -+-----+ -----+	0301, 0314 = 0305, 0315 = 0314, 0315
+--+++++- +-----+--+ -+-+-----+ -----+	0304, 0311 = 0304, 0324
+--+++++- +-----+--+ -+-+-----+ -----+	0304, 0312 = 0304, 0316
+--+++++- +-----+--+ -+-+-----+ -----+	0301, 0304 = 0304, 0328
+--+++++- +-----+--+ -+-+-----+ -----+	0307, 0308 = 0308, 0340
+--+++++- +-----+--+ -+-+-----+ -----+	0308, 0325 = 0308, 0328
+--+++++- +-----+--+ -+-+-----+ -----+	0307, 0323 = 0323, 0340
+--+++++- +-----+--+ -+-+-----+ -----+	0307, 0313 = 0313, 0340
+--+++++- +-----+--+ -+-+-----+ -----+	0307, 0311 = 0311, 0340
+--+++++- +-----+--+ -+-+-----+ -----+	0307, 0316 = 0316, 0340
+--+++++- +-----+--+ -+-+-----+ -----+	0307, 0337 = 0337, 0340
+--+++++- +-----+--+ -+-+-----+ -----+	0307, 0315 = 0315, 0340
+--+++++- +-----+--+ -+-+-----+ -----+	0307, 0320 = 0320, 0340
+--+++++- +-----+--+ -+-+-----+ -----+	0307, 0322 = 0322, 0340
+--+++++- +-----+--+ -+-+-----+ -----+	0307, 0325 = 0307, 0328 = 0328, 0340
+--+++++- +-----+--+ -+-+-----+ -----+	0301, 0321 = 0307, 0330
+--+++++- +-----+--+ -+-+-----+ -----+	0307, 0318 = 0318, 0340
+--+++++- +-----+--+ -+-+-----+ -----+	0307, 0319 = 0319, 0340
+--+++++- +-----+--+ -+-+-----+ -----+	0307, 0332 = 0332, 0340
+--+++++- +-----+--+ -+-+-----+ -----+	0307, 0333 = 0333, 0340
+--+++++- +-----+--+ -+-+-----+ -----+	0307, 0334 = 0334, 0340
+--+++++- +-----+--+ -+-+-----+ -----+	0307, 0336 = 0336, 0340
+--+++++- +-----+--+ -+-+-----+ -----+	0307, 0339 = 0339, 0340
+--+++++- +-----+--+ -+-+-----+ -----+	0301, 0307 = 0301, 0340
+--+++++- +-----+--+ -+-+-----+ -----+	0307, 0310 = 0310, 0340
+--+++++- +-----+--+ -+-+-----+ -----+	0307, 0312 = 0312, 0340
+--+++++- +-----+--+ -+-+-----+ -----+	0301, 0317 = 0317, 0337 = 0335, 0337
+--+++++- +-----+--+ -+-+-----+ -----+	0306, 0311 = 0311, 0335
+--+++++- +-----+--+ -+-+-----+ -----+	0312, 0331 = 0316, 0331
+--+++++- +-----+--+ -+-+-----+ -----+	0312, 0326 = 0316, 0326
+--+++++- +-----+--+ -+-+-----+ -----+	0306, 0312 = 0306, 0316

101.113-24u/04u – without *Taq* polymeraseLot No.: **65F**

Lot-specific information

+--+--	++-----	-+-+----	-+----+	0325, 0326 = 0326, 0328 = 0331, 0333
+--+--	++-----	-+-+----	-----+	0301, 0331 = 0306, 0325 = 0306, 0328 = 0328, 0331
+--+--	++-----	-+-----	-+----+	0301, 0326 = 0306, 0333 = 0326, 0333
+--+--	+--+-----	-+-----+	-----+	0311, 0323 = 0323, 0324
+--+--	+--+-----	-+-----	-----+	0311, 0313 = 0313, 0324
+--+--	+--+-----	-+-----+	-----+	0312, 0323 = 0316, 0323
+--+--	+--+-----	-+-----	-----+	0312, 0313 = 0313, 0316
+--+--	+--+-----	-+-+-----	-----+	0323, 0325 = 0323, 0328
+--+--	+--+-----	-+-+-----	-----+	0313, 0325 = 0313, 0328
+--+--	+--+-----	-+-----+	-----+	0301, 0323 = 0313, 0319 = 0313, 0323 = 0319, 0323 = 0323, 0323
+--+--	+--+-----	-+-----	-----+	0301, 0313 = 0313, 0313
+--+--	+--+-----	-+-+-----	-----+	0310, 0325 = 0310, 0328
+--+--	+--+-----	-+-----	-----+	0311, 0312 = 0311, 0316 = 0316, 0324
+--+--	+--+-----	-+-----	-----+	0311, 0337 = 0324, 0337
+--+--	+--+-----	-+-----	-----+	0311, 0315 = 0315, 0324
+--+--	+--+-----	++-----	-----+	0311, 0320 = 0320, 0324
+--+--	+--+-----	-+-+-----	-----+	0311, 0322 = 0322, 0324
+--+--	+--+-----	-+-+-----	-----+	0311, 0325 = 0311, 0328 = 0324, 0325 = 0324, 0328
+--+--	+--+-----	-+-+-----	-----+	0311, 0330 = 0324, 0330
+--+--	+--+-----	-+-+-----	-----+	0311, 0318 = 0318, 0324
+--+--	+--+-----	-+-+-----	-----+	0311, 0319 = 0319, 0324
+--+--	+--+-----	-+-----	+-----+	0311, 0332 = 0324, 0332
+--+--	+--+-----	-+-----	-+-----+	0311, 0333 = 0324, 0333
+--+--	+--+-----	-+-----	---+---+	0311, 0334 = 0324, 0334
+--+--	+--+-----	-+-----	---+---+	0311, 0336 = 0324, 0336
+--+--	+--+-----	-+-----	-----++	0311, 0339 = 0324, 0339
+--+--	+--+-----	-+-----	-----+	0301, 0311 = 0301, 0324
+--+--	+--+-----	-----	-----+	0311, 0311 = 0311, 0324
+--+--	+--+-----	-+-+-----	-----+	0312, 0337 = 0316, 0337
+--+--	+--+-----	-+-+-----	-----+	0312, 0315 = 0315, 0316
+--+--	+--+-----	++-----	-----+	0312, 0320 = 0316, 0320
+--+--	+--+-----	-+-+-----	-----+	0312, 0322 = 0316, 0322
+--+--	+--+-----	-+-+-----	-----+	0312, 0325 = 0312, 0328 = 0316, 0325 = 0316, 0328
+--+--	+--+-----	-+-+-----	-----+	0312, 0330 = 0316, 0330
+--+--	+--+-----	-+-+-----	-----+	0312, 0318 = 0316, 0318
+--+--	+--+-----	-+-+-----	-----+	0312, 0319 = 0316, 0319
+--+--	+--+-----	-+-----	+-----+	0312, 0332 = 0316, 0332
+--+--	+--+-----	-+-----	-+-----+	0312, 0333 = 0316, 0333
+--+--	+--+-----	-+-----	---+---+	0312, 0334 = 0316, 0334
+--+--	+--+-----	-+-----	---+---+	0312, 0336 = 0316, 0336
+--+--	+--+-----	-+-----	-----++	0312, 0339 = 0316, 0339
+--+--	+--+-----	-+-----	-----+	0301, 0312 = 0301, 0316 = 0312, 0316 = 0316, 0316
+--+--	+--+-----	-+-+-----	-----+	0325, 0337 = 0328, 0337
+--+--	+--+-----	-+-----	-----+	0301, 0337 = 0337, 0337

101.113-24u/04u – without *Taq* polymeraseLot No.: **65F**

Lot-specific information

+--+--+	+-----+	-+-+---	-----+	0315, 0325 = 0315, 0328
+--+--+	+-----+	-+-+---	-----+	0301, 0315 = 0315, 0315
+--+--+	+-----	++-+---	-----+	0320, 0325 = 0320, 0328
+--+--+	+-----	++-+---	-----+	0301, 0320 = 0320, 0320
+--+--+	+-----	-+++---	-----+	0322, 0325 = 0322, 0328
+--+--+	+-----	-++----	-----+	0301, 0322 = 0322, 0322
+--+--+	+-----	-+-+---	-----+	0318, 0325 = 0318, 0328
+--+--+	+-----	-+-+---	-----+	0319, 0325 = 0319, 0328
+--+--+	+-----	-+-+---	+-----+	0325, 0332 = 0328, 0332
+--+--+	+-----	-+-+---	-+-----+	0325, 0333 = 0328, 0333
+--+--+	+-----	-+-+---	--+-----+	0325, 0334 = 0328, 0334
+--+--+	+-----	-+-+---	---+-----+	0325, 0336 = 0328, 0336
+--+--+	+-----	-+-+---	-----++	0325, 0339 = 0328, 0339
+--+--+	+-----	-+-+---	-----+	0301, 0325 = 0301, 0328 =
				0325, 0328 = 0328, 0328
+--+--+	+-----	-+-----	-----+	0301, 0318 = 0318, 0318
+--+--+	+-----	-+-----	-----+	0301, 0319 = 0319, 0319
+--+--+	+-----	-+-----	+-----+	0301, 0332 = 0332, 0332
+--+--+	+-----	-+-----	-+-----+	0301, 0333 = 0333, 0333
+--+--+	+-----	-+-----	--+-----+	0301, 0334 = 0334, 0334
+--+--+	+-----	-+-----	---+-----+	0301, 0336 = 0336, 0336
+--+--+	+-----	-+-----	-----++	0301, 0339 = 0339, 0339
+--+--+	++-----	-+-----	-+-----+	0307, 0326 = 0326, 0340
+--+--+	++-----	-+-----	-----+	0306, 0307 = 0306, 0340
+--+---	++-+---	-----	-+-----+	0324, 0326 = 0326, 0335
+--+---	++-+---	-----	-----+	0306, 0324 = 0306, 0335
+--+---	++-+---	-----	-+-----+	0306, 0326 = 0326, 0326
+---++-	+-----	-+-+---	-----+	0304, 0304 = 0304, 0325
+---++-	+-----	-+-+---	-----+	0321, 0330 = 0330, 0340
-+++--+	-+-+---	-----+	-----+	0302, 0324 = 0303, 0327
-+++--+	-+-+---	-+-----+	-----+	0302, 0307 = 0303, 0329
-++-+-+	+++-----	-+-----+	-----+	0302, 0309 = 0309, 0338
-++-+-+	-++-+-+	-----+	-----+	0327, 0338 = 0335, 0338
-++-+-+	-++-+-+	-+-----+	-----+	0329, 0338 = 0338, 0340
-++-+-+	-++-+-+	-----+	-----+	0302, 0338 = 0338, 0338
-++-+-+	-+-+---	-+-----+	-----+	0327, 0329 = 0327, 0340 =
				0329, 0335
-++-+-+	-+-+---	-----+	-----+	0302, 0327 = 0302, 0335
-++-+-+	-+-+---	-+-----+	-----+	0302, 0329 = 0302, 0340
-++-+-+	-----	-+-----+	-----+	0329, 0329 = 0329, 0340
-++-+-+	-+-+---	-----+	-----+	0327, 0327 = 0327, 0335
-+-+---	-+-+---	-----+	-----+	0303, 0324 = 0303, 0335
-+-+---	-+-+---	-+-----+	-----+	0303, 0307 = 0303, 0340
--+-+---	+--+-----	-+-----+	-----+	0305, 0309 = 0309, 0309
--+-+---	+--+-----	-+-----	-----+	0305, 0314 = 0314, 0314
---+---	---+---	-+-----	-----+	0307, 0324 = 0324, 0340
---+---	-----	-+-----	-----+	0307, 0307 = 0307, 0340
-----+	-----	-+-+---	-----+	0321, 0321 = 0321, 0340
-----	-+-+---	-----	-----+	0317, 0317 = 0317, 0335

0301 = 030101-030106

0302 = 030201-030202

0305 = 030501-030502

SPECIFICITY TABLE

DRB1*03 SSP subtyping

Specificities and sizes of the PCR products of the 30 primer mixes used for DRB1*03 SSP subtyping

Primer Mix	Size of spec. PCR product ¹	Size of control band ²	Amplified DRB1*03 alleles ³	Other amplified DRB alleles ⁴
1	220 bp	515 bp	030101-030106, 0304, 0306, 0308, 0310-0313, 0315, 0318-0320, 0316 [?] , 0322, 0323, 0325, 0326, 0328, 0330-0334, 0336, 0337, 0339	0906, 1159, 1327, 1371, 1482
2⁵	85 bp	430 bp	030201-0303, 0327, 0329, 0338	1315, 1319, 1326, 1353, 1357, 1385, 1386, 1402-140302, 1406, 1412, 1413, 1418-1420, 1424, 1427, 1429, 1440, 1441, 1447-1449, 1451, 1463, 1467, 1477, 1478, 1481, 1483
3	190 bp	515 bp	030201-030202, 030501-030502, 0309, 0314, 0327, 0329, 0338	1109, 1120, 130201-130203, 130501-130502, 1326, 1329, 1331, 1334, 1336, 1339, 1341, 135002, 1356, 1363, 1365, 1367, 1373, 1374, 1385, 1402-140302, 1409, 1413, 1419, 1424, 1427, 1430, 1446-1448, 1451, 1463, 1467, 1608, DRB3*0108, DRB3*0206, DRB3*0220
4	190 bp	430 bp	030101-030106, 0303, 0306-0308, 0310-0313, 0315, 0316 [?] , 0318-0320, 0322-0324, 0326, 0328, 0332-0334, 0336, 0337, 0339	1116, 1140, 1158, 1159, 130101-130103, 1306, 1309, 1310, 1315, 1318, 1320, 1327, 1328, 1332, 1335, 1340, 1342, 1343, 1351, 1353, 1359, 1361, 1369, 1371, 1377-1380, 1387, 1406, 1412, 1417, 1418, 1421, 1429, 1433, 1452, 1459, 1464, 1478, 1480, 1481, 1483

5⁵	75 bp	430 bp	0304	
6⁵	105 bp	430 bp	030101-030106, 0304-030502, 0308-0315, 0316 ^{weakly} , 0318- 0320, 0322, 0323, 0325, 0328, 0330, 0332-0334, 0336, 0337, 0339	1159, 1327, 1341, 1371
7^{5,6}	135 bp	430 bp	0308	1159
8	185 bp	430 bp	030201-0303, 0307, 0321, 0329, 0338, 0340	0116, 0422, 1107, 1476, 1479, 1525, DRB3*0109, DRB3*0204, DRB3*0219, DRB3*0222, DRB3*0303
9⁵	135 bp	430 bp	030101-030106, 0304-0306, 0309, 0311, 0313 ^{weakly} , 0314-0316, 0318- 0320, 0322, 0323, 0325, 0326, 0328, 0330-0334, 0336, 0337, 0339	090202, 1327, 1341, 1371, DRB3*0107
10⁵	145 bp	430 bp	030201-0303, 0306, 0317, 0326, 0327, 0331, 0335, 0338	0820, 1117, 1137, 1141, 1148, 1150, 1152, 1156, 130301-130302, 130701- 1308, 1312, 1313, 1319, 1326, 1332, 1333, 1336- 1338, 1340, 1347-1349, 1353, 1355, 1358, 1360, 1365, 1370, 1372, 1376, 1381, 1384, 1385, 140101- 140301, 140501-1409, 1412- 1414, 1416, 1418-1420, 1422-1427, 1429, 143201- 143202, 1434, 1436-1441, 1443-1445, 1447-1449, 1451, 1454-1456, 1458- 1460, 1462, 1463, 1467, 1469, 1470, 1474, 1475, 1477, 1478, 1480-1483
11^{5,7}	105, 180 bp	515 bp	0309, 0313, 0323, 0338	1339, 1354, 1377-1379, 1448, 1464
12⁵	135 bp	430 bp	0310	1482
13	225 bp	430 bp	0311, 0317, 0324, 0327, 0335	1438
14^{5,8}	105, 135	515 bp	0312, 0316	0908

	bp			
15	155 bp	430 bp	0317, 0337	1139, 1143, 1150, 1459
16	195 bp	430 bp	0314, 0315	0905, 1159, 1327, 1341, 1371
17	255 bp	515 bp	0320	110601-110602, 1121, 1147, 1157, 1358, 1381, 1429
18⁵	120 bp	430 bp	030101-030106, 0304-030502, 0307-0310, 0312-0316, 0318-0323, 0325, 0328-0330, 0332-0334, 0336, 0337, 0339, 0340	1107, 1525
19	175 bp	515 bp	0322	
20^{5,9}	75, 120 bp	430 bp	0304 ^{weakly} , 0325, 0328, 0331	
21⁵	70 bp	430 bp	0321, 0330	111001-111002, 111201-111202 ^{weakly} , 111301-111302, 1205, 1214, 1215, 1357, 1364, 1383, 1435, 1465, 1472
22⁵	95 bp	430 bp	0318	
23^{5,10}	65, 165 bp	430 bp	0309, 0319, 0323	
24⁵	65 bp	430 bp	030201-0303, 0327, 0329, 0338	1315, 1326, 1353, 1385, 1478, 1402-140302, 1406, 1412, 1413, 1418, 1419, 1424, 1427, 1429, 1446-1448, 1451, 1452, 1463, 1467, 1478, 1481, DRB3*0220
25	180 bp	430 bp	0332	
26⁵	115 bp	515 bp	0326, 0333	1144, 1149, 1150
27	230 bp	515 bp	0334	
28	160 bp	430 bp	0336	
29	180 bp	430 bp	0339	
30	175 bp	430 bp	03010101-0307, 0309, 0311-0340	0820, 130101-1316, 1318-1342, 1344, 1346-1366, 1368-1387, 1402-140302, 140501-1406, 1409, 1412-1414, 1417-1421, 142301, 1424, 1427, 1429, 1430, 1433, 1436, 1437, 1440-1445, 1447, 1448, 1451, 1456, 1459, 1463-1465, 1467, 1477, 1478, 1480, 1481, 1483

¹Alleles are assigned by the presence of specific PCR product(s). However, the sizes of the specific PCR products may be helpful in the interpretation of DRB1*03SSP subtypings.

When the primers in a primer mix can give rise to specific PCR products of more than one length this is indicated if the size difference is 20 base pairs or more. Size differences shorter than 20 base pairs are not given. For high resolution SSP kits the respective length of the specific PCR product(s) of the alleles amplified by these primer mixes are given.

Nonspecific amplifications, i.e. a ladder or a smear of bands, may sometimes be seen. GC-rich primers have a higher tendency of giving rise to nonspecific amplifications than other primers.

PCR fragments longer than the control bands may sometimes be observed. Such bands should be disregarded and do not influence the interpretation of the SSP typings.

PCR fragments migrating faster than the control bands, but slower than a 400 bp fragment may be seen in some gel read-outs. Such bands can be disregarded and do not influence the interpretation of the SSP typings.

Some primers may give rise to primer oligomer artifacts. Sometimes this phenomenon is an inherent feature of the primer pair(s) of a primer mix. More often it is due to other factors such as too low amount of DNA in the PCR reactions, taking too long time in setting up the PCR reactions, working at elevated room temperature or using thermal cyclers that are not pre-heated.

²The internal positive control primer pairs amplify segments of the human growth hormone gene. The two different control primer pairs give rise to either an internal positive control band of 430 base pairs, for most wells, or a band of 515 base pairs, for some wells.

Well number 1 contains the primer pair giving rise to the longer, 515 bp, internal positive control band in order to help in the correct orientation of the DRB1*03 subtyping.

In addition, wells number 3, 11, 14, 17, 19, 26 and 27 contain the primer pair giving rise to the longer, 515 bp, internal positive control band in order to allow kit identification.

In the presence of a specific amplification the intensity of the control band often decreases.

³For several DRB alleles only partial second exon nucleotide sequences are available. In these instances it is not known whether some of the primers of the SSP sets are completely matched with the target sequences or not. We assume that unknown sequences in the first hyperpolymorphic region of the second exon of DRB alleles are conserved within allelic groups and that unknown sequences of codons 87 to 92 are identical with the DRB1*0101 consensus sequence.

⁴Due to the sharing of sequence motifs within the DR52 group of DRB1 alleles, several DRB1*11 and DRB1*12 as well as many DRB1*13 and DRB1*14 alleles are amplified by some of the DRB1*03 primer mixes. In addition, the DRB1*0116 allele, the DRB1*0422 allele, the DRB1*0820 allele, the DRB1*090202, 0905, 0906 and 0908 alleles, the DRB1*1525 allele and DRB1*1608 allele are amplified by a few of the DRB1*03 primer mixes. Finally, the DRB3*0107 allele is amplified by primer mix 9, the DRB3*0108 and DRB3*0206 alleles are amplified by primer mix 3, the DRB3*0109, DRB3*0204, DRB3*0219, DRB3*0222 and DRB3*0303 alleles by primer mix 8 and the DRB1*0220 allele by primer mixes 3 and 24.

⁵Specific PCR fragments shorter than 150 base pairs have a lower intensity than longer PCR bands.

⁶Primer mix 7 may have a tendency of giving rise to non specific amplifications.

⁷Primer mix 11: Specific PCR fragment of 105 bp in the DRB1*0309 and 0323 alleles. Specific PCR fragment of 180 bp in the DRB1*0313 and 0338 and the 1339, 1354, 1377-1379, 1448 and 1464 alleles.

⁸Primer mix 14: Specific PCR fragment of 105 bp in the DRB1*0316 allele. Specific PCR fragment of 135 bp in the DRB1*0312 and in the DRB1*0908 allele.

⁹Primer mix 20: Specific PCR fragment of 75 bp in the DRB1*0304^{weakly}, 0325 and 0331 alleles. Specific PCR fragment of 120 bp in the DRB1*0328 allele.

¹⁰Primer mix 23: Specific PCR fragment of 65 bp in the DRB1*0309 and 0323 alleles. Specific PCR fragment of 165 bp in the DRB1*0319 allele.

'?', nucleotide sequence information not available for the primer matching sequence.

INTERPRETATION TABLE																
DRB1*03 SSP subtyping																
Amplification patterns of the DRB1*0301 to *0340 alleles																
	Well ⁴															
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16
Length of spec.	220	85	190	190	75	105	135	185	135	145	105	135	225	105	155	195
PCR product											180			135		
Length of int.	515	430	515	430	430	430	430	430	430	430	515	430	430	515	430	430
pos. control ¹																
5'-primer ²	26	13	37	37	26	26	26	26	26	13	13	26	13	26	13	26
	5'-g TA ³ 5'-g TC ³ 5'-Ag A ³ 5'-Ag A ³ 5'-g TA ³ 5'-g TA ³ 5'-g TA ³ 5'-g TT ³ 5'-g TA ³ 5'-g TC ³ 5'-g TC ³ 5'-g TA ³ 5'-g TC ³ 5'-g TA ³ 5'-g TC ³ 5'-g TA ³															
3'-primer(s) ³	86	28	86	86	37	47	58	74	57	47	34	57	74	48	51	77
	5'-C CA ³ 5'-CT C ³ 5'-C AC ³ 5'-C CA ³ 5'-C gg ³ 5'-g gA ³ 5'-C CT ³ 5'-C CC ³ 5'-C AT ³ 5'-g gT ³ 5'-C CC ³ 5'-C Ag ³ 5'-C CT ³ 5'-CCA ³ 5'-C CC ³ 5'-A gg ³															
											60			57		
											5'-A gg ³			5'-gCT ³		
Well No.	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16
DRB1 allele																
*03010101-030106	1			4		6			9							
*030201-030202		2	3					8		10						
*0303		2		4				8		10						
*0304	1				5	6			9							
*030501-030502			3			6			9							
*0306	1			4					9	10						
*0307				4				8								
*0308	1			4		6	7									
*0309			3			6			9		11					
*0310	1			4		6						12				
*0311	1			4		6			9				13			
*0312	1			4		6								14		
*0313	1			4		6			w		11					
*0314			3			6			9							16
*0315	1			4		6			9							16
*0316	1			4		6			9					14		
*0317										10			13		15	
*0318	1			4		6			9							
*0319	1			4		6			9							
*0320	?			?		w			9							
*0321								8								
*0322	1			4		6			9							
*0323	1			4		6			9		11					
*0324				4									13			
Well No.	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16

INTERPRETATION TABLE														
DRB1*03 SSP subtyping														
Amplification patterns of the DRB1*0301 to *0340 alleles														
Well ⁴														
17	18	19	20	21	22	23	24	25	26	27	28	29	30	
255	120	175	75	70	95	65	65	180	115	230	160	180	175	Length of spec.
			120			165								PCR product
515	430	515	430	430	430	430	430	430	515	515	430	430	430	Length of int.
														pos. control ¹
13	47	32	26	37	26	26	28	27	13	13	33	27	13	5'-primer ²
5'-g TC ³ 5'-g TT ³ 5'-TC g ³ 5'-g TA ³ 5'-g TT ³ 5'-g TA ³ 5'-g TA ³ 5'-g Ag ³ 5'-AC g ³ 5'-g TC ³ 5'-g TC ³ 5'-AAA ³ 5'-C Cg ³ 5'-g TC ³														
85	73	77	37	47	45	34	37	73	38	76	73	73	58	3'-primer(s) ³
5'-C Ag ³ 5'-g gC ³ 5'-A gT ³ 5'-gTA ³ 5'-g gA ³ 5'-CCg ³ 5'-C CC ³ 5'-gTT ³ 5'-g gC ³ 5'-g Cg ³ 5'-gTT ³ 5'-g gC ³ 5'-g gC ³ 5'-C gg ³														
			52			67							58	
			5'-CTT ³			5'-gAT ³							5'-C Ag ³	
17	18	19	20	21	22	23	24	25	26	27	28	29	30	Well No.
														DRB1 allele
	18												30	*03010101-030106
							24						30	*030201-030202
							24						30	*0303
	18		w										30	*0304
	18												30	*030501-030502
													30	*0306
	18												30	*0307
	18													*0308
	18					23							30	*0309
	18													*0310
													30	*0311
	18												30	*0312
	18												30	*0313
	18												30	*0314
	18												30	*0315
	18												30	*0316
													30	*0317
	18				22								30	*0318
	18					23							30	*0319
17	18												30	*0320
	18			21									30	*0321
	18	19											30	*0322
	18					23							30	*0323
													30	*0324
17	18	19	20	21	22	23	24	25	26	27	28	29	30	Well No.

Length of spec.	220	85	190	190	75	105	135	185	135	145	105	135	225	105	155	195
PCR product											180			135		
Well No.	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16
*0325	1					6			9							
*0326	1			4					9	10						
*0327		2	3							10			13			
*0328	1			4		6			9							
*0329		2	3					8								
*0330	1					6			9							
*0331	1								9	10						
*0332	1			4		6			9							
*0333	1			4		6			9							
*0334	1			4		6			9							
*0335										10			13			
*0336	1			4		6			9							
*0337	1			4		6			9						15	
*0338		2	3					8		10	11					
*0339	1			4		6			9							
*0340								8								
*0116, 0422, 1476, 1479								8								
*0820, 130301-130302, 130701-1308, 1312, 1313, 1333, 1337, 1338, 1347-1349, 1355, 1360, 1370, 1372, 1376, 1384, 140501-140503, 1414, 142301, 1436, 1437, 1443-1445, 1456										10						
*090202									9							
*0905																16
*0906	1															
*0908														14		
*110601-110602, 1121, 1147, 1157																
*1107, 1525								8								
*1109, 1120, 1367, 1608			3													
*111001-111002, 111301-111302, 1205, 1214, 1215, 1435, 1472																
Well No.	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16

255	120	175	75	70	95	65	65	180	115	230	160	180	175	Length of spec.
			120			165								PCR product
17	18	19	20	21	22	23	24	25	26	27	28	29	30	Well No.
	18		20										30	*0325
									26				30	*0326
							24						30	*0327
	18		20										30	*0328
	18						24						30	*0329
	18			21									30	*0330
			20										30	*0331
	18							25					30	*0332
	18								26				30	*0333
	18									27			30	*0334
													30	*0335
	18										28		30	*0336
	18												30	*0337
							24						30	*0338
	18											29	30	*0339
	18												30	*0340
														*0116, 0422, 1476, 1479
													30	*0820, 130301-130302, 130701-1308, 1312, 1313, 1333, 1337, 1338, 1347-1349, 1355, 1360, 1370, 1372, 1376, 1384, 140501-140503, 1414, 142301, 1436, 1437, 1443-1445, 1456
														*090202
														*0905
														*0906
														*0908
17														*110601-110602, 1121, 1147, 1157
	18													*1107, 1525
														*1109, 1120, 1367, 1608
				21										*111001-111002, 111301-111302, 1205, 1214, 1215, 1435, 1472
17	18	19	20	21	22	23	24	25	26	27	28	29	30	Well No.

Length of spec.	220	85	190	190	75	105	135	185	135	145	105	135	225	105	155	195
PCR product											180			135		
Well No.	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16
*111201-111202																
*1116, 1140, 1158, 1343	4															
*1117, 1137, 1141, 1148, 1152, 1156, *140101-140103, 140701-1408, 1416, 1422, 142302, 1425, 1426, 143201-143202, 1434, 1439, 1454, 1455, 1458, 1460, 1462, 1469, 1470, 1474, 1475										10						
*1139, 1143													15			
*1144, 1149																
*1150									10				15			
*1159	1			4		6	7									16
*130101-130103, 1306, 1309, 1310, 1318, 1320, 1328, 1335, 1342, 1351, 1359, 1361, 1369, 1380, 1387, 1417, 1421, 1433	4															
*130201-130203, 130501-130502, 1329, 1331, 1334, 135002, 1356, 1363, 1373, 1374, 1430			3													
*1304, 131101-131102, 131401-131403, 1316, 1321-1325, 1330, 1344, 1346, 135001, 1352, 1362, 1366, 1368, 1375, 1382, 1442																
*1315		2		4												
*1319, 1420, 1440, 1441, 1477	2								10							
Well No.	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16

255	120	175	75	70	95	65	65	180	115	230	160	180	175	Length of spec.
			120			165								PCR product
17	18	19	20	21	22	23	24	25	26	27	28	29	30	Well No.
				w										*111201-111202
														*1116, 1140, 1158, 1343
														*1117, 1137, 1141, 1148, 1152, 1156, *140101-140103, 140701-1408, 1416, 1422, 142302, 1425, 1426, 143201-143202, 1434, 1439, 1454, 1455, 1458, 1460, 1462, 1469, 1470, 1474, 1475
														*1139, 1143
								26						*1144, 1149
								26						*1150
														*1159
														*130101-130103, 1306, 1309, 1310, 1318, 1320, 1328, 1335, 1342, 1351, 1359, 1361, 1369, 1380, 1387, 1417, 1421, 1433
													30	*130201-130203, 130501-130502, 1329, 1331, 1334, 135002, 1356, 1363, 1373, 1374, 1430
													30	*1304, 131101-131102, 131401-131403, 1316, 1321-1325, 1330, 1344, 1346, 135001, 1352, 1362, 1366, 1368, 1375, 1382, 1442
							24						30	*1315
													30	*1319, 1420, 1440, 1441, 1477
17	18	19	20	21	22	23	24	25	26	27	28	29	30	Well No.

Length of spec.	220	85	190	190	75	105	135	185	135	145	105	135	225	105	155	195
PCR product											180			135		
Well No.	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16
*1326, 1385, 1402, 140301, 1413, 1419, 1424, 1427, 1447, 1451, 1463, 1467		2	3							10						
*1327, 1371	1			4		6			9							16
*1332, 1340, 1480				4						10						
*1336, 1365, 1409			3							10						
*1339			3								11					
*1341			3			6			9							16
*1353, 1406, 1412, 1418, 1478, 1481		2		4						10						
*1354											11					
*1357		2														
*1358, 1381										10						
*1364, 1383, 1465																
*1377-1379, 1464				4							11					
*1386		2														
*140302		2	3													
*1429		2		4						10						
*1438										10			13			
*1446			3													
*1448		2	3							10	11					
*1449		2								10						
*1452				4												
*1459				4						10					15	
*1482	1									10		12				
*1483		2		4						10						
DRB1 allele																
Well No.	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16
DRB3*0107									9							
DRB3*0108, 0206			3													
DRB3*0109, 0204, 0219, 0222, 0303							8									
DRB3*0220			3													
Well No.	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16

255	120	175	75	70	95	65	65	180	115	230	160	180	175	Length of spec.
			120			165								PCR product
17	18	19	20	21	22	23	24	25	26	27	28	29	30	Well No.
							24						30	*1326, 1385, 1402, 140301, 1413, 1419, 1424, 1427, 1447, 1451, 1463, 1467
													30	*1327, 1371
													30	*1332, 1340, 1480
													30	*1336, 1365, 1409
													30	*1339
													30	*1341
							24						30	*1353, 1406, 1412, 1418, 1478, 1481
													30	*1354
				21									30	*1357
17													30	*1358, 1381
				21									30	*1364, 1383, 1465
													30	*1377-1379, 1464
													30	*1386
							24						30	*140302
17							24						30	*1429
														*1438
							24							*1446
							24						30	*1448
														*1449
							24							*1452
													30	*1459
														*1482
													30	*1483
														DRB1 allele
17	18	19	20	21	22	23	24	25	26	27	28	29	30	Well No.
														DRB3*0107
														DRB3*0108, 0206
														DRB3*0109, 0204, 0219, 0222, 0303
							24							DRB3*0220
17	18	19	20	21	22	23	24	25	26	27	28	29	30	Well No.

¹The internal positive control primer pairs amplify segments of the human growth hormone gene. The two different control primer pairs give rise to either an internal positive control band of 430 base pairs, for most wells, or a band of 515 base pairs, for some wells.

Well number 1 contains the primer pair giving rise to the longer, 515 bp, internal positive control band in order to help in the correct orientation of the DRB1*03 subtyping.

In addition, wells number 3, 11, 14, 17, 19, 26 and 27 contain the primer pair giving rise to the longer, 515 bp, internal positive control band in order to allow kit identification.

²The codon, in the 2nd exon, matching the specificity-determining 3'-end of the primer is given. Codon numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given. Empty spaces indicate codon boundaries.

³The codon, in the 2nd exon, matching the specificity-determining 3'-end of the primer is given in the anti-sense direction. Codon numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given. Empty spaces indicate codon boundaries.

⁴Primer mix 11: Specific PCR fragment of 105 bp in the DRB1*0309 and 0323 alleles. Specific PCR fragment of 180 bp in the DRB1*0313 and 0338 and the 1339, 1354, 1377-1379, 1448 and 1464 alleles.

Primer mix 14: Specific PCR fragment of 105 bp in the DRB1*0316 allele. Specific PCR fragment of 135 bp in the DRB1*0312 and in the DRB1*0908 allele.

Primer mix 20: Specific PCR fragment of 75 bp in the DRB1*0304^{weakly}, 0325 and 0331 alleles. Specific PCR fragment of 120 bp in the DRB1*0328 allele.

Primer mix 23: Specific PCR fragment of 65 bp in the DRB1*0309 and 0323 alleles. Specific PCR fragment of 165 bp in the DRB1*0319 allele.

"w", may be weakly amplified.

'?', nucleotide sequence information not available for the primer matching sequence.

CELL LINE VALIDATION SHEET																
DRB1*03 SSP subtyping																
						Well										
						1	2	3	4	5	6	7	8	9	10	11
					Prod. No.:	200729601	200844102	200729603	200729604	200729605	200729606	200729607	200729608	200729609	200729610	200957111
						200729612	200729613	200729614	200729615	200729616	200729617	200729618	200729619	200729620	200729621	200729622
	IHWC cell line	DRB1														
1	9001 SA	*0101				-	-	-	-	-	-	-	-	-	-	-
2	9280 LK707	*1502	*0405			-	-	-	-	-	-	-	-	-	-	-
3	9011 E4181324	*1502				-	-	-	-	-	-	-	-	-	-	-
4	9275 GU373	*0301				+	-	-	+	-	+	-	-	+	-	-
5	9009 KAS011	*1601				-	-	-	-	-	-	-	-	-	-	-
6	9353 SM	*0407	*0803			-	-	-	-	-	-	-	-	-	-	-
7	9020 QBL	*0301				+	-	-	+	-	+	-	-	+	-	-
8	9007 DEM	*0401	*1602			-	-	-	-	-	-	-	-	-	-	-
9	9026 YAR	*0402				-	-	-	-	-	-	-	-	-	-	-
10	9107 LKT3	*0405				-	-	-	-	-	-	-	-	-	-	-
11	9051 PITOUT	*0701				-	-	-	-	-	-	-	-	-	-	-
12	9052 DBB	*0701				-	-	-	-	-	-	-	-	-	-	-
13	9004 JESTHOM	*0101				-	-	-	-	-	-	-	-	-	-	-
14	9071 OLGA	*0802				-	-	-	-	-	-	-	-	-	-	-
15	9075 DKB	*0901				-	-	-	-	-	-	-	-	-	-	-
16	9037 SWEIG007	*1101				-	-	-	-	-	-	-	-	-	-	-
17	9282 CTM3953540	*0301	*1301			+	-	-	+	-	+	-	-	+	-	-
18	9257 32367	*0901	*1101			-	-	-	-	-	-	-	-	-	-	-
19	9038 BM16	*1201				-	-	-	-	-	-	-	-	-	-	-
20	9059 SLE005	*1302				-	-	+	-	-	-	-	-	-	-	-
21	9064 AMALA	*1402				-	+	+	-	-	-	-	-	-	+	-
22	9056 KOSE	*1302	*1401			-	-	+	-	-	-	-	-	-	+	-
23	9124 IHL	*0803	*1414			-	-	-	-	-	-	-	-	-	+	-
24	9035 JBUSH	*1101				-	-	-	-	-	-	-	-	-	-	-
25	9049 IBW9	*0701				-	-	-	-	-	-	-	-	-	-	-
26	9285 WT49	*0301				+	-	-	+	-	+	-	-	+	-	-
27	9191 CH1007	*0405	*1001			-	-	-	-	-	-	-	-	-	-	-
28	9320 BEL5GB	*0416	*0701			-	-	-	-	-	-	-	-	-	-	-
29	9050 MOU	*0701				-	-	-	-	-	-	-	-	-	-	-
30	9021 RSH	*0302				-	+	+	-	-	-	-	+	-	+	-
31	9019 DUCAF	*0301				+	-	-	+	-	+	-	-	+	-	-
32	9297 HAG	*1303				-	-	-	-	-	-	-	-	-	+	-
33	9098 MT14B	*0404				-	-	-	-	-	-	-	-	-	-	-
34	9104 DHIF	*1101				-	-	-	-	-	-	-	-	-	-	-
35	9302 SSTO	*0403				-	-	-	-	-	-	-	-	-	-	-
36	9024 KT17	*0403	*0406			-	-	-	-	-	-	-	-	-	-	-
37	9065 HHKB	*1301				-	-	-	+	-	-	-	-	-	-	-
38	9099 LZL	*1402				-	+	+	-	-	-	-	-	-	+	-
39	9315 CML	*0301	*0401			+	-	-	+	-	+	-	-	+	-	-
40	9134 WHONP199	*0701	*0901			-	-	-	-	-	-	-	-	-	-	-
41	9055 H0301	*1302				-	-	+	-	-	-	-	-	-	-	-
42	9066 TAB089	*0803				-	-	-	-	-	-	-	-	-	-	-
43	9076 T7526	*0901				-	-	-	-	-	-	-	-	-	-	-
44	9057 TEM	*1401				-	-	-	-	-	-	-	-	-	+	-
45	9239 SHJO	*0701				-	-	-	+	-	-	-	-	-	-	-
46	9013 SCHU	*1501				-	-	-	-	-	-	-	-	-	-	-
47	9045 TUBO	*1104	*1201			-	-	-	-	-	-	-	-	-	-	-
48	9303 TER-ND	*0103				-	-	-	-	-	-	-	-	-	-	-

CELL LINE VALIDATION SHEET																		
DRB1*03 SSP subtyping																		
					Well													
					17	18	19	20	21	22	23	24	25	26	27	28	29	30
				Prod. No.:	200729617	200844118	200844119	200729620	200729621	200729622	200729623	200729624	200844125	200844126	200844127	200957128	200844129	200957130
	IHWC cell line		DRB1															
1	9001	SA	*0101		-	-	-	-	-	-	-	-	-	-	-	-	-	-
2	9280	LK707	*1502	*0405	-	-	-	-	-	-	-	-	-	-	-	-	-	-
3	9011	E4181324	*1502		-	-	-	-	-	-	-	-	-	-	-	-	-	-
4	9275	GU373	*0301		-	+	-	-	-	-	-	-	-	-	-	-	-	+
5	9009	KAS011	*1601		-	-	-	-	-	-	-	-	-	-	-	-	-	-
6	9353	SM	*0407	*0803	-	-	-	-	-	-	-	-	-	-	-	-	-	-
7	9020	QBL	*0301		-	+	-	-	-	-	-	-	-	-	-	-	-	+
8	9007	DEM	*0401	*1602	-	-	-	-	-	-	-	-	-	-	-	-	-	-
9	9026	YAR	*0402		-	-	-	-	-	-	-	-	-	-	-	-	-	-
10	9107	LKT3	*0405		-	-	-	-	-	-	-	-	-	-	-	-	-	-
11	9051	PITOUT	*0701		-	-	-	-	-	-	-	-	-	-	-	-	-	-
12	9052	DBB	*0701		-	-	-	-	-	-	-	-	-	-	-	-	-	-
13	9004	JESTHOM	*0101		-	-	-	-	-	-	-	-	-	-	-	-	-	-
14	9071	OLGA	*0802		-	-	-	-	-	-	-	-	-	-	-	-	-	-
15	9075	DKB	*0901		-	-	-	-	-	-	-	-	-	-	-	-	-	-
16	9037	SWEIG007	*1101		-	-	-	-	-	-	-	-	-	-	-	-	-	-
17	9282	CTM3953540	*0301	*1301	-	-	-	-	-	-	-	-	-	-	-	-	-	+
18	9257	32367	*0901	*1101	-	-	-	-	-	-	-	-	-	-	-	-	-	-
19	9038	BM16	*1201		-	-	-	-	-	-	-	-	-	-	-	-	-	-
20	9059	SLE005	*1302		-	-	-	-	-	-	-	-	-	-	-	-	-	+
21	9064	AMALA	*1402		-	-	-	-	-	-	-	+	-	-	-	-	-	+
22	9056	KOSE	*1302	*1401	-	-	-	-	-	-	-	-	-	-	-	-	-	+
23	9124	IHL	*0803	*1414	-	-	-	-	-	-	-	-	-	-	-	-	-	+
24	9035	JBUSH	*1101		-	-	-	-	-	-	-	-	-	-	-	-	-	-
25	9049	IBW9	*0701		-	-	-	-	-	-	-	-	-	-	-	-	-	-
26	9285	WT49	*0301		-	+	-	-	-	-	-	-	-	-	-	-	-	+
27	9191	CH1007	*0405	*1001	-	-	-	-	-	-	-	-	-	-	-	-	-	-
28	9320	BEL5GB	*0416	*0701	-	-	-	-	-	-	-	-	-	-	-	-	-	-
29	9050	MOU	*0701		-	-	-	-	-	-	-	-	-	-	-	-	-	-
30	9021	RSH	*0302		-	-	-	-	-	-	-	+	-	-	-	-	-	+
31	9019	DUCAF	*0301		-	+	-	-	-	-	-	-	-	-	-	-	-	+
32	9297	HAG	*1303		-	-	-	-	-	-	-	-	-	-	-	-	-	+
33	9098	MT14B	*0404		-	-	-	-	-	-	-	-	-	-	-	-	-	-
34	9104	DHIF	*1101		-	-	-	-	-	-	-	-	-	-	-	-	-	-
35	9302	SSTO	*0403		-	-	-	-	-	-	-	-	-	-	-	-	-	-
36	9024	KT17	*0403	*0406	-	-	-	-	-	-	-	-	-	-	-	-	-	-
37	9065	HHKB	*1301		-	-	-	-	-	-	-	-	-	-	-	-	-	+
38	9099	LZL	*1402		-	-	-	-	-	-	-	+	-	-	-	-	-	+
39	9315	CML	*0301	*0401	-	+	-	-	-	-	-	-	-	-	-	-	-	+
40	9134	WHONP199	*0701	*0901	-	-	-	-	-	-	-	-	-	-	-	-	-	-
41	9055	H0301	*1302		-	-	-	-	-	-	-	-	-	-	-	-	-	+
42	9066	TAB089	*0803		-	-	-	-	-	-	-	-	-	-	-	-	-	-
43	9076	T7526	*0901		-	-	-	-	-	-	-	-	-	-	-	-	-	-
44	9057	TEM	*1401		-	-	-	-	-	-	-	-	-	-	-	-	-	-
45	9239	SHJO	*0701		-	-	-	-	-	-	-	-	-	-	-	-	-	-
46	9013	SCHU	*1501		-	-	-	-	-	-	-	-	-	-	-	-	-	-
47	9045	TUBO	*1104	*1201	-	-	-	-	-	-	-	-	-	-	-	-	-	-
48	9303	TER-ND	*0103		-	-	-	-	-	-	-	-	-	-	-	-	-	-

CERTIFICATE OF ANALYSIS

Olerup SSP® DRB1*03 SSP

Product number: 101.113-24u/04u – without *Taq* pol.
Lot number: 65F
Expiry date: 2011-March-01
Number of tests: 24 tests – Product No. 101.113-24u
 4 tests – Product No. 101.113-04u
Number of wells per test: 30

Well specifications:

Well No.	Production No.	Well No.	Production No.	Well No.	Production No.
1	2007-296-01	13	2007-296-13	25	2008-441-25
2	2008-441-02	14	2007-296-14	26	2008-441-26
3	2007-296-03	15	2007-296-15	27	2008-441-27
4	2007-296-04	16	2007-296-16	28	2009-571-28
5	2007-296-05	17	2007-296-17	29	2008-441-29
6	2007-296-06	18	2008-441-18	30	2009-571-30
7	2007-296-07	19	2008-441-19		
8	2007-296-08	20	2007-296-20		
9	2007-296-09	21	2007-296-21		
10	2007-296-10	22	2007-296-22		
11	2009-571-11	23	2007-296-23		
12	2007-296-12	24	2007-296-24		

The specificity of each primer solution of the kit has been tested against 48 well characterized IHWC cell line DNAs.

No DNAs carrying the alleles to be amplified by primer solutions 7, 11 to 17, 19 to 23 and 25 to 29 were available. The specificities of the primers in primer solutions 7, 11 to 17, 21 and 26 were tested by separately adding one additional 5'-primer, respectively one additional 3'-primer. In primer solutions 19, 25, 28 and 29 it was only possible to test the 3'-primers, the 5'-primers were not possible to test. In primer solutions 20, 22, 23 and 27 it was only possible to test the 5'-primers, the 3'-primers were not possible to test. In primer solutions 11 and 14 one of the 3'-primer was not possible to test. In primer solution 21 one of the 5'-primers was not possible to test.

Results: No false positive or false negative amplifications were obtained.

Date of approval: 2009-May-26

Approved by:

Quality Control, Supervisor

Declaration of Conformity

Product name: *Olerup* SSP® DRB1*03
Product number: 101.113-24u/04u
Lot number: 65F
Intended use: DRB1*03 high resolution histocompatibility testing

Manufacturer: *Olerup* SSP AB
Hasselstigen 1
SE-133 33 Saltsjöbaden, Sweden
Phone: +46-8-717 88 27
Fax: +46-8-717 88 18

We, *Olerup* SSP AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) ISO 9001:2000 and ISO 13485:2003, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex II List B, conformity assessed using Annex IV, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at *Olerup* SSP AB, Hasselstigen 1, SE-133 33 Saltsjöbaden, Sweden.

The Authorized Representative located within the Community is: *Olerup* SSP AB.

Notified Body: Lloyd's Register Quality Assurance Limited, Hiramford, Middlemarch Office Village, Siskin Drive, Coventry CV3 4FJ, United Kingdom. (Notified Body number: 0088.)

Saltsjöbaden, Sweden
2009-May-26

Olle Olerup
Managing Director

ADDRESSES:**Manufacturer:****Olerup SSP AB**, Hasselstigen 1, SE-133 33 Saltsjöbaden, Sweden.**Tel:** +46-8-717 88 27**Fax:** +46-8-717 88 18**E-mail:** info-ssp@olerup.com**Web page:** <http://www.olerup.com>**Distributed by:****Olerup GmbH**, Löwengasse 47 / 6, AT-1030 Vienna, Austria.**Tel:** +43-1-710 15 00**Fax:** +43-1-710 15 00 10**E-mail:** support-at@olerup.com**Web page:** <http://www.olerup.com>**Olerup Inc.**, 901 S. Bolmar St., Suite R, West Chester, PA 19382**Tel:** 1-877-OLERUP1**Fax:** 610-344-7989**E-mail:** info.us@olerup.com**Web page:** <http://www.olerup.com>

For information on *Olerup* SSP distributors worldwide, contact **Olerup GmbH**.