

Olerup SSP[®] DRB3

Product number:	101.121-24/04 – including <i>Taq</i> pol.
Lot number:	71M
Expiry date:	2014-March-01
Number of tests:	24 tests – Product No. 101.121-24 4 tests – Product No. 101.121-04
Number of wells per test:	30
Storage - pre-aliquoted primers:	dark at -20°C
- PCR Master Mix:	-20°C
- Adhesive PCR seals	RT
- Product Insert	RT

This Product Description is only valid for Lot No. 71M.

CHANGES COMPARED TO THE PREVIOUS *OLERUP SSP*[®] DRB3 Lot

The DRB3 primer set as well as the specificity and interpretation tables have been updated for the DRB alleles described since the previous *Olerup SSP*[®] DRB3 lot (Lot No. 82K) was made.

Three wells have been added to the DRB3 kit,
wells **28** to **30**.

The primers of the wells detailed below have been exchanged, added or modified compared to the previous lot.

Well	5'-primer	3'-primer	rationale
1	-	Added	Increased yield of specific PCR product.
7	-	Added	3'-primer added for the DRB3*01:15 allele.
16	Moved	Moved	Primer pair moved to well 29.
25	-	Added	3'-primer added for the DRB3*02:26 allele.
28	New	New	New primer pair for the DRB1*02:27 allele.
29	Added	Added	Primer pair from well 16.
30	New	New	New primer pair for the DRB1*02:28 allele.

PRODUCT DESCRIPTION

DRB3 SSP subtyping

CONTENT

The primer set contains 5'- and 3'-primers for identifying the DRB3*01:01 to DRB3*03:03 alleles.

PLATE LAYOUT

Each test consists of 30 PCR reactions in a 32 well cut PCR plate. Wells 31 and 32 are empty.

1	2	3	4	5	6	7	8
9	10	11	12	13	14	15	16
17	18	19	20	21	22	23	24
25	26	27	28	29	30	empty	empty

The 32 well cut PCR plate is marked with 'DRB3' in silver/gray ink.

Well No. 1 is marked with the Lot No. '71M'.

A faint row of numbers is seen between wells 1 and 2 or wells 7 and 8 of the PCR trays. These stem from the manufacture of the trays, and should be disregarded.

The PCR plates are covered with a PCR-compatible foil.

Please note: When removing each 32 well PCR plate, make sure that the remaining plates stay covered. Use a scalpel or a similar instrument to carefully cut the foil between the plates.

INTERPRETATION

The interpretation of DRB3 subtypings will be influenced by thirteen DRB1*03 alleles, two DRB1*04, four DRB1*08 alleles, the DRB1*10, five DRB1*11 alleles, three DRB1*12, eleven DRB1*13 alleles and nine DRB1*14 alleles. Thus, the interpretation of DRB3 subtypings is only marginally influenced by other DRB genes.

UNIQUELY IDENTIFIED ALLELES

All the DRB3 alleles, i.e. **DRB3*01:01 to DRB3*01:15, DRB3*02:01 to DRB3*02:28 and DRB3*03:01 to DRB3*03:03**, recognized by the HLA Nomenclature Committee in July 2011¹ will give rise to unique amplification patterns by the primers in the DRB3 subtyping kit.

The DRB3 subtyping kit cannot distinguish the DRB3* 01:01:02:01 to 01:01:05 alleles, the DRB3* 02:02:01 to 02:02:05 alleles or the DRB3* 03:01:01 to 03:01:03 alleles.

¹DRB alleles listed on the IMGT/HLA web page 2011-July-14, release 3.5.0, www.ebi.ac.uk/imgt/hla.

RESOLUTION IN HOMO- AND HETEROZYGOTES

A total of 57 alleles generate 46 amplification patterns that can be combined in 1081 homozygous and heterozygous combinations. 308 of these genotypes do not give rise to unique amplification patterns. The different lengths of the specific PCR products were not considered in these calculations.

+++++---	----+---	---+-----	-----	*01:02, *01:03 = *01:02, *01:10
+++++--+	----+---	---+-----	-----	*01:02, *01:05 = *01:02, *01:15
+++++---	----+---	---+-----	-----	*01:01:02:01, *01:02 = *01:02, *01:02
+++-++--	----+---	---+-----	-----	*01:03, *01:04 = *01:04, *01:10
+++-+-+--	+----+--	---+-----	-----	*01:03, *01:07 = *01:07, *01:10
+++-+-+--	----+---	---+-----	-----	*01:03, *01:05 = *01:03, *01:15 = *01:05, *01:10 = *01:10, *01:15
+++-+---+	----+---	+---+-----	-----	*01:03, *01:08 = *01:08, *01:10
+++-+---+	----+---	---+-----	-----	*01:03, *01:06 = *01:06, *01:10
+++-+---	----+---	---+-----	-----	*01:03, *01:13 = *01:10, *01:13
+++-+---	----+---	---+-----	-----	*01:03, *01:11 = *01:10, *01:11
+++-+---	----+---	---+-----	+-----	*01:03, *01:12 = *01:10, *01:12
+++-+---	----+---	---+-----	-----	*01:01:02:01, *01:03 = *01:01:02:01, *01:10 = *01:03, *01:10 = *01:10, *01:10
+++--++-	----+---	---+-----	-----	*01:04, *01:05 = *01:04, *01:15
+++--++-	----+---	---+-----	-----	*01:01:02:01, *01:04 = *01:04, *01:04
+++--+++	----+---	+---+-----	-----	*01:05, *01:08 = *01:08, *01:15
+++--+++	----+---	---+-----	-----	*01:05, *01:06 = *01:06, *01:15
+++--++-	+----+--	---+-----	-----	*01:01:02:01, *01:07 = *01:05, *01:07 = *01:05, *02:05
+++--++-	---+---+--	---+-----+	-----	*01:05, *03:01:01 = *01:15, *03:01:01
+++--++-	----+---	---+-----+	-----	*01:05, *03:03 = *01:15, *03:03
+++--++-	----+---	---+-----	-----	*01:05, *01:13 = *01:13, *01:15
+++--++-	----+---	---+-----	-----	*01:05, *01:11 = *01:11, *01:15
+++--++-	----+---	---+-----	+-----	*01:05, *01:12 = *01:12, *01:15
+++--++-	----+---	---+-----	-+-----	*01:05, *01:14 = *01:14, *01:15
+++--++-	----+---	---+-----	-----	*01:01:02:01, *01:05 = *01:01:02:01, *01:15 = *01:05, *01:05 = *01:05, *01:15
+++--+++	+---+---+	+---+-----	-----	*01:06, *02:16 = *01:08, *02:16
+++--++-	+---+---+	+---+-----	-----	*01:06, *02:23 = *01:08, *02:08 = *01:08, *02:23
+++--++-	+---+---+	+---+-----	-----	*01:06, *02:06 = *01:08, *02:02:01:01 = *01:08, *02:06
+++--++-	+---+---+	+---+-----	-----	*01:06, *02:20 = *01:08, *02:20
+++--++-	----+---	+---+-----	-----	*01:01:02:01, *01:08 = *01:06, *01:08 = *01:08, *01:08
+++--++-	----+---	---+-----	-----	*01:01:02:01, *01:06 = *01:06, *01:06
+++--++-	+---+---+	---+---+	-----	*01:11, *02:11 = *01:13, *02:12
+++--++-	+---+---+	---+---+	+-----	*01:12, *02:11 = *01:13, *02:26
+++--++-	+---+---+	---+---+	-----	*01:01:02:01, *02:11 = *01:13, *02:02:01:01 = *01:13, *02:11
+++--++-	+---+---+	---+---+	+-----	*01:11, *02:26 = *01:12, *02:12
+++--++-	+---+---+	---+---+	-----	*01:01:02:01, *02:12 = *01:11, *02:02:01:01 = *01:11, *02:12

Lot No.: **71M**

Lot-specific Information

www.olerup-ssp.com

+++-----	++-++--	---+-----	+-----	*01:01:02:01, *02:26 = *01:12, *02:02:01:01 = *01:12, *02:26
+++-----	---+-----	---+-----	-----	*01:01:02:01, *01:13 = *01:13, *01:13
+++-----	---+-----	---+-----	-----	*01:01:02:01, *01:11 = *01:11, *01:11
+++-----	---+-----	---+-----	+-----	*01:01:02:01, *01:12 = *01:12, *01:12
++-----	+++-----	---+-----	-----	*02:19, *03:01:01 = *02:22, *03:01:01
++-----	+++-----	+-----	-----	*02:06, *03:01:01 = *02:16, *03:01:01
++-----	+++-----	-+-----	-----	*02:07, *03:01:01 = *02:21, *03:01:01
++-----	+++-----	---+-----	-----	*02:08, *03:01:01 = *02:18, *03:01:01
++-----	+++-----	---+-----	-----	*02:02:01:01, *03:01:01 = *02:25, *03:01:01
++-----	+++-----	---+-----	-----	*02:19, *03:03 = *02:22, *03:03
++-----	+++-----	+-----	-----	*02:06, *03:03 = *02:16, *03:03
++-----	+++-----	-+-----	-----	*02:07, *03:03 = *02:21, *03:03
++-----	+++-----	---+-----	-----	*02:08, *03:03 = *02:18, *03:03
++-----	+++-----	---+-----	-----	*02:02:01:01, *03:03 = *02:25, *03:03
+-----	+++-----	+-----	-----	*01:07, *02:16 = *01:15, *02:16
+-----	+++-----	+-----	-----	*01:07, *02:06 = *01:07, *02:20
+-----	+++-----	-+-----	-----	*01:07, *02:09 = *01:15, *02:21
+-----	+++-----	---+-----	-----	*01:07, *02:08 = *01:07, *02:14
+-----	+++-----	---+-----	-----	*01:07, *02:10 = *01:15, *02:02:01:01
+-----	+++-----	---+-----	-----	*01:07, *01:15 = *01:15, *02:05
+-----	+++-----	-----	-----	*01:07, *01:07 = *01:07, *02:05
+-----	+++-----	+-----	-----	*02:05, *02:06 = *02:05, *02:20
+-----	+++-----	---+-----	-----	*02:05, *02:08 = *02:05, *02:14
-----	+++++-----	-+-----	-----	*01:09, *02:04 = *01:09, *02:24
-----	+++++-----	-----+-----	-----	*02:04, *02:19 = *02:04, *02:25 = *02:19, *02:24
-----	+++++-----	-----	-----	*02:02:01:01, *02:04 = *02:04, *02:22 = *02:22, *02:24
-----	+++++-----	-----	-----	*02:04, *02:04 = *02:04, *02:24
-----	+++++-----	-----	---+-----	*02:01, *02:01 = *02:01, *02:24
-----	+++++-----	---+-----	-----	*02:19, *03:02 = *02:22, *03:02
-----	+++++-----	+-----	-----	*02:06, *03:02 = *02:16, *03:02
-----	+++++-----	-+-----	-----	*02:07, *03:02 = *02:21, *03:02
-----	+++++-----	---+-----	-----	*02:08, *03:02 = *02:18, *03:02
-----	+++++-----	---+-----	-----	*02:02:01:01, *03:02 = *02:25, *03:02
-----	+++++-----	+-----	-----	*02:06, *02:17 = *02:17, *02:20
-----	+++++-----	---+-----	-----	*02:08, *02:17 = *02:14, *02:17
-----	+++++-----	-----	-----	*02:02:01:01, *02:17 = *02:03, *02:17 = *02:17, *02:17
-----	+++++-----	++-----	-----	*02:09, *02:16 = *02:16, *02:21
-----	+++++-----	+-----	-----	*02:08, *02:16 = *02:14, *02:16 = *02:16, *02:23
-----	+++++-----	+-----	-----	*02:02:01:01, *02:16 = *02:03, *02:16 = *02:06, *02:16 = *02:10, *02:16 = *02:16, *02:16 = *02:16, *02:20
-----	+++++-----	++-----	-----	*01:09, *02:06 = *01:09, *02:20
-----	+++++-----	+-----	-----	*02:06, *02:19 = *02:19, *02:20
-----	+++++-----	+-----	-----	*02:06, *02:22 = *02:20, *02:22
-----	+++++-----	++++-----	-----	*01:09, *02:08 = *01:09, *02:14
-----	+++++-----	-+-----	-----	*01:09, *02:19 = *01:09, *02:21 = *01:09, *02:25 = *02:09, *02:19 = *02:09, *02:22

Lot No.: **71M**

Lot-specific Information

www.olerup-ssp.com

-----	++---++-	-++-----	-----	*01:09, *02:02:01:01 = *01:09, *02:07 = *01:09, *02:22
-----	++---++-	-++-----	-----	*02:07, *02:19 = *02:19, *02:21 = *02:21, *02:22
-----	++---++-	-++-----	-----	*02:08, *02:19 = *02:14, *02:19 = *02:18, *02:19 = *02:18, *02:22
-----	++---++-	-++-----	-----	*02:08, *02:22 = *02:14, *02:22
-----	++---++-	-++-----	-----	*02:02:01:01, *02:19 = *02:19, *02:19 = *02:19, *02:22 = *02:19, *02:25 = *02:22, *02:25
-----	++---++-	-++-----	-----	*02:02:01:01, *02:22 = *02:22, *02:22
-----	++---++-	++---++-	-----	*02:06, *02:09 = *02:09, *02:20
-----	++---++-	++---++-	-----	*02:06, *02:21 = *02:20, *02:21
-----	++---++-	++---++-	-----	*02:06, *02:07 = *02:07, *02:20
-----	++---++-	++---++-	-----	*02:06, *02:18 = *02:18, *02:20 = *02:18, *02:23 = *02:23, *02:25
-----	++---++-	++---++-	-----	*02:02:01:01, *02:23 = *02:06, *02:08 = *02:06, *02:14 = *02:06, *02:23 = *02:08, *02:20 = *02:08, *02:23 = *02:14, *02:23 = *02:20, *02:23 = *02:23, *02:23
-----	++---++-	++---++-	-----	*02:06, *02:10 = *02:10, *02:20
-----	++---++-	++---++-	-----	*02:06, *02:11 = *02:11, *02:20
-----	++---++-	++---++-	-----	*02:06, *02:12 = *02:12, *02:20
-----	++---++-	++---++-	-----	*02:06, *02:13 = *02:13, *02:20
-----	++---++-	++---++-	-----	*02:06, *02:25 = *02:20, *02:25
-----	++---++-	++---++-	-----	*02:06, *02:26 = *02:20, *02:26
-----	++---++-	++---++-	-----	*02:06, *02:27 = *02:20, *02:27
-----	++---++-	++---++-	-----	*02:06, *02:15 = *02:15, *02:20
-----	++---++-	++---++-	-----	*02:06, *02:28 = *02:20, *02:28
-----	++---++-	++---++-	-----	*02:02:01:01, *02:06 = *02:02:01:01, *02:20 = *02:06, *02:06 = *02:06, *02:20
-----	++---++-	++---++-	-----	*02:08, *02:09 = *02:09, *02:14 = *02:09, *02:18
-----	++---++-	++---++-	-----	*02:07, *02:18 = *02:08, *02:21 = *02:14, *02:21 = *02:18, *02:21
-----	++---++-	++---++-	-----	*02:07, *02:08 = *02:07, *02:14
-----	++---++-	++---++-	-----	*02:02:01:01, *02:09 = *02:07, *02:09 = *02:09, *02:21 = *02:09, *02:25 = *02:10, *02:21
-----	++---++-	++---++-	-----	*02:02:01:01, *02:21 = *02:07, *02:21 = *02:07, *02:25 = *02:21, *02:21 = *02:21, *02:25
-----	++---++-	++---++-	-----	*02:02:01:01, *02:07 = *02:07, *02:07
-----	++---++-	++---++-	-----	*02:08, *02:10 = *02:10, *02:14
-----	++---++-	++---++-	-----	*02:08, *02:11 = *02:11, *02:14
-----	++---++-	++---++-	-----	*02:08, *02:12 = *02:12, *02:14
-----	++---++-	++---++-	-----	*02:08, *02:13 = *02:13, *02:14
-----	++---++-	++---++-	-----	*02:02:01:01, *02:18 = *02:08, *02:18 = *02:08, *02:25 = *02:14, *02:18 = *02:14, *02:25 = *02:18, *02:18 = *02:18, *02:25
-----	++---++-	++---++-	-----	*02:08, *02:26 = *02:14, *02:26
-----	++---++-	++---++-	-----	*02:08, *02:27 = *02:14, *02:27
-----	++---++-	++---++-	-----	*02:08, *02:15 = *02:14, *02:15
-----	++---++-	++---++-	-----	*02:08, *02:28 = *02:14, *02:28
-----	++---++-	++---++-	-----	*02:02:01:01, *02:08 = *02:02:01:01, *02:14 =

Lot No.: **71M**

Lot-specific Information

www.olerup-ssp.com

-----	++--+-	-----	-----	*02:08, *02:08 = *02:08, *02:14
-----	++--+-	-----	-----	*02:02:01:01, *02:11 = *02:11, *02:11
-----	++--+-	-----	-----	*02:02:01:01, *02:12 = *02:12, *02:12
-----	++--+-	-----	-----	*02:02:01:01, *02:13 = *02:13, *02:13
-----	++--+-	-----	-----	*02:02:01:01, *02:25 = *02:25, *02:25
-----	++--+-	-----	-----	*02:02:01:01, *02:26 = *02:26, *02:26
-----	++--+-	-----	-----	*02:02:01:01, *02:28 = *02:28, *02:28
-----	++--+-	-----	-----	*02:15, *02:27 = *02:27, *02:27
-----	++--+-	-----	-----	*01:09, *01:09 = *01:09, *02:10
-----	++--+-	-----	-----	*02:09, *02:09 = *02:09, *02:10

*01:01:02:01 = *01:01:02:01-01:01:05

*02:02:01 = *02:02:01-02:02:05

*03:01:01 = *03:01:01-03:01:03

SPECIFICITY TABLE

DRB3 SSP subtyping

Specificities and sizes of the PCR products of the 30 primer mixes used for DRB3 SSP subtyping

Primer Mix	Size of spec. PCR product ¹	Size of control band ²	Amplified DRB3 alleles ³	Amplified DRB1 ⁴ alleles
1⁵	100 bp	515 bp	*01:01:02:01-01:08, 01:10-01:13, 01:15, 02:05, 03:01:01-03:01:03, 03:03	*03:42, 11:30 ^w , 13:67, 14:46
2⁵	125 bp	430 bp	*01:01:02:01-01:06, 01:08, 01:10-01:14, 03:01:01-03:01:03, 03:03	*03:38, 04:82 ^w , 04:99 ^w , 08:07 ^w , 08:19 ^w , 08:25 ^w , 08:34 ^w , 12:27 ^w , 13:31 ^w , 13:46 ^w , 13:54 ^w , 13:77 ^w , 13:100 ^w , 14:48
3⁵	95 bp	430 bp	*01:01:02:01-01:02, 01:04-01:08, 01:10-01:13, 01:15	*03:42
4	265 bp	430 bp	*01:02	
5^{5,8}	95 bp, 125 bp	430 bp	*01:03, 01:10	*14:46
6	190 bp	430 bp	*01:04	
7^{5,9}	90 bp, 180 bp	515 bp	*01:05, 01:07, 01:15	*03:42, 14:46
8⁵	120 bp	430 bp	*01:06, 01:08	*03:42, 14:46
9⁶	165 bp	430 bp	*01:07, 02:01-02:08, 02:11-02:28	
10⁶	185 bp	515 bp	*01:09, 02:01-02:04, 02:06-02:14, 02:16-02:26, 02:28, 03:02	*12:22 ^w
11	270 bp	515 bp	*02:01, 02:04, 02:24, 03:01:01-03:02	*03:42
12^{5,10}	90 bp, 145 bp	515 bp	*02:01, 02:04, 02:17, 02:24	*03:37, 11:43, 11:50, 12:34, 14:59, 14:86, 14:96
13⁶	270 bp	515 bp	*01:01:02:01-01:13, 01:15, 02:02:01-01-02:03, 02:05-02:13, 02:15-02:23, 02:25-02:28, 03:03	*11:30, 13:67, 14:46

Lot No.: **71M**

Lot-specific Information

www.olerup-ssp.com

14	185 bp	515 bp	*01:07, 01:09, 02:02:01:01- 02:02:05, 02:05- 02:13, 02:15-02:19, 02:21-02:23, 02:25- 02:28	*10:01:01 ^W -10:01:03 ^W , 10:03 ^W
15^{5,11}	120 bp, 180 bp, 210 bp	430 bp	*02:03, 02:16-02:17, 03:01:01-03:03	
16	150 bp	430 bp	*01:09, 02:04, 02:19, 02:22	*03:44
17^{5,12}	120 bp, 180 bp	430 bp	*01:08, 02:06, 02:16, 02:20, 02:23	*03:42, 13:67, 14:46
18⁵	100 bp	430 bp	*01:09, 02:07, 02:09, 02:21	
19^{7,13}	180 bp, 270 bp	515 bp	*02:08, 02:14, 02:18, 02:23	*11:30
20¹⁴	165 bp, 180 bp	430 bp	*01:01:02:01-01:06, 01:08-01:13, 01:15, 02:09-02:10, 02:16, 03:01:01-03:03	*03:42, 11:30, 13:67, 14:46
21^{5,15}	120 bp, 210 bp	430 bp	*01:13, 02:11	
22^{7,16}	185 bp, 245 bp	430 bp	*01:11, 02:12	
23⁷	195 bp	430 bp	*02:13	
24	185 bp	430 bp	*02:09, 02:18-02:19, 02:21, 02:25, 03:01:01-03:03	*11:30
25¹⁷	225 bp, 260 bp	430 bp	*01:12, 02:26	
26	235 bp	430 bp	*01:14	*03:01:02, 03:02:02, 03:05:03, 03:11:02, 03:13:02, 03:17, 03:24, 03:27, 03:35, 11:27:01, 11:84, 12:22, 13:33:02- 13:33:03, 13:61:01, 13:96:02, 14:38, 14:47, 14:50, 14:98
27⁵	90 bp	430 bp	*02:01	
28	235 bp	430 bp	*02:27	*11:30, 13:67, 14:46
29	160 bp	430 bp	*02:15, 02:27	*10:01:01 ^W -10:03 ^W , 13:106
30	210 bp	430 bp	*02:28	

¹Alleles are assigned by the presence of specific PCR product(s). However, the sizes of the specific PCR products may be helpful in the interpretation of DRB3 SSP typings. When the primers in a primer mix can give rise to specific PCR products of more than one length this is indicated if the size difference is 20 base pairs or more. Size differences shorter

than 20 base pairs are not given. For high resolution SSP kits the respective lengths of the specific PCR product(s) of the alleles amplified by these primer mixes are given.

Nonspecific amplifications, i.e. a ladder or a smear of bands, may sometimes be seen. GC-rich primers have a higher tendency of giving rise to nonspecific amplifications than other primers.

PCR fragments longer than the control bands may sometimes be observed. Such bands should be disregarded and do not influence the interpretation of the SSP typings.

PCR fragments migrating faster than the control bands, but slower than a 400 bp fragment may be seen in some gel read-outs. Such bands can be disregarded and do not influence the interpretation of the SSP typings.

Some primers may give rise to primer oligomer artifacts. Sometimes this phenomenon is an inherent feature of the primer pair(s) of a primer mix. More often it is due to other factors such as too low amount of DNA in the PCR reactions, taking too long time in setting up the PCR reactions, working at elevated room temperature or using thermal cyclers that are not pre-heated.

²The internal positive control primer pairs amplify segments of the human growth hormone gene. The two different control primer pairs give rise to either an internal positive control band of 430 base pairs, for most wells, or a band of 515 base pairs, for some wells.

Well number 1 contains the primer pair giving rise to the longer, 515 bp, internal positive control band in order to help in the correct orientation of the DRB3 subtyping.

In addition, wells number 7, 10 to 14 and 19 contain the primer pair giving rise to the longer, 515 bp, internal positive control band in order to allow kit identification.

In the presence of a specific amplification the intensity of the control band often decreases.

³For several DRB alleles only partial second exon nucleotide sequences are available. In these instances it is not known whether some of the primers of the SSP sets are completely matched with the target sequences or not. We assume that unknown sequences in the first hyperpolymorphic region of the second exon of DRB alleles are conserved within allelic groups and that unknown sequences of codons 87 to 92 are identical with the DRB1*0101 consensus sequence.

⁴Due to the sharing of sequence motifs between DRB3 and DRB1 alleles, primer mixes 1 to 3, 5, 7, 8, 10 to 14, 16, 17, 19, 20, 24, 26, 28 and 29 will amplify DRB1 alleles.

⁵Short specific PCR fragments are less intense and not as sharp as longer specific bands.

⁶Primer mixes 9, 10 and 13 have a tendency to giving rise to nonspecific amplifications.

⁷Primer mixes 19, 22 and 23 may give rise to primer oligomer formation.

⁸Primer mix 5: Specific PCR fragment of 95 bp in the DRB3*01:03 and the DRB1*14:46 alleles. Specific PCR fragment of 125 bp in the DRB3*01:10 allele.

⁹Primer mix 7: Specific PCR fragment of 90 bp in the DRB3*01:05 allele. Specific PCR fragment of 180 bp in the DRB3*01:07 and 01:15 and in the DRB1*03:42 and DRB1*14:46 alleles.

¹⁰Primer mix 12: Specific PCR fragment of 90 bp in the DRB3*02:17 allele. Specific PCR fragment of 145 bp in the DRB3*02:01, 02:04 and 02:24 and in the DRB1*03:37, 11:43, 11:50, 12:34, 14:59, 14:86 and 14:96 alleles.

¹¹Primer mix 15: Specific PCR fragment of 120 bp DRB3*02:03 and 03:01:01-03:03 alleles. Specific PCR fragment of 180 bp in the DRB3*02:16 allele. Specific PCR fragment of 210 bp in the DRB3*02:17 allele.

¹²Primer mix 17: Specific PCR fragment of 120 bp in the DRB3*01:08, 02:06 and 02:20 and in the DRB1*03:42, 13:67 and 14:46 alleles. Specific PCR fragment of 180 bp in the DRB3*02:16 and 02:23 alleles.

¹³Primer mix 19: Specific PCR fragment of 180 bp in the DRB3*02:08, 02:18 and 02:23 and the DRB1*11:30 alleles. Specific PCR fragment of 270 bp in the DRB3*02:14 allele.

¹⁴Primer mix 20: Specific PCR fragment of 165 bp in the DRB3*01:01:02:01-01:06, 01:08-01:13, 01:15, 02:09-02:10 and 03:01:01-03:03 and the DRB1*03:42, 11:30, 13:67 and 14:46 alleles. Specific PCR fragment of 180 bp in the DRB3*02:16 allele.

¹⁵Primer mix 21: Specific PCR fragment of 120 bp in the DRB3*01:13 allele. Specific PCR fragment of 210 bp in the DRB3*02:11 allele.

¹⁶Primer mix 22: Specific PCR fragment of 185 bp in the DRB3*01:11 allele. Specific PCR fragment of 245 bp in the DRB3*02:12 allele.

¹⁷Primer mix 25: Specific PCR fragment of 225 bp in the DRB3*02:26 allele. Specific PCR fragment of 260 bp in the DRB3*01:12 allele.

‘w’, might be weakly amplified.

INTERPRETATION TABLE																
DRB3 SSP subtyping																
Amplification patterns of the DRB3*01:01 to *03:03 alleles																
	Well ⁵															
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16
Length of spec. PCR product	100	125	95	265	95	190	90	120	165	185	270	90	270	185	120	150
					125		180					145			180	
															210	
Length of int. pos. control ¹	515	430	430	430	430	430	515	430	430	515	515	515	515	515	430	430
5'-primer(s) ²	10(116)	30(175)	11(119)	11(118)	11(119)	8(110)	11(119)	11(119)	10(116)	30(175)	10(116)	51(239)	10(116)	38(200)	11(119)	38(200)
	5'-gCT 3'	5'-gAT 3'	5'-gCg 3'	5'-TgT 3'	5'-gCg 3'	5'-CTC 3'	5'-gCg 3'	5'-gCg 3'	5'-gCT 3'	5'-gAC 3'	5'-gCT 3'	5'-gAg 3'	5'-gCT 3'	5'-CgC 3'	5'-gCT 3'	5'-CgC 3'
			11(119)						10(116)				10(116)	38(200)		
			5'-gTg 3'						5'-gCT 3'				5'-gCT 3'	5'-CgC 3'		
3'-primer(s) ³	30(175)	57(257)	28(171)	86(344)	28(171)	57(257)	28(169)	38(199)	51(239)	77(317)	86(344)	67(286)	86(344)	86(344)	37(197)	74(308)
	5'-gTA 3'	5'-CgA 3'	5'-CTg 3'	5'-CAC 3'	5'-CTC 3'	5'-CgA 3'	5'-gTT 3'	5'-CAC 3'	5'-CCC 3'	5'-AgT 3'	5'-CCA 3'	5'-gAA 3'	5'-CAC 3'	5'-CAC 3'	5'-Cgg 3'	5'-CCC 3'
	30(175)	57(257)	28(171)		39(202)		57(257)			77(318)		86(344)			37(197)	
	5'-gTA 3'	5'-CAA 3'	5'-ATg 3'		5'-gCT 3'		5'-CAT 3'			5'-TAA 3'		5'-CCA 3'			5'-CgA 3'	
															57(257)	
															5'-CAG 3'	
															67(286)	
															5'-gAA 3'	
Well No.	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16

INTERPRETATION TABLE														
DRB3 SSP subtyping														
Amplification patterns of the DRB3*01:01 to *03:03 alleles														
Well ⁵														
17	18	19	20	21	22	23	24	25	26	27	28	29	30	
120	100	180	165	120	185	195	185	225	235	90	235	160	210	Length of spec. PCR product
180		270	180	210	245			260						
430	430	515	430	430	430	430	430	430	430	430	430	430	430	Length of int. pos. control ¹
10(116)	38(200)	10(116)	10(116)	10(116)	9(112)	26(163)	11(119)	10(116)	12(122)	164(577)	10(116)	38(200)	123(455)	5'-primer(s) ²
5' -gCT 3'	5' -CgC 3'	5' -gCT 3'	5' -gCT 3'	5' -gCT 3'	5' -TgC 3'	5' -ggC 3'	5' -gCT 3'	5' -gCT 3'	5' -TAC 3'	5' -CAT 3'	5' -gCT 3'	5' -CgC 3'	5' -CTg 3'	
					5' -CAT 3' 29(173)									
37(196)	57(257)	55(251)	51(239)	37(196)	77(318)	77(318)	58(260)	71(299)	77(317)	179(624)	74(308)	77(317)	179(624)	3'-primer(s) ³
5' -gTT 3'	5' -CgA 3'	5' -gCA 3'	5' -CCg 3'	5' -gAg 3'	5' -TAA 3'	5' -TAA 3'	5' -CCT 3'	5' -gCC 3'	5' -AAT 3'	5' -ACg 3'	5' -CCg 3'	5' -Agg 3'	5' -ACg 3'	
37(196)		57(256)	57(257)	67(286)			60(266)	84(337)						
5' -gTT 3'		5' -gCT 3'	5' -CAg 3'	5' -gAT 3'			5' -Agg 3'	5' -CCg 3'						
55(251)		58(260)												
5' -gCA 3'		5' -CCT 3'												
57(257)		86(344)												
5' -CAg 3'		5' -CAg 3'												
17	18	19	20	21	22	23	24	25	26	27	28	29	30	Well No.

Lot No.: **71M**

Lot-specific Information

www.olerup-ssp.com

Length of spec. PCR product	100	125	95	265	95	190	90	120	165	185	270	90	270	185	120	150
					125		180					145			180	
															210	
Well No.	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16
DRB3 allele ⁴																
*01:01:02:01-01:01:05	1	2	3										13			
*01:02	1	2	3	4									13			
*01:03	1	2			5								13			
*01:04	1	2	3			6							13			
*01:05	1	2	3				7						13			
*01:06	1	2	3					8					13			
*01:07	1		3				7		9				13	14		
*01:08	1	2	3					8					13			
*01:09										10			13	14		16
*01:10	1	2	3		5								13			
*01:11	1	2	3										13			
*01:12	1	2	3										13			
*01:13	1	2	3										13			
*01:14		2														
*01:15	1		3				7						13			
*02:01									9	10	11	12				
*02:02:01:01-02:02:05									9	10			13	14		
*02:03									9	10			13		15	
*02:04									9	10	11	12				16
*02:05	1								9				13	14		
*02:06									9	10			13	14		
*02:07									9	10			13	14		
*02:08									9	10			13	14		
*02:09										10			13	14		
*02:10										10			13	14		
*02:11									9	10			13	14		
*02:12									9	10			13	14		
*02:13									9	10			13	14		
*02:14									9	10						
*02:15									9				13	14		
*02:16									9	10			13	14	15	
*02:17									9	10		12	13	14	15	
*02:18									9	10			13	14		
*02:19									9	10			13	14		16
*02:20									9	10			13			
*02:21									9	10			13	14		
*02:22									9	10			13	14		16
Well No.	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16

Lot No.: **71M**

Lot-specific Information

www.olerup-ssp.com

120	100	180	165	120	185	195	185	225	235	90	235	160	210	Length of spec. PCR product
180		270	180	210	245			260						
17	18	19	20	21	22	23	24	25	26	27	28	29	30	Well No. DRB3 allele ⁴
			20											*01:01:02:01- 01:01:05
			20											*01:02
			20											*01:03
			20											*01:04
			20											*01:05
			20											*01:06
														*01:07
17			20											*01:08
	18		20											*01:09
			20											*01:10
			20		22									*01:11
			20					25						*01:12
			20	21										*01:13
									26					*01:14
			20											*01:15
									27					*02:01
														*02:02:01:01- 02:02:05
														*02:03
														*02:04
17														*02:05
	18													*02:06
		19												*02:07
														*02:08
	18		20				24							*02:09
			20											*02:10
				21										*02:11
					22									*02:12
						23								*02:13
		19												*02:14
												29		*02:15
17			20											*02:16
														*02:17
		19					24							*02:18
							24							*02:19
17														*02:20
	18						24							*02:21
														*02:22
17	18	19	20	21	22	23	24	25	26	27	28	29	30	Well No.

Lot No.: **71M**

Lot-specific Information

www.olerup-ssp.com

Length of spec.	100	125	95	265	95	190	90	120	165	185	270	90	270	185	120	150
PCR product					125		180					145			180	
															210	
Well No.	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16
*02:23									9	10			13	14		
*02:24									9	10	11	12				
*02:25									9	10			13	14		
*02:26									9	10			13	14		
*02:27									9				13	14		
*02:28									9	10			13	14		
*03:01:01-03:01:03	1	2									11				15	
*03:02										10	11				15	
*03:03	1	2											13		15	
<i>DRB1</i> *03:01:02, 03:02:02, 03:05:03, 03:11:02, 03:13:02, 03:17, 03:24, 03:35, 11:27:01, 11:84, 13:33:02-13:33:03, 13:61:01, 13:96:02, 14:38, 14:47, 14:50, 14:98																
<i>DRB1</i> *03:37, 11:43, 11:50, 12:34, 14:59, 4:86, 14:96												12				
<i>DRB1</i> *03:38, 14:48		2														
<i>DRB1</i> *03:42	1		3				7	8			11					
<i>DRB1</i> *03:44																16
<i>DRB1</i> *04:82,04:99, 08:07, 08:19, 08:25, 08:34, 12:27, 13:31, 13:46, 13:54, 13:77, 13:100		w														
<i>DRB1</i> *10:01:01- 10:01:03, 10:03														w		
<i>DRB1</i> *10:02																
<i>DRB1</i> *11:30	w												13			
<i>DRB1</i> *12:22										w						
<i>DRB1</i> *13:67	1												13			
<i>DRB1</i> *13:106																
<i>DRB1</i> *14:46	1				5		7	8					13			
Well No.	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16

120	100	180	165	120	185	195	185	225	235	90	235	160	210	Length of spec. PCR product
180		270	180	210	245			260						
17	18	19	20	21	22	23	24	25	26	27	28	29	30	Well No.
17		19												*02:23
														*02:24
							24							*02:25
								25						*02:26
											28	29		*02:27
												30		*02:28
			20				24							*03:01:01-03:01:03
			20				24							*03:02
			20				24							*03:03
								26						DRB1*03:01:02, 03:02:02, 03:05:03, 03:11:02, 03:13:02, 03:17, 03:24, 03:35, 11:27:01, 11:84, 13:33:02-13:33:03, 13:61:01, 13:96:02, 14:38, 14:47, 14:50, 14:98
														DRB1*03:37, 11:43, 11:50, 12:34, 14:59, 4:86, 14:96
														DRB1*03:38, 14:48
17			20											DRB1*03:42
														DRB1*03:44
														DRB1*04:82,04:99, 08:07, 08:19, 08:25, 08:34, 12:27, 13:31, 13:46, 13:54, 13:77, 13:100
												w		DRB1*10:01:01- 10:01:03, 10:03
												w		DRB1*10:02
		19	20				24				28			DRB1*11:30
								26						DRB1*12:22
17			20								28			DRB1*13:67
												29		DRB1*13:106
17			20								28			DRB1*14:46
17	18	19	20	21	22	23	24	25	26	27	28	29	30	Well No.

Lot No.: 71M

Lot-specific Information

www.olerup-ssp.com

¹The internal positive control primer pairs amplify segments of the human growth hormone gene. The two different control primer pairs give rise to either an internal positive control band of 430 base pairs, for most wells, or a band of 515 base pairs, for some wells.

Well number 1 contains the primer pair giving rise to the longer, 515 bp, internal positive control band in order to help in the correct orientation of the DRB3subtyping.

In addition, wells number 7, 10 to 14 and 19 contain the primer pair giving rise to the longer, 515 bp, internal positive control band in order to allow kit identification.

In the presence of a specific amplification the intensity of the control band often decreases.

²The codon, and in parenthesis the nucleotide, in the 2nd and 3rd exon, matching the specificity-determining 3'-end of the primer is given. Codon and nucleotide numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.

³The codon, and in parenthesis the nucleotide, in the 2nd and 3rd exon, matching the specificity-determining 3'-end of the primer is given in the anti-sense direction. Codon and nucleotide numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.

⁴DRB3*010101 allele has been shown to be identical to DRB3*01:01:02:01

⁵Primer mix 5: Specific PCR fragment of 95 bp in the DRB3*01:03 and the DRB1*14:46 alleles. Specific PCR fragment of 125 bp in the DRB3*01:10 allele.

Primer mix 7: Specific PCR fragment of 90 bp in the DRB3*01:05 allele. Specific PCR fragment of 180 bp in the DRB3*01:07 and 01:15 and in the DRB1*03:42 and DRB1*14:46 alleles.

Primer mix 12: Specific PCR fragment of 90 bp in the DRB3*02:17 allele. Specific PCR fragment of 145 bp in the DRB3*02:01, 02:04 and 02:24 and in the DRB1*03:37, 11:43, 11:50, 12:34, 14:59, 14:86 and 14:96 alleles.

Primer mix 15: Specific PCR fragment of 120 bp DRB3*02:03 and 03:01:01-03:03 alleles. Specific PCR fragment of 180 bp in the DRB3*02:16 allele. Specific PCR fragment of 210 bp in the DRB3*02:17 allele.

Primer mix 17: Specific PCR fragment of 120 bp in the DRB3*01:08, 02:06 and 02:20 and in the DRB1*03:42, 13:67 and 14:46 alleles. Specific PCR fragment of 180 bp in the DRB3*02:16 and 02:23 alleles.

Primer mix 19: Specific PCR fragment of 180 bp in the DRB3*02:08, 02:18 and 02:23 and the DRB1*11:30 alleles. Specific PCR fragment of 270 bp in the DRB3*02:14 allele.

Primer mix 20: Specific PCR fragment of 165 bp in the DRB3*01:01:02:01-01:06, 01:08-01:13, 01:15, 02:09-02:10 and 03:01:01-03:03 and the DRB1*03:42, 11:30, 13:67 and 14:46 alleles. Specific PCR fragment of 180 bp in the DRB3*02:16 allele.

Primer mix 21: Specific PCR fragment of 120 bp in the DRB3*01:13 allele. Specific PCR fragment of 210 bp in the DRB3*02:11 allele.

Primer mix 22: Specific PCR fragment of 185 bp in the DRB3*01:11 allele. Specific PCR fragment of 245 bp in the DRB3*02:12 allele.

Primer mix 25: Specific PCR fragment of 225 bp in the DRB3*02:26 allele. Specific PCR fragment of 260 bp in the DRB3*01:12 allele.

'w', might be weakly amplified.

CELL LINE VALIDATION SHEET																					
DRB3 SSP subtyping kit																					
					Prod. No.	Well															
						1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16
						201190801	201080502	201080503	201080504	201080505	201080506	201190807	201080508	201080509	201080510	201080511	201080512	201080513	201080514	201080515	201190816
	IHWC cell line	DRB3																			
1	9001 SA					-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
2	9280 LK707					-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
3	9011 E4181324					-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
4	9275 GU373	*02:02				-	-	-	-	-	-	-	-	+	+	-	-	+	+	-	-
5	9009 KAS011					-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
6	9353 SM					-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
7	9020 QBL	*02:02				-	-	-	-	-	-	-	-	+	+	-	-	+	+	-	-
8	9025 DEU					-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
9	9026 YAR					-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
10	9107 LKT3					-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
11	9051 PITOUT					-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
12	9052 DBB					-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
13	9004 JESTHOM					-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
14	9071 OLGA					-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
15	9075 DKB					-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
16	9037 SWEIG007	*02:02				-	-	-	-	-	-	-	-	+	+	-	-	+	+	-	-
17	9282 CTM3953540	*01:01				+	+	+	-	-	-	-	-	-	-	-	-	+	-	-	-
18	9257 32367	*02:02				-	-	-	-	-	-	-	-	+	+	-	-	+	+	-	-
19	9038 BM16	*02:02				-	-	-	-	-	-	-	-	+	+	-	-	+	+	-	-
20	9059 SLE005	*03:01				+	+	-	-	-	-	-	-	-	-	+	-	-	-	+	-
21	9064 AMALA	*01:01				+	+	+	-	-	-	-	-	-	-	-	-	+	-	-	-
22	9056 KOSE	*02:02	*03:01			+	+	-	-	-	-	-	-	+	+	+	-	+	+	+	-
23	9124 IHL	*02:02				-	-	-	-	-	-	-	-	+	+	-	-	+	+	-	-
24	9035 JBUSH	*02:02				-	-	-	-	-	-	-	-	+	+	-	-	+	+	-	-
25	9049 IBW9					-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
26	9285 WT49	*02:02				-	-	-	-	-	-	-	-	+	+	-	-	+	+	-	-
27	9191 CH1007 ¹					-	-	-	-	-	-	-	-	-	-	-	-	-	w	-	-
28	9320 BEL5GB					-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
29	9050 MOU					-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
30	9021 RSH	*01:01				+	+	+	-	-	-	-	-	-	-	-	-	+	-	-	-
31	9019 DUCAF	*02:02				-	-	-	-	-	-	-	-	+	+	-	-	+	+	-	-
32	9297 HAG	*01:01				+	+	+	-	-	-	-	-	-	-	-	-	+	-	-	-
33	9098 MT14B					-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
34	9104 DHIF	*02:02				-	-	-	-	-	-	-	-	+	+	-	-	+	+	-	-
35	9302 SSTO					-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
36	9024 KT17					-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
37	9065 HHKB	*01:01				+	+	+	-	-	-	-	-	-	-	-	-	+	-	-	-
38	9099 LZL	*01:01				+	+	+	-	-	-	-	-	-	-	-	-	+	-	-	-
39	9315 CML	*01:01				+	+	+	-	-	-	-	-	-	-	-	-	+	-	-	-
40	9134 WHONP199					-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
41	9055 H0301	*03:01				+	+	-	-	-	-	-	-	-	-	+	-	-	-	+	-
42	9066 TAB089					-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
43	9076 T7526					-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
44	9057 TEM	*02:01				-	-	-	-	-	-	-	-	+	+	+	+	-	-	-	-
45	9239 SHJO					-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
46	9013 SCHU					-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
47	9045 TUBO	*02:02				-	-	-	-	-	-	-	-	+	+	-	-	+	+	-	-
48	9303 TER-ND					-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-

CELL LINE VALIDATION SHEET																														
DRB3 SSP subtyping kit																														
					Well																									
					17	18	19	20	21	22	23	24	25	26	27	28	29	30												
				Prod. No.	201080517	201080518	201080519	201080520	201080521	201190822	201190823	201080524	201190825	201080526	201080527	201190828	201190829	201190830												
	IHC cell line		DRB3																											
1	9001	SA			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-											
2	9280	LK707			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-											
3	9011	E4181324			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-											
4	9275	GU373	*02:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-											
5	9009	KAS011			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-											
6	9353	SM			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-											
7	9020	QBL	*02:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-											
8	9025	DEU			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-											
9	9026	YAR			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-											
10	9107	LKT3			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-											
11	9051	PITOUT			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-											
12	9052	DBB			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-											
13	9004	JESTHOM			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-											
14	9071	OLGA			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-											
15	9075	DKB			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-											
16	9037	SWEIG007	*02:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-											
17	9282	CTM3953540	*01:01		-	-	-	+	-	-	-	-	-	-	-	-	-	-	-											
18	9257	32367	*02:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-											
19	9038	BM16	*02:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-											
20	9059	SLE005	*03:01		-	-	-	+	-	-	-	+	-	-	-	-	-	-	-											
21	9064	AMALA	*01:01		-	-	-	+	-	-	-	-	-	-	-	-	-	-	-											
22	9056	KOSE	*02:02	*03:01	-	-	-	+	-	-	-	+	-	-	-	-	-	-	-											
23	9124	IHL	*02:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-											
24	9035	JBUSH	*02:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-											
25	9049	IBW9			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-											
26	9285	WT49	*02:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-											
27	9191	CH1007			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-											
28	9320	BEL5GB			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-											
29	9050	MOU			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-											
30	9021	RSH	*01:01		-	-	-	+	-	-	-	-	-	-	-	-	-	-	-											
31	9019	DUCAF	*02:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-											
32	9297	HAG	*01:01		-	-	-	+	-	-	-	-	-	-	-	-	-	-	-											
33	9098	MT14B			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-											
34	9104	DHIF	*02:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-											
35	9302	SSTO			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-											
36	9024	KT17			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-											
37	9065	HHKB	*01:01		-	-	-	+	-	-	-	-	-	-	-	-	-	-	-											
38	9099	LZL	*01:01		-	-	-	+	-	-	-	-	-	-	-	-	-	-	-											
39	9315	CML	*01:01		-	-	-	+	-	-	-	-	-	-	-	-	-	-	-											
40	9134	WHONP199			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-											
41	9055	H0301	*03:01		-	-	-	+	-	-	-	+	-	-	-	-	-	-	-											
42	9066	TAB089			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-											
43	9076	T7526			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-											
44	9057	TEM	*02:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-											
45	9239	SHJO			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-											
46	9013	SCHU			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-											
47	9045	TUBO	*02:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-											
48	9303	TER-ND			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-											

CERTIFICATE OF ANALYSIS

Olerup SSP® DRB3 SSP

Product number: 101.121-24/04 – including *Taq* pol.
Lot number: 71M
Expiry date: 2014-March-01
Number of tests: 24 tests – Product No. 101.121-24
4 tests – Product No. 101.121-04
Number of wells per test: 30

Well specifications:

Well No.	Production No.	Well No.	Production No.	Well No.	Production No.
1	2011-908-01	11	2010-805-11	21	2010-805-21
2	2010-805-02	12	2010-805-12	22	2011-908-22
3	2010-805-03	13	2010-805-13	23	2011-908-23
4	2010-805-04	14	2010-805-14	24	2010-805-24
5	2010-805-05	15	2010-805-15	25	2011-908-25
6	2010-805-06	16	2011-908-16	26	2010-805-26
7	2011-908-07	17	2010-805-17	27	2010-805-27
8	2010-805-08	18	2010-805-18	28	2011-908-28
9	2010-805-09	19	2010-805-19	29	2011-908-29
10	2010-805-10	20	2010-805-20	30	2011-908-30

The specificity of each primer solution of the kit has been tested against 48 well characterized cell line DNAs.

No DNAs carrying the alleles to be amplified by primer solutions 4 to 8, 16 to 19, 21 to 23 and 25 to 30 were available. The specificities of the primers in primer solutions 5, 7, 8, 16 to 19, 21, 25, 26, 28 and 29 were tested by separately adding one additional 5'-primer, respectively one additional 3'-primer. In primer solutions 4, 6, 22, 23, 27 and 30 it was only possible to test the 3'-primer, the 5'-primer was not possible to test. One or two of the 3'-primers in primer solutions 1 to 3, 5, 7, 12, 15, 17, 19, 21 and 25 were not possible to test. In primer solutions 2, 3, 4, 6, 22, 23, 27 and 30 one or two of the 5'-primers was not possible to test. Additional 3'-primers in primer solutions 15, 20 and 24 were tested by separately adding one 5'-primer.

Results: No false positive or false negative amplifications were obtained.

Date of approval: 2011- October-11

Approved by:

Quality Control, Supervisor

Declaration of Conformity

Product name: *Olerup* SSP® DRB3
Product number: 101.121-24/04
Lot number: 71M

Intended use: DRB3 high resolution histocompatibility testing

Manufacturer: *Olerup* SSP AB
Franzengatan 5
SE-112 51 Stockholm, Sweden
Phone: +46-8-717 88 27
Fax: +46-8-717 88 18

We, *Olerup* SSP AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) ISO 9001:2008 and ISO 13485:2003, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex II List B, conformity assessed using Annex IV, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at *Olerup* SSP AB, Franzengatan 5, SE-112 51 Stockholm, Sweden.

Notified Body: Lloyd's Register Quality Assurance Limited, Hiramford, Middlemarch Office Village, Siskin Drive, Coventry CV3 4FJ, United Kingdom.
(Notified Body number: 0088.)

Stockholm, Sweden
2011-October-11

Ann-Cathrin Jareman
Head of QA and Regulatory Affairs

Lot No.: **71M**

Lot-specific Information

www.olerup-ssp.com

ADDRESSES:

Manufacturer:

Olerup SSP AB, Franzengatan 5, SE-112 51 Stockholm, Sweden.

Tel: +46-8-717 88 27

Fax: +46-8-717 88 18

E-mail: info-ssp@olerup.com

Web page: <http://www.olerup-ssp.com>

Distributed by:

Olerup GmbH, Löwengasse 47 / 6, AT-1030 Vienna, Austria.

Tel: +43-1-710 15 00

Fax: +43-1-710 15 00 10

E-mail: support-at@olerup.com

Web page: <http://www.olerup.com>

Olerup Inc., 901 S. Bolmar St., Suite R, West Chester, PA 19382

Tel: 1-877-OLERUP1

Fax: 610-344-7989

E-mail: info.us@olerup.com

Web page: <http://www.olerup.com>

For information on *Olerup* SSP distributors worldwide, contact **Olerup GmbH**.