

101.572-12 – including *Taq* pol., IFU-01
101.572-12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **5E0**

Lot-specific information

Olerup SSP® HLA-B*57:01

Product number:	101.572-12 – including <i>Taq</i> polymerase 101.572-12u – without <i>Taq</i> polymerase
Lot number:	5E0
Expiry date:	2019-05-01
Number of tests:	12
Number of wells per test:	22+1
Storage - pre-aliquoted primers:	dark at -20°C
- PCR Master Mix:	-20°C
- Adhesive PCR seals	RT
- Product Insert	RT

This Product Description is only valid for Lot No. 5E0.

Complete product documentation consists of generic Instructions for Use (IFU), lot specific Product Insert, Worksheet and Certificate.

CHANGES COMPARED TO THE PREVIOUS OLERUP SSP® HLA-B*57:01 Lot (0D4).

Positive and negative control DNAs are included in the kit.

Two wells have been added to HLA-B*57:01, wells **22 and 23**.

The HLA-B*57:01 specificity and interpretation tables have been updated for the HLA-B alleles described since the previous *Olerup SSP®* HLA-B*57:01 lot was made (**Lot No. 0D4**). The kit design is based on IMGT/HLA database 3.25.0.

As of lot series V, the Specificity Table is included in the lot-specific Product Insert, and the Interpretation Table is included in the Worksheet.

101.572-12 – including *Taq* pol., IFU-01

101.572-12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **5E0**

Lot-specific information

The primers of the wells detailed below have been exchanged, added or modified compared to the previous lot.

Well	5'-primer	3'-primer	rationale
12	Moved	Moved	Primer pairs moved to well 21 for decreased tendency of primer oligomer formation.
18	Added	Added	Primer pair added for the B*57:81 allele.
21	Added	Added	Negative Control moved to well 23, primer pairs added from well 12.
22	New	New	New primer pair added for the B*57:67:02 allele.
23	-	-	Negative Control added from well 21.

Change in revision R01 compared to R00:

1. Primer mix 4 does not amplify the B*07:120 and the B*40:150 alleles. This has been corrected in the Specificity and Interpretation Tables.

101.572-12 – including *Taq* pol., IFU-01
101.572-12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **5E0**

Lot-specific information

Well **23** contains Negative Control primer pairs, that will amplify more than 95% of the Olerup SSP® HLA Class I, DRB, DQB1, DPB1 and DQA1 amplicons as well as all the amplicons generated by the control primer pairs matching the human growth hormone gene.

HLA-specific PCR product sizes range from 75 to 200 base pairs.
The PCR product generated by the positive control primer pair is 430 base pairs.

Length of PCR product	105	200	105	80	75	80	85
5'-primer¹	164	340	440	45	45	43	36
	5'-CAC ^{3'}	5'-Agg ^{3'}	5'-TTA ^{3'}	5'-Tgg ^{3'}	5'-Tgg ^{3'}	5'-Tgg ^{3'}	5'-TAC ^{3'}
							36
							5'-TAT ^{3'}
3'-primer²	231	2nd I	507	59	58	57	47
	5'-TgC ^{3'}	5'-AAA ^{3'}	5'-TTg ^{3'}	5'-CTC ^{3'}	5'-ggC ^{3'}	5'-CTC ^{3'}	5'-ACA ^{3'}
							48
							5'-gCA ^{3'}
							48
							5'-gCC ^{3'}
							52
							5'-TgT ^{3'}
A*	+	+	+				
B*	+	+	+				
C*	+	+	+				
DRB1				+	+		
DRB3				+	+		
DRB5				+			
DQB1					+		
DPB1						+	
DQA1							+

¹The nucleotide position for HLA class I genes and the codon for HLA class II genes, in the 2nd or 3rd exon, matching the specificity-determining 3'-end of the primer is given. Nucleotide and codon numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.

²The nucleotide position for HLA class I genes and the codon for HLA class II genes, in the 2nd or 3rd exon or the 2nd intron, matching the specificity-determining 3'-end of the primer is given in the anti-sense direction. Nucleotide and codon numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.

101.572-12 – including *Taq* pol., IFU-01
101.572-12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **5E0**

Lot-specific information

PRODUCT DESCRIPTION

HLA-B*57:01 SSP subtyping

CONTENT

The primer set contains 5'- and 3'-primers for identifying the B*57:01:01 to 57:01:22 alleles.

The primer solutions consist of specific primer mixes, i.e. group-specific primers as well as a **control primer pair** matching non-allelic sequences.

Positive and negative control DNAs are included in the kit:

DNA 1; a B*57-positive DNA as a positive control, **IHW 9052, DBB, B*57:01**.

DNA 2; a B*58-positive DNA as a negative control, **IHW 9285, WT49, B*58:01**.

(A B*58:01-positive DNA was chosen as negative control, as this is most similar to the B*57 group of alleles in the primer matching regions.)

The kit contains enough control DNAs to perform 1 test set up of positive and negative control. If more than 1 test set up per kit are run other positive and negative DNA samples can be used as controls (e.g. positive and negative samples from previous tests).

Legal notice: The DNA in this kit is included under licensing agreement between Olerup SSP AB and Public Health England.

PLATE LAYOUT

Each test consists of 23 PCR reactions in a 24 well cut PCR plate. Well 24 is empty.

1	2	3	4	5	6	7	8
9	10	11	12	13	14	15	16
17	18	19	20	21	22	NC	empty

The 24 well cut PCR plate is marked with 'B*57:01' in silver/gray ink.

Well No. 1 is marked with the Lot No. '5E0'.

Wells 1 to 22 – HLA-B*57:01 primers.

Well 23 – Negative Control (NC).

A faint row of numbers is seen between wells 1 and 2 or wells 7 and 8 of the PCR trays. These stem from the manufacture of the trays, and should be disregarded.

The PCR plates are heat-sealed with a PCR-compatible foil.

Please note: When removing each 24 well PCR plate, make sure that the remaining plates stay sealed. Use a scalpel or a similar instrument to carefully cut the foil between the plates.

INTERPRETATION

Due to the sharing of sequence motifs between HLA-B alleles some non-HLA-B*57 alleles will be amplified by primer mixes 1 to 18, 20 and 22. In addition, a few HLA-A and HLA-C alleles will be amplified by primer mixes 5, 7, 10, 11, 13 to 17 and 22.

For further details see Specificity Table.

101.572-12 – including *Taq* pol., IFU-01

101.572-12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: 5E0

Lot-specific information

UNIQUELY IDENTIFIED ALLELES

HLA-B*57:01 will give rise to a unique amplification pattern by the primers in the HLA-B*57:01 kit^{1,2}.

The HLA-B*57:01 typing kit cannot distinguish the silent mutations in the B*57:01:01 to B*57:01:22 alleles.

¹HLA-B alleles listed on the IMGT/HLA web page 2016-July-14, release 3.25.0, www.ebi.ac.uk/imgt/hla.

²Alleles that have been deleted from or renamed in the official WHO HLA Nomenclature up to and including the last IMGT/HLA database release can be retrieved from web page <http://hla.alleles.org/alleles/deleted.html>.

101.572-12 – including *Taq* pol., IFU-01

101.572-12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **5E0**

Lot-specific information

SPECIFICITY TABLE

HLA-B*57:01 SSP subtyping

Specificities and sizes of the PCR products of the 22+1 primer mixes used for HLA-B*57:01 SSP subtyping

Primer Mix	Size of spec. PCR product ¹	Size of control band ²	Amplified HLA-B*57:01 alleles	Other amplified HLA-B alleles ^{3,4}
1⁵	90 bp	800 bp	*57:01:01-57:01:22	*57:02:01-57:04:01, 57:05-57:15, 57:17-57:19, 57:21-57:35, 57:37-57:44, 57:46-57:50, 57:52-57:61, 57:63-57:68, 57:70-57:83, 58:36
2⁶	220 bp	800 bp	*57:01:01-57:01:22	*57:03:01-57:03:02, 57:06-57:08, 57:10, 57:14:01-57:18, 57:20-57:23, 57:25-57:27, 57:29, 57:31-57:41, 57:43-57:51, 57:53-57:60, 57:62, 57:64-57:81, 35:208, 35:317, 40:30, 40:34, 44:153, 55:14, 58:14
3⁵	95 bp 170 bp 215 bp	800 bp		*57:04:01-57:04:02, 57:32, 44:153 *57:06, 57:18 *57:27
4⁵	100 bp	1070 bp		*57:15, 57:20, 57:29, 57:62, 15:214, 18:81, 35:250, 51:165
5⁵	90 bp 165 bp 200 bp 245 bp	800 bp		*57:16, 57:34, 58:54 *57:23 *57:46, 57:64 *57:07, 57:26, 44:153, 55:14, C*06:72
6⁵	90 bp	1070 bp		*57:02:01-57:03:02, 57:07, 57:09, 57:12, 57:17, 57:39, 57:42, 57:46, 57:57, 57:63, 57:66, 57:70, 57:80, 40:30, 40:34 *57:08, 57:65
7⁵	205 bp 100 bp 140 bp 185 bp 215 bp 250 bp	1070 bp		*57:53 *57:13, 57:31, 07:227, 40:30, 40:34, C*06:72 *57:50 *57:09, 57:24 *57:13, 57:14:01-57:14:02, 57:25, 07:227, 40:30, 40:34
8	185 bp	1070 bp		*57:10, 57:39, 57:44, 57:51, 57:73, 07:219, 58:41
9⁵	110 bp 150 bp	800 bp		*57:33 *57:21, 57:40, 57:74, 14:20, 35:127, 51:186, 53:41
10	205 bp 240 bp	1070 bp		*57:13, 57:22, 57:57, 57:63, 07:227, 55:14, C*06:72^w *57:43
11⁵	100 bp	1070 bp	*57:01:01-57:01:22	*57:06, 57:08, 57:10, 57:13-57:16, 57:18-57:27, 57:29-57:31,

101.572-12 – including *Taq* pol., IFU-01

101.572-12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **5E0**

Lot-specific information

			57:33-57:38, 57:40 ^w , 57:41, 57:43-57:45, 57:47-57:52, 57:54-57:56, 57:58-57:62, 57:64-57:65, 57:67:01-57:69, 57:71-57:75, 57:77-57:79N, 57:81, 35:208, 35:317, 55:14, 58:14, C*06:72
12⁵	75 bp	1070 bp	*57:35
	140 bp		*57:08, 57:60, 15:87
13⁵	105 bp	1070 bp	*57:45, 57:51, 57:69, 57:71, A*02:285
	155 bp		*57:58, 57:80, 07:219, 58:12, A*02:42, A*02:310
	200 bp		*57:49
14⁵	100 bp	1070 bp	*57:59, 15:116, 15:352, 40:63, 40:92, 40:324, 44:169, 44:182, 46:43, C*03:129, C*03:234, C*17:07
	125 bp		*57:47, 57:54, 15:33, 15:248, 49:22, 51:126, C*03:87:01-03:87:02, C*05:27, C*05:39, C*08:115
	205 bp		*57:77, 44:36, C*03:232
15	145 bp	1070 bp	*57:56, 57:78, 15:340, 35:300, 40:218, 46:28, 51:58
	185 bp		*57:48, A*23:72, A*33:12
	295 bp		*57:55
16⁵	65 bp	1070 bp	*57:67:01, 58:14, C*06:72
17⁵	90 bp	1070 bp	*57:04:01-57:04:02, 57:41, 44:153, C*06:72
	180 bp		*57:68
18	170 bp	1070 bp	*57:37, 57:79N
	275 bp		*57:81, 44:13, 44:67, 51:170
19⁵	100 bp	1070 bp	*57:76
	215 bp		*57:72
20	155 bp	1070 bp	*57:75, 40:327
21⁵	100 bp	1070 bp	*57:36
	135 bp		*57:38
22⁵	125 bp	1070 bp	*57:05, 57:67:01-57:67:02, 07:235, 13:01:01-13:04, 13:06-13:09, 13:11-13:20, 13:22:01-13:23, 13:25-13:34, 13:36-13:61, 13:63N-13:64, 13:66-13:70, 13:72-13:78, 13:80-13:93, 14:22, 14:37, 15:16:01:01-15:16:03, 15:20, 15:34, 15:42, 15:62, 15:67, 15:85-15:86, 15:95, 15:137, 15:194, 15:222, 15:242:02, 15:254, 15:303, 15:362, 15:393, 18:22, 18:52, 18:105, 27:14, 27:107, 35:01:01:01-35:01:42, 35:01:44-35:05:02, 35:06-35:08:08, 35:10-35:11:03, 35:13-35:15:02, 35:19-35:21, 35:23-35:30, 35:33-35:38, 35:40N-35:42:02, 35:45-35:50, 35:52, 35:54-35:58, 35:60,

101.572-12 – including *Taq* pol., IFU-01
101.572-12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **5E0**

Lot-specific information

	35:62-35:63, 35:65Q-35:66, 35:68:01-35:72, 35:74-35:78, 35:80-35:85, 35:89-35:98, 35:100- 35:101:02, 35:103-35:112, 35:115- 35:116, 35:119-35:124, 35:126- 35:134N, 35:136-35:148, 35:150:01-35:163, 35:166- 35:173N, 35:175-35:178, 35:180- 35:184, 35:186-35:195, 35:197- 35:198, 35:200, 35:202-35:204, 35:206-35:207, 35:209-35:214, 35:216N-35:229, 35:231-35:232, 35:235-35:241, 35:243-35:250, 35:252-35:259, 35:261-35:264, 35:267-35:275, 35:277-35:316, 35:318-35:320, 35:322, 40:28, 44:10, 44:15, 44:18, 44:48, 45:01:01-45:19, 46:11, 46:18, 46:33, 46:53, 46:67, 48:02:01- 48:02:03, 48:25, 49:01:01-49:14, 49:16-49:39, 49:41-49:43, 50:01:01, 50:01:03-50:02, 50:05- 50:19, 50:31-50:47, 51:15, 51:56:01-51:56:03, 51:157, 51:158:02, 53:01:01-53:06, 53:08:01-53:43, 54:01:01- 54:01:05, 54:01:07-54:05N, 54:07- 54:08N, 54:10, 54:12-54:35, 55:01:01-55:01:05, 55:01:07- 55:03, 55:05, 55:07, 55:09-55:13, 55:15-55:16, 55:18-55:19, 55:21- 55:37, 55:39-55:48, 55:50, 55:52- 55:71, 55:73-55:78, 56:01:01-01- 56:01:03, 56:01:05-56:02:02, 56:07-56:11, 56:13-56:17, 56:19N, 56:22-56:26, 56:28N-56:30, 56:33- 56:47, 56:50-56:51, 58:01:01-01- 58:01:02, 58:01:04-58:02:02, 58:04-58:07, 58:09, 58:11-58:19, 58:21-58:26, 58:28:01-58:29, 58:31N-58:45:02, 58:47-58:58, 58:60, 58:62-58:81, 59:01:01-01- 59:01:01:02, 59:03-59:09, 82:01- 82:03, C*01:32:01-01:32:02, C*02:56, C*04:180:01-04:180:02, C*06:20, C*07:81, C*07:168, C*07:450, C*08:123, C*15:126, C*16:98			
23 ⁷	-	-	-	Negative Control

¹Alleles are assigned by the presence of specific PCR product(s). However, the sizes of the specific PCR products may be helpful in the interpretation of HLA-B*57:01 SSP typings. When the primers in a primer mix can give rise to HLA-specific PCR products of more than one length this is indicated if the size difference is more than 20 base pairs. Size differences of 20 base pairs or less are not given. For high resolution SSP kits, the alleles listed are specified according to amplicon length.

101.572-12 – including Taq pol., IFU-01
101.572-12u – without Taq pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: 5E0

Lot-specific information

Nonspecific amplifications, i.e. a ladder or a smear of bands, may sometimes be seen. GC-rich primers have a higher tendency of giving rise to nonspecific amplifications than other primers.

PCR fragments longer than the control bands may sometimes be observed. Such bands should be disregarded and do not influence the interpretation of the SSP typings.

PCR fragments migrating faster than the control bands, but slower than a 400 bp fragment may be seen in some gel read-outs. Such bands can be disregarded and do not influence the interpretation of the SSP typings.

Some primers may give rise to primer oligomer artifacts. Sometimes this phenomenon is an inherent feature of the primer pair(s) of a primer mix. More often it is due to other factors such as too low amount of DNA in the PCR reactions, taking too long time in setting up the PCR reactions, working at elevated room temperature or using thermal cyclers that are not pre-heated.

²The internal positive control primer pairs amplify segments of the human growth hormone gene. The internal positive control bands are 1070 or 800 base pairs respectively, well distribution as outlined in the table. Well number 1 contains the shorter, 800 bp, internal positive control band. The well distribution of the internal controls can help in orientation of the kit on gel photo, as well as allow for kit identification. In the presence of a specific amplification the intensity of the control band often decreases.

³For several HLA Class I alleles 1st and/or 4th exon(s) and beyond, as well as intron nucleotide sequences, are not available. In these instances it is not known whether some of the primers of the SSP sets are completely matched with the target sequences or not. Assumption is made that unknown sequences in these regions are conserved within allelic groups.

⁴Due to the sharing of sequence motifs between HLA-B alleles some non-HLA-B*57 alleles will be amplified by primer mixes 1 to 18, 20 and 22. In addition, a few HLA-A and HLA-C alleles will be amplified by primer mixes 5, 7, 10, 11, 13 to 17 and 22.

⁵HLA-specific PCR products shorter than 125 base pairs have a lower intensity and are less sharp than longer PCR products.

⁶Primer mix 2 may give rise to a lower yield of HLA-specific PCR product than the other B*57:01 primer mixes.

⁷Primer mix 23 contains a negative control, which will amplify more than 95% of HLA amplicons as well as the amplicons generated by the control primer pairs matching the human growth hormone gene. HLA-specific PCR product sizes range from 75 to 200 base pairs and the PCR product generated by the HGH positive control primer pair is 430 base pairs.

'w', may be weakly amplified.

101.572-12 – including *Taq* pol., IFU-01
101.572-12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **5E0**

Lot-specific information

PRIMER SPECIFICATION

Well No.	1	2	3	4	5	6	7	8	9	10	11	12
Length of spec.	90	220	95	100	90	90	100	185	110	205	100	75
PCR product			170		165	205	140		150	240		140
			215		200		185					
					245		215					
							250					
Length of int. pos. control ¹	800	800	800	1070	800	1070	1070	1070	800	1070	1070	1070
5'-primer(s) ²	209	362	362	209	130	320	362	97	349	362	362	209
	5'-ggC 3'	5'-ggT 3'	5'-ggT 3'	5'-ggC 3'	5'-AgT 3'	5'-CCC 3'	5'-ggT 3'	5'-TCC 3'	5'-CTT 3'	5'-ggT 3'	5'-ggT 3'	5'-ggC 3'
				704	200	362		103	352			
				5'-TgT 3'	5'-TCg 3'	5'-ggT 3'		5'-CCT 3'	5'-ACg 3'			
					209			122	353			
					5'-ggA 3'			5'-CCT 3'	5'-CAA 3'			
					362				392			
					5'-ggT 3'				5'-CgA 3'			
3'-primer(s) ³	256	539	418	259	256	2 nd I	421	256	463	527	419	244
	5'-CCC 3'	5'-TCA 3'	5'-gTC 3'	5'-CTT 3'	5'-CCC 3'	5'-TCg 3'	5'-ggT 3'	5'-CCC 3'	5'-gCT 3'	5'-CCT 3'	5'-Cgg 3'	5'-CTT 3'
			481	263	523	412	463			559	419	302
			5'-gTA 3'	5'-gTT 3'	5'-ACA 3'	5'-gTT 3'	5'-gCg 3'			5'-Cgg 3'	5'-CAg 3'	5'-ggC 3'
			500	271	559	521	505					320
			5'-ggA 3'	5'-CAC 3'	5'-CgT 3'	5'-ggA 3'	5'-gCC 3'					5'-Agg 3'
			537	774	572		538					
			5'-Agg 3'	5'-ggT 3'	5'-gCg 3'		5'-gTC 3'					
							559					
							5'-CTC 3'					
							583					
							5'-gTg 3'					
Well No.	1	2	3	4	5	6	7	8	9	10	11	12

Well No.	13	14	15	16	17	18	19	20	21	22
Length of spec.	105	100	145	65	90	170	100	155	100	125
PCR product	155	125	185		180	275	215		135	
	200	205	295							
Length of int. pos. control ¹	1070	1070	1070	1070	1070	1070	1070	1070	1070	1070
5'-primer(s) ²	97	395	106	362	362	255	362	106	209	379
	5'-TCg 3'	5'-gCT 3'	5'-CCA 3'	5'-ggT 3'	5'-ggT 3'	5'-gAg 3'	5'-ggT 3'	5'-CCA 3'	5'-ggC 3'	5'-ACC 3'
	142	467	395			877				
	5'-TCT 3'	5'-CTg 3'	5'-gCC 3'			5'-AC 3'				
	193	485	757			878				
	5'-CCA 3'	5'-CAA 3'	5'-CCA 3'			5'-gCA 3'				
		499								
		5'-TCT 3'								
3'-primer(s) ³	256	559	212	387	409	2 nd I	419	220	268	463
	5'-CCC 3'	5'-CAg 3'	5'-ggT 3'	5'-TCg 3'	5'-ATA 3'	5'-TCg 3'	5'-CgA 3'	5'-TgT 3'	5'-gTg 3'	5'-gCT 3'
			214		502	916	534		302	
			5'-CCA 3'		5'-CTT 3'	5'-gAC 3'	5'-Tgg 3'		5'-ggg 3'	
			539							
			5'-TCA 3'							
			916							
			5'-gAC 3'							
Well No.	13	14	15	16	17	18	19	20	21	22

101.572-12 – including *Taq* pol., IFU-01

101.572-12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: 5E0

Lot-specific information

¹The internal positive control primer pairs amplify segments of the human growth hormone gene. The internal positive control bands are 1070 or 800 base pairs respectively, well distribution as outlined in the table. Well number 1 contains the shorter, 800 bp, internal positive control band. The well distribution of the internal controls can help in orientation of the kit on gel photo, as well as allow for kit identification. In the presence of a specific amplification the intensity of the control band often decreases.

²The nucleotide position matching the specificity-determining 3'-end of the primer is given. Nucleotide numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.

³The nucleotide position matching the specificity-determining 3'-end of the primer is given in the anti-sense direction. Nucleotide numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.

101.572-12 – including *Taq* pol., IFU-01

101.572-12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **5E0**

Lot-specific information

CELL LINE VALIDATION SHEET																				
HLA-B*57:01 SSP typing kit ²																				
					Well															
				Prod. No.:	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16
					201327601	201327602	201327603	201327604	201327605	201327606	201441007	201558608	201558609	201673510	201327611	201673512	201327613	201558614	201558615	201441016
	IHWC cell line ¹		B*																	
1	9001	SA	*07:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
2	9280	LK707	*52:01	*73:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
3	9011	E4181324	*52:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
4	9275	GU373	*15:10	*53:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
5	9009	KAS011	*37:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
6	9353	SM	*39:01	*51:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
7	9020	QBL	*18:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
8	9025	DEU	*35:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
9	9026	YAR	*38:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
10	9107	LKT3	*54:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
11	9051	PITOUT	*44:03		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
12	9052	DBB	*57:01		+	+	-	-	-	-	-	-	-	-	+	-	-	-	-	-
13	9025	JESTHOM	*27:05		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
14	9071	OLGA	*15:01	*15:20	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
15	9075	DKB	*40:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
16	9037	SWEIG007	*40:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
17	9282	CTM3953540	*08:01	*55:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
18	9257	32367	*14:01	*56:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
19	9038	BM16	*18:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
20	9059	SLE005	*40:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
21	9064	AMALA	*15:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
22	9056	KOSE	*35:03		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
23	9124	IHL	*40:02	*56:02	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
24	9035	JBUSH	*38:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
25	9049	IBW9	*14:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
26	9285	WT49	*58:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
27	9191	CH1007	*07:05	*51:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
28	9320	BEL5GB	*44:02	*44:03	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
29	9050	MOU	*44:03		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
30	9021	RSH	*42:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
31	9019	DUCAF	*18:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
32	9297	HAG	*41:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
33	9098	MT14B	*40:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
34	9104	DHIF	*38:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
35	9302	SSTO	*44:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
36	9024	KT17	*15:01	*35:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
37	9065	HHKB	*07:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
38	9099	LZL	*15:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
39	9315	CML	*08:01	*27:05	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
40	9134	WHONP199	*13:02	*46:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
41	9055	H0301	*14:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
42	9066	TAB089	*46:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
43	9076	T7526	*46:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
44	9057	TEM	*38:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
45	9239	SHJO	*42:01	*50:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
46	9013	SCHU	*07:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
47	9045	TUBO	*51:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
48	9303	TER-ND	*35:01	*44:03	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-

101.572-12 – including *Taq* pol., IFU-01

101.572-12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **5E0**

Lot-specific information

CELL LINE VALIDATION SHEET											
HLA-B*57:01 SSP typing kit ²											
					Prod. No.:	Well					
						17	18	19	20	21	22
						201558617	201673518	201558619	201558620	201673521	201673522
	IHWC cell line ¹			B*							
1	9001	SA	*07:02			-	-	-	-	-	-
2	9280	LK707	*52:01	*73:01		-	-	-	-	-	-
3	9011	E4181324	*52:01			-	-	-	-	-	-
4	9275	GU373	*15:10	*53:01		-	-	-	-	-	+
5	9009	KAS011	*37:01			-	-	-	-	-	-
6	9353	SM	*39:01	*51:01		-	-	-	-	-	-
7	9020	QBL	*18:01			-	-	-	-	-	-
8	9025	DEU	*35:01			-	-	-	-	-	+
9	9026	YAR	*38:01			-	-	-	-	-	-
10	9107	LKT3	*54:01			-	-	-	-	-	+
11	9051	PITOUT	*44:03			-	-	-	-	-	-
12	9052	DBB	*57:01			-	-	-	-	-	-
13	9025	JESTHOM	*27:05			-	-	-	-	-	-
14	9071	OLGA	*15:01	*15:20		-	-	-	-	-	+
15	9075	DKB	*40:01			-	-	-	-	-	-
16	9037	SWEIG007	*40:02			-	-	-	-	-	-
17	9282	CTM3953540	*08:01	*55:01		-	-	-	-	-	+
18	9257	32367	*14:01	*56:01		-	-	-	-	-	+
19	9038	BM16	*18:01			-	-	-	-	-	-
20	9059	SLE005	*40:01			-	-	-	-	-	-
21	9064	AMALA	*15:01			-	-	-	-	-	-
22	9056	KOSE	*35:03			-	-	-	-	-	+
23	9124	IHL	*40:02	*56:02		-	-	-	-	-	+
24	9035	JBUSH	*38:01			-	-	-	-	-	-
25	9049	IBW9	*14:02			-	-	-	-	-	-
26	9285	WT49	*58:01			-	-	-	-	-	+
27	9191	CH1007	*07:05	*51:01		-	-	-	-	-	-
28	9320	BEL5GB	*44:02	*44:03		-	-	-	-	-	-
29	9050	MOU	*44:03			-	-	-	-	-	-
30	9021	RSH	*42:01			-	-	-	-	-	-
31	9019	DUCAF	*18:01			-	-	-	-	-	-
32	9297	HAG	*41:02			-	-	-	-	-	-
33	9098	MT14B	*40:01			-	-	-	-	-	-
34	9104	DHIF	*38:01			-	-	-	-	-	-
35	9302	SSTO	*44:02			-	-	-	-	-	-
36	9024	KT17	*15:01	*35:01		-	-	-	-	-	+
37	9065	HHKB	*07:02			-	-	-	-	-	-
38	9099	LZL	*15:01			-	-	-	-	-	-
39	9315	CML	*08:01	*27:05		-	-	-	-	-	-
40	9134	WHONP199	*13:02	*46:01		-	-	-	-	-	+
41	9055	H0301	*14:02			-	-	-	-	-	-
42	9066	TAB089	*46:01			-	-	-	-	-	-
43	9076	T7526	*46:01			-	-	-	-	-	-
44	9057	TEM	*38:01			-	-	-	-	-	-
45	9239	SHJO	*42:01	*50:01		-	-	-	-	-	+
46	9013	SCHU	*07:02			-	-	-	-	-	-
47	9045	TUBO	*51:01			-	-	-	-	-	-
48	9303	TER-ND	*35:01	*44:03		-	-	-	-	-	+

101.572-12 – including *Taq* pol., IFU-01

101.572-12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: 5E0

Lot-specific information

¹The provided cell line HLA specificities are retrieved from the <http://www.ihwg.org/hla> web site. The specificity of an individual cell line may thus be subject to change.

²The specificity of each primer solution in the kit has been tested against 48 well characterized cell line DNAs and where applicable, additional cell line DNAs.

No DNAs carrying the alleles to be amplified by primer solutions 3 to 5, 7 to 10 and 12 to 21 were available. The specificities of the primers in primer solutions 3 to 5, 7, 8, 10, 12 to 17 and 19 were tested by separately adding additional 5'-primers, respectively 3'-primers. In primer solutions 9 and 18 it was only possible to test the 3'-primers, the 5'-primers were not possible to test.

In primer solutions 20 and 21 it was only possible to test the 5'-primer, the 3'-primers were not possible to test.

In primer solutions 3 to 7, 10 to 12, 15 and 19 one to three 3'-primers were not possible to test, and in primer solutions 4, 5, 8, 14 and 15 one to three 5'-primers were not possible to test.

101.572-12 – including *Taq* pol., IFU-01

101.572-12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: 5E0

Lot-specific information

101.572-12 – including *Taq* pol., IFU-01

101.572-12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **5E0**

Lot-specific information

ADDRESSES:

Manufacturer:

Olerup SSP AB, Franzengatan 5, SE-112 51 Stockholm, Sweden.

Tel: +46-8-717 88 27

Fax: +46-8-717 88 18

E-mail: olerup-se@caredx.com

Web page: <http://www.olerup.com>

Distributed by:

Olerup GmbH, Löwengasse 47 / 6, AT-1030 Vienna, Austria.

Tel: +43-1-710 15 00

Fax: +43-1-710 15 00 10

E-mail: olerup-at@caredx.com

Web page: <http://www.olerup.com>

Olerup Inc., 901 S. Bolmar St., Suite R, West Chester, PA 19382

Tel: 1-877-OLERUP1

Fax: 610-344-7989

E-mail: olerup-us@caredx.com

Web page: <http://www.olerup.com>

For information on *Olerup* distributors worldwide, contact **Olerup GmbH**.