

Olerup SSP[®] DQA1

Product number: 101.231-24/04 – including *Taq* pol.
101.231-24u/04u – without *Taq* pol.
Lot number: 16F
Expiry date: 2010-October-01
Number of tests: 24 tests – Product No. 101.231-24
4 tests – Product No. 101.231-04
Number of wells per test: 32

CHANGES COMPARED TO THE PREVIOUS DQA1 LOT (Y46):

The DQA1 primer set is unchanged.

THE NUMBER OF WELLS is unchanged.

ALLELE COVERAGE:

All the currently recognized DQA1 alleles, i.e. DQA1*0101 to 0107, DQA1*0201, DQA1*0301 to 0303, DQA1*0401 to DQA1*0404, DQA1*0501 to 0509 and DQA1*0601 to DQA1*0602, give rise to unique amplification patterns; www.ebi.ac.uk/imgt/hla, 2008-October-10, release 2.23.0.

RESOLUTION IN DQA1 HOMOZYGOTES:

Good.

MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:

No comments received.