

101.231-24/04 – including *Taq* polymerase  
 101.231-24u/04u – without *Taq* polymerase

Lot No.: **3F0**

## Olerup SSP<sup>®</sup> DQA1

**Product number:** 101.231-24/04 – including *Taq* pol.  
 101.231-24u/04u – without *Taq* pol.  
**Lot number:** 3F0  
**Expiry date:** 2021-06-01  
**Number of tests:** 24 tests – Product No. 101.231-24/24u  
 4 tests – Product No. 101.231-04/04u  
**Number of wells per test:** 31+1

### CHANGES COMPARED TO THE PREVIOUS DQA1 LOT (9D6):

Well	5'-primer	3'-primer	rationale
19	-	Added	3'-primer added for the DQA1*01:15N allele.
28	Added	Added	Primer pair added for the DQA1*01:14 allele.

THE NUMBER OF WELLS is unchanged.

### ALLELE COVERAGE:

All the currently recognized DQA1 alleles<sup>1</sup>, i.e. DQA1\*01:01 to 01:15N, DQA1\*02:01, DQA1\*03:01 to 03:03, DQA1\*04:01 to DQA1\*04:04, DQA1\*05:01 to 05:11 and DQA1\*06:01 to 06:02, give rise to unique amplification patterns; [www.ebi.ac.uk/imgt/hla](http://www.ebi.ac.uk/imgt/hla), 2017-January-20, release 3.27.0.

<sup>1</sup>Alleles that have been deleted from or renamed in the official WHO HLA Nomenclature up to and including the last IMGT/HLA database release can be retrieved from web page <http://hla.alleles.org/alleles/deleted.html>.

### RESOLUTION IN DQA1 HOMOZYGOTES:

Good.

### MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:

No comments received.

Change in revision R01 compared to R00:

1. The expiration date has been altered due to extension of shelf-life.