

101.231-24/04 – including *Taq* polymerase

101.231-24u/04u – without *Taq* polymerase

Lot No.: 7L8

Olerup SSP[®] DQA1

Product number: 101.231-24/04 – including *Taq* pol.
101.231-24u/04u – without *Taq* pol.

Lot number: 7L8

Expiry date: 2025-01-01

Number of tests: 24 tests – Product No. 101.231-24/24u
4 tests – Product No. 101.231-04/04u

Number of wells per test: 31+1

CHANGES COMPARED TO THE PREVIOUS DQA1 LOT (9K2):

Well	5'-primer	3'-primer	rationale
29	Added	Added	Primer pair added for improving the resolution of the DQA1*05:02 allele.

THE NUMBER OF WELLS is unchanged.

ALLELE COVERAGE:

All the DQA1 alleles^{1,2}, i.e. **DQA1*01:01 to 01:52, DQA1*02:01 to DQA1*02:14, DQA1*03:01 to 03:16, DQA1*04:01 to DQA1*04:08, DQA1*05:01 to 05:29Q and DQA1*06:01 to 06:02**, recognized by the HLA Nomenclature Committee in October 2020 will be amplified by the primers in the DQA1 typing kit, www.ebi.ac.uk/imgt/hla, 2020-October-15, release 3.42.0.

¹Alleles that have been deleted from or renamed in the official WHO HLA Nomenclature up to and including the last IMGT/HLA database release can be retrieved from web page <http://hla.alleles.org/alleles/deleted.html>.

²The DQA1 kit enables separation of DQA1 alleles as listed in the IMGT/HLA database 3.34.0.

RESOLUTION IN DQA1 HOMOZYGOTES:

Good.

MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:

No comments received.