

Olerup SSP[®] HLA-A*30

Product number: 101.429-12 – including *Taq* polymerase
101.429-12u – without *Taq* polymerase
Lot number: 78F
Expiry date: 2011-March-01
Number of tests: 12
Number of wells per test: 24

CHANGES COMPARED TO THE PREVIOUS HLA-A*30 LOT (38E).

Well	5'-primer	3'-primer	rationale
3	-	Added	Primer added for the A*3027N allele.
7	Added	Added	Primer pair added for the A*3028 allele.
22	New	New	New primer pair for the A*3027N allele.
23	New	New	Primer pair added for the A*3026 allele.
24	New	New	Primer pair added for the A*3025 allele.

THE NUMBER OF WELLS has been increased from 21 to 24.

ALLELE COVERAGE:

A*3001 to A*3028, i.e. all the currently recognized HLA-A*30 alleles, give rise to unique amplification patterns; www.ebi.ac.uk/imgt/hla, 2009-January-16, release 2.24.0.

RESOLUTION IN HLA-A*30 HOMO- AND HETEROZYGOTES:

The A*3001,3001, A*3001,3002 and A*3002,3002 genotypes give rise to unique amplification patterns.

INFLUENCE ON THE INTERPRETATION OF HLA-A*30 SUBTYPINGS BY NON-HLA-A*30 ALLELES:

None frequently occurring.

MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:

No comments received.