

Olerup SSP[®] HLA-Cw*02

Product number: 101.622-12 – including *Taq* polymerase
101.622-12u – without *Taq* polymerase
Lot number: 74F
Expiry date: 2011-April-01
Number of tests: 12
Number of wells per test: 16

CHANGES COMPARED TO THE PREVIOUS HLA-Cw*02 LOT (12E):

Well	5'-primer	3'-primer	rationale
4	Added	-	Primer added for the Cw*0221 allele.
5	Added	-	Primer added for the Cw*0222 allele.
10	Added	-	Primer added for the Cw*0221 allele.
11	-	-	Exchanged positive control primer pair.
12	Added	-	Primer added for increased yield of specific PCR product.
13	Added	Added	Primer pair added for the Cw*0223 allele.
15	Added	Added	Primer pair added for the Cw*0221 allele.

THE NUMBER OF WELLS is unchanged.

ALLELE COVERAGE:

Cw*0202 to Cw*0223, i.e. all the currently recognized HLA-Cw*02 alleles, give rise to unique amplification patterns; www.ebi.ac.uk/imgt/hla, 2009-January-16, release 2.24.0.

RESOLUTION IN HLA-Cw*02 HOMO- AND HETEROZYGOTES:

The Cw*0202,0202 and Cw*0202,0210 genotypes give rise to identical amplification patterns.

INFLUENCE ON THE INTERPRETATION OF HLA-Cw*02 SUBTYPINGS BY NON-HLA-Cw*02 ALLELES:

None frequently occurring.

MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:

No comments received.